



INSTRUMENT PROCESSING SHEET

Agency Ocala PD

S/N 80-006214

Florida Department of Law Enforcement

Date In 1/16/2018

DI Completion Date 1/18/18

1/18/18 Ship

P/U H/D CMI EE

Intake Performed By JA

Annual
 Registration
 Return from CMI / EE

Visual Inspection:
 Case Handle
 Keyboard Dry Gas Shelf
 Feet Breath Tube
 Ports Screws Tight

Other Equipment/ Accessories:
 Power cord Printer Cable
 Static Bag 12V DC Cable

Notes: _____

Quality Checks Performed By JA

Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 203
 Flow Verification (L/s)

Flow Column # ATP105
 32 mm 148 (.139 - .169)
 36 mm 171 (.156 - .190)
 53 mm 242 (.228 - .278)
 103 mm 512 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Flow Calibration Performed By _____

Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)

Flow Column # _____
 32 mm _____ (.139 - .169)
 36 mm _____ (.156 - .190)
 53 mm _____ (.228 - .278)
 103 mm _____ (.447 - .547)

Final Release Date

FDLE

JAN 18 2018

Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	G2835	201707D 07/25/2019
0.080	SD1013	201707E 07/25/2019
0.200	SD1025	201707C 07/24/2019
0.080 DGS	N/A	AG-708807 3/29/19

Maintenance Performed By JA

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other Software Reloaded

Temperature Checks Performed By JA

Lab Temp °C 20.9
 External Digital Therm. ID#: 300504
 34°C +/-2 Serial #: G2835
 34°C +/-2 Serial #: SD1013
 34°C +/-2 Serial #: SD1025

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By JA

Barometric Pressure ID# 28427
 Gauge 1029 Instrument 1030
 Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	G2880
Interferent	G2840
0.050	G2835
0.080	SD1013
0.200	SD1025

Attachments

Form 41 Post-Stability Checks
 Stability Checks Flow Calibration
 Calibration Certificate Form 40
 Calibration Adjustment Other Cal Cert

Notes/Suggested Service: _____

ATP Notified me of a data error w/ forms. I updated software to correct issue.

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

JA 1/18/2018 J. Graham 1/18/18
 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OCALA POLICE DEPT
Time of Inspection: 10:16

Date of Inspection: 01/18/2018

Serial Number: 80-006214
Software: 8100.27

✓

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG708807 Exp: 03/29/2019
0.000	0.051	0.084	0.201	0.081
0.000	0.051	0.084	0.204	0.081
0.000	0.051	0.084	0.202	0.080
0.000	0.051	0.083	0.203	0.080
0.000	0.051	0.083	0.203	0.080
0.000	0.051	0.083	0.203	0.080
0.000	0.051	0.084	0.203	0.080
0.000	0.052	0.084	0.203	0.080
0.000	0.051	0.083	0.203	0.080
0.000	0.051	0.084	0.204	0.080

Standard Deviations	0.0003	0.0005	0.0008	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

gpm

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Handwritten Signature]

JAKE L SHANAHAN

Signature and Printed Name

01/18/2018
Date

*1/18/18
JA*

80-006214
 Stability Checks
 1/18/18

OCALA POLICE DEPT
 Intoxilyzer - Alcotest Analyzer
 Model 8000 SN 80-006214
 01/18/2018
 Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	07:35
Control Test	0.079	07:35
Air Blank	0.000	07:37
Control Test	0.080	07:38
Air Blank	0.080	07:39
Control Test	0.080	07:39
Air Blank	0.000	07:39
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DWS

Operator's Signature

1/18/18
 SD

OCALA POLICE DEPT
 Intoxilyzer - Alcotest Analyzer
 Model 8000 SN 80-006214
 01/18/2018
 Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	07:47
Control Test	0.049	07:47
Air Blank	0.000	07:48
Control Test	0.050	07:48
Air Blank	0.000	07:49
Control Test	0.049	07:49
Air Blank	0.000	07:50
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature

OCALA POLICE DEPT
 Intoxilyzer - Alcotest Analyzer
 Model 8000 SN 80-006214
 01/18/2018
 Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	07:54
Control Test	0.080	07:54
Air Blank	0.000	07:55
Control Test	0.082	07:55
Air Blank	0.000	07:55
Control Test	0.080	07:56
Air Blank	0.000	07:56
Control Test Stats		
Average	0.0807	
Std Dev	0.0012	
Rel Std Dev(%)	1.4314	

Operator's Signature

OCALA POLICE DEPT
 Intoxilyzer - Alcotest Analyzer
 Model 8000 SN 80-006214
 01/18/2018
 Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	08:04
Control Test	0.201	08:04
Air Blank	0.000	08:05
Control Test	0.200	08:05
Air Blank	0.000	08:06
Control Test	0.201	08:07
Air Blank	0.000	08:07
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

Operator's Signature

ADM



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006214, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006214</u>	UNCERTAINTY* ±
Owning Agency:	<u>OCALA POLICE DEPT</u>	0.05 g/ 210 L
Calibration Date:	<u>01/18/2018</u>	0.08 g/ 210 L
Calibration Time:	<u>10:16</u>	0.20 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.

Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/18/2018

Date

JAKE L SHANAHAN,
 Department Inspector

FDLE/ATP Form 69 January 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

1/18/18
JLS

Blm

Jake, here is the instrument I emailed you about. Just a refresher, when I go to put in agency code to conduct breath test there is a DATA error. Also, I conducted the December AI and when I went to print it showed that ATP form 40 was not found. If you can email me the AI report I would appreciate it. Also just FYI the upload was not completed.

Thanks William. *Ocala Pd*
Cell 352 427 9654

William

RECEIVED
JAN 16 2018
FDLE
Alcohol Testing Program

*1/18/18
JD*