



**INSTRUMENT PROCESSING SHEET**

Agency Broward County Sheriff's Office

S/N 80-006924

Florida Department of Law Enforcement

Date In 11/27/2017 DI Completion Date 1/17/2018

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DEAL</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DEAL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>167</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.183</u> (.156 - .190) 53 mm <u>.253</u> (.228 - .278) 103 mm <u>.523</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3963</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG626604 09/22/2018</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3963	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG626604 09/22/2018	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>DEAL</u> <input checked="" type="checkbox"/> Lab Temp °C <u>20.88c</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3963</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	SD3967	201707D 07/25/2019																																																												
0.080	SD3963	201707E 07/25/2019																																																												
0.200	SD3969	201707C 07/24/2019																																																												
0.080 DGS	N/A	AG626604 09/22/2018																																																												
<b>Final Release Date</b> <b>FDLE</b>  <b>JAN 30 2018</b>  <b>Alcohol Testing Program</b>																																																														
<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>DEAL</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1022</u> Instrument <u>1023</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3963</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3963	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	SD3965																																																													
Interferent	SD3966																																																													
0.050	SD3967																																																													
0.080	SD3963																																																													
0.200	SD3969																																																													
Notes/Suggested Service: <u>E-mailed (2018 Annual)</u> <div style="border: 1px solid blue; padding: 5px; display: inline-block; color: blue; font-weight: bold; font-size: 1.2em;"> <input checked="" type="checkbox"/> APPROVED         </div>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  <div style="display: flex; justify-content: space-between;"> <div style="text-align: center;"> <u>Poam 1/29/18</u>              Tech Review / Date           </div> <div style="text-align: center;"> <u>J. Debra 1/30/18</u>              Admin Review / Date           </div> </div>																																																													

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 09:17

Date of Inspection: 01/17/2018

Serial Number: 80-006924  
Software: 8100.27

*loc*

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.079
0.000	0.048	0.081	0.201	0.080
0.000	0.048	0.080	0.201	0.079
0.000	0.049	0.081	0.201	0.079
0.000	0.048	0.080	0.200	0.079
0.000	0.048	0.081	0.199	0.080
0.000	0.048	0.080	0.199	0.078
0.000	0.049	0.081	0.200	0.079

Standard Deviations	0.0004	0.0007	0.0008	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

*QPM*

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David E Reyes-Rivera*

DAVID E REYES-RIVERA

Signature and Printed Name

01/17/2018  
Date

*1/30/18  
JA*

<b>TYPE OF TEST</b>	<b>SERIAL NUMBER</b>	<b>AGENCY</b>	<b>DATE</b>	<b>PERFORMED BY</b>
Stabilities	80-006924	Broward County Sheriff's Office	01/17/2018	<i>Dee</i>

0.05g/210L	0.08g/210L	0.20g/210L	0.08g/210L	DGS 0.08g/210L																																																																																																																																																
<b>SN: SD3967 Temp: 34.06c</b> <b>0.047 to 0.053</b> <input checked="" type="checkbox"/>	<b>SN: SD3963 Temp: 34.06c</b> <b>0.077 to 0.083</b> <input checked="" type="checkbox"/>	<b>SN: SD3969 Temp: 34.08c</b> <b>0.194 to 0.206</b> <input checked="" type="checkbox"/>	<b>SN: SD3963 Temp: 34.06c</b> <b>0.077 to 0.083</b> <input checked="" type="checkbox"/>	<b>Lot AG626604</b> <b>0.077 to 0.083</b> <input checked="" type="checkbox"/>																																																																																																																																																
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 01/17/2018 Software: 8100.27	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 01/17/2018 Software: 8100.27	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 01/17/2018 Software: 8100.27	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 01/17/2018 Software: 8100.27	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 01/17/2018 Software: 8100.27																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:34</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:35</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:36</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:38</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:38</td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:34	Control Test	0.049	06:35	Air Blank	0.000	06:35	Control Test	0.049	06:36	Air Blank	0.000	06:36	Control Test	0.049	06:37	Air Blank	0.000	06:38	Control Test	0.049	06:38	Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:40</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:41</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:42</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:43</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:43</td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:40	Control Test	0.080	06:40	Air Blank	0.000	06:41	Control Test	0.080	06:42	Air Blank	0.000	06:42	Control Test	0.080	06:43	Air Blank	0.000	06:43	Control Test	0.080	06:43	Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:45</td></tr> <tr><td>Control Test</td><td>0.198</td><td>06:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:46</td></tr> <tr><td>Control Test</td><td>0.197</td><td>06:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:47</td></tr> <tr><td>Control Test</td><td>0.197</td><td>06:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr> <tr><td>Control Test</td><td>0.197</td><td>06:48</td></tr> <tr><td>Average</td><td>0.1973</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:45	Control Test	0.198	06:45	Air Blank	0.000	06:46	Control Test	0.197	06:47	Air Blank	0.000	06:47	Control Test	0.197	06:48	Air Blank	0.000	06:48	Control Test	0.197	06:48	Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:51</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:52</td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:50	Control Test	0.079	06:50	Air Blank	0.000	06:51	Control Test	0.079	06:51	Air Blank	0.000	06:52	Control Test	0.078	06:52	Air Blank	0.000	06:52	Control Test	0.078	06:52	Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		<p style="text-align: right;"><i>Dee</i> Operator's Signature</p>
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	06:34																																																																																																																																																		
Control Test	0.049	06:35																																																																																																																																																		
Air Blank	0.000	06:35																																																																																																																																																		
Control Test	0.049	06:36																																																																																																																																																		
Air Blank	0.000	06:36																																																																																																																																																		
Control Test	0.049	06:37																																																																																																																																																		
Air Blank	0.000	06:38																																																																																																																																																		
Control Test	0.049	06:38																																																																																																																																																		
Average	0.0490																																																																																																																																																			
Std Dev	0.0000																																																																																																																																																			
Rel Std Dev(%)	0.0000																																																																																																																																																			
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	06:40																																																																																																																																																		
Control Test	0.080	06:40																																																																																																																																																		
Air Blank	0.000	06:41																																																																																																																																																		
Control Test	0.080	06:42																																																																																																																																																		
Air Blank	0.000	06:42																																																																																																																																																		
Control Test	0.080	06:43																																																																																																																																																		
Air Blank	0.000	06:43																																																																																																																																																		
Control Test	0.080	06:43																																																																																																																																																		
Average	0.0800																																																																																																																																																			
Std Dev	0.0000																																																																																																																																																			
Rel Std Dev(%)	0.0000																																																																																																																																																			
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	06:45																																																																																																																																																		
Control Test	0.198	06:45																																																																																																																																																		
Air Blank	0.000	06:46																																																																																																																																																		
Control Test	0.197	06:47																																																																																																																																																		
Air Blank	0.000	06:47																																																																																																																																																		
Control Test	0.197	06:48																																																																																																																																																		
Air Blank	0.000	06:48																																																																																																																																																		
Control Test	0.197	06:48																																																																																																																																																		
Average	0.1973																																																																																																																																																			
Std Dev	0.0006																																																																																																																																																			
Rel Std Dev(%)	0.2926																																																																																																																																																			
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	06:50																																																																																																																																																		
Control Test	0.079	06:50																																																																																																																																																		
Air Blank	0.000	06:51																																																																																																																																																		
Control Test	0.079	06:51																																																																																																																																																		
Air Blank	0.000	06:52																																																																																																																																																		
Control Test	0.078	06:52																																																																																																																																																		
Air Blank	0.000	06:52																																																																																																																																																		
Control Test	0.078	06:52																																																																																																																																																		
Average	0.0787																																																																																																																																																			
Std Dev	0.0006																																																																																																																																																			
Rel Std Dev(%)	0.7339																																																																																																																																																			
<p style="text-align: right;"><i>Dee</i> Operator's Signature</p>	<p style="text-align: right;"><i>Dee</i> Operator's Signature</p>	<p style="text-align: right;"><i>Dee</i> Operator's Signature</p>	<p style="text-align: right;"><i>Dee</i> Operator's Signature</p>	<p style="text-align: right;"><i>Dee</i> Operator's Signature</p>																																																																																																																																																

*1/30/18*  
*Dee*

*Dee*



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 4700 Terminal Drive, Suite 1  
 Ft. Myers, FL 33907

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006924, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006924</u>	UNCERTAINTY* ±
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.05 g/ 210 L 0.004
Calibration Date:	<u>01/17/2018</u>	0.08 g/ 210 L 0.005
Calibration Time:	<u>09:17</u>	0.20 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.  
 Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/17/2018 Date *David Reyes-Rivera*  
**DAVID E REYES-RIVERA,**  
 Department Inspector

FDLE/ATP Form 69 January 2018  
 Issuing Authority: Alcohol Testing Program

*1/30/18*  
*DR*  
*Page*

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 06:04

Date of Inspection: 01/17/2018

Serial Number: 80-006924  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED

*DM*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David Reyes Rivera*

DAVID E REYES-RIVERA

Signature and Printed Name

01/17/2018  
Date

*1/30/18  
JA*