FDLE
Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Florida Highway Patrol Instrument Serial Number: 80-006261

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
☐ Registration ☐ Request for Registration Attached ☐ Annual ☐ Inspection After Repair ☐ Other: _____ ☐ Barometric Pressure: _____ Gauge _____ Instrument ☐ Instrument Set Up Verified	☑ Agency Inspection Notification ☐ Agency Contact ☐ Other _____ Date of Notification/Contact: 07/30/2018 Agency Inspector: Trooper Joseph Farley	The following records were audited: ☑ Agency Inspection Reports ☐ Breath Test Instrument Repair/Maintenance Records ☐ Instrument Registration ☐ Dry Gas Standard Certificate(s) of Analysis ☐ Other: _____ ☐ Other: _____

Comments:
Trooper Joseph Farley was contacted on July 30th, 2018 to notify him that the Alcohol Reference Solutions used during agency inspections in 2017 and 2018 are not approved by the FDLE for use. Trooper Farley was informed that this is a Rules violation and corrective measures are

Instrument/Area:	Equipment:	Supplies:
☐ Clean/Dry ☐ Secure ☐ Limited Access ☐ Other: _____	☐ Proper Number of Simulators ☐ Checked Simulators for Air Leak Resistant Seal and Proper Temperature ☐ Class A Glassware	☐ Distilled/Deionized Water ☐ Mouth Alcohol Solution ☐ Acetone Stock Solution ☑ Alcohol Reference Solution ☐ Dry Gas Standard ☐ Mouth Pieces

Comments:
required immediately. Trooper Farley and his co-workers perform inspections on the three (3) instruments assigned together. The same solutions are used for inspections on all of their instruments. Farley was instructed to do an inspection with approved solutions immediately.

Action:	Comments:
☐ Instrument Complies with Chapter 11D-8, FAC ☑ Instrument Does Not Comply with Chapter 11D-8, FAC	The instrument appears to have been working accurately and reliably regardless of the unapproved nature of the solutions used.
☐ Remain in Evidentiary Use ☑ Return to/Place into Evidentiary Use ☐ Remove from Evidentiary Use ☐ Remain Out of Evidentiary Use	
☑ Conduct an Agency Inspection Before Evidentiary Use	

Signature of Alcohol Testing Program Staff Member Date: 7/31/18

Original – FDLE Copy – Agency