

Bell, Danielle

80-000833 RMB

From: RG SKIPPER <rskipper@hcsso.tampa.fl.us>
Sent: Monday, July 16, 2018 9:23 AM
To: Bell, Danielle
Cc: TIMOTHY BELL
Subject: 80-000833 Removed From Service

Follow Up Flag: FollowUp
Flag Status: Flagged

Inspector Bell,

Instrument 80-000833 was removed from evidential service this date due to Diagnostic Fail (Temperature Regulation Fail).

An "out of service" agency inspection cannot be conducted due to this condition.

This instrument will be transported to Enforcement Electronics for servicing.

This email memorializes our telephone conversation, this date.

--

Roger G. Skipper, CBTU Supervisor
Hillsborough County Sheriff's Office
Central Breath Testing Unit
1201 Orient Road
Tampa, Florida 33619

813-247-8361 (office)
813-242-1870 (fax)

Public Disclosure:

Under Florida law, email addresses are public record. If you do not want your e-mail address released in response to a public records request, do not send electronic mail to this entity. Instead, contact the Hillsborough County Sheriff's Office via telephone at 813-247-8000 or US Mail at P.O. Box 3371, Tampa, FL 33601.



RECEIVED

FEB 16 2018

FDLE

Alcohol Testing Program

01/22/18 20:04:18

Ticket #026263

Status Open
Priority Normal
Department DMB ORJ
Create Date 01/22/18 15:04:12

Name TIMOTHY BELL
Email tbell@hcsso.tampa.fl.us
Phone (813) 936-9099
Source Web

Ticket Details

Department: ORJ
Location: ORJ - Building 10 - CBT

CBT Staff Lockers Keys/Locks

01/22/18 15:04:12

TIMOTHY BELL

Keys to some of the staff lockers in CBT are not working. Example: locker #8. Also, need a spare key for locker #8 and #9. See Mr. Skipper as he can get you access to the restricted access area and has the set of locker keys.