



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

October 20, 2018

Mrs. Shayla Platt
Florida Department of Law Enforcement
Alcohol Testing Program
PO Box 1489
Tallahassee, FL 32302

REF: October 2, 2018 Inspection

Dear Mrs. Platt:

I am enclosing a copy of the Agency Inspection Reports for the month of October 22, 2018. Instrument 80-001043 and 80-001054 both have passed their inspections.

A handwritten signature in cursive, appearing to read "Sgt. S. Abeln".

Sgt. S. Abeln
Agency Inspector

Cc: File

RECEIVED

OCT 25 2018

FDLE
Alcohol Testing Program



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

August 27, 2018

Mrs. Shayla Platt
Florida Department of Law Enforcement
Alcohol Testing Program
PO Box 1489
Tallahassee, FL 32302

REF: August 27, 2018 Inspection

Dear Mrs. Platt:

I am enclosing a copy of the Agency Inspection Reports for the month of August 2018. Instrument 80-001043 and 80-001054 both have passed their inspections. However, 80-001043 had a RFI issue when Deputy Carver came into the testing room just as I was starting the dry gas standard portion of my inspection with his radio on. I asked him to turn his radio off and I repeated that portion without any issue. It is documented on the enclosed Form 38.

Sgt. Gregory Glasco
Agency Inspector

Cc: File



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

RECEIVED

JUL 05 2018

FDLE
Alcohol Testing Program

June 29, 2018

Mrs. Shayla Platt
Florida Department of Law Enforcement
Alcohol Testing Program
PO Box 1489
Tallahassee, FL 32302

REF: June 29, 2018 Inspection

Dear Mrs. Platt:

I am enclosing a copy of the Agency Inspection Reports for the month of June 2018. Instrument 80-001043 and 80-001054 both have passed their inspections.

Sgt. Shon W. Abeln
Agency Inspector

Cc: File



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

May 27, 2018

Mrs. Shayla Platt
Florida Department of Law Enforcement
Alcohol Testing Program
PO Box 1489
Tallahassee, FL 32302

REF May 27, 2018 Inspection

Dear Mrs. Platt:

I am enclosing a copy of the Agency Inspection Reports for the month of May 2018. Instrument 80-001043 and 80-001054 both have passed their inspections.

A handwritten signature in black ink, appearing to read "Shon W. Abeln".

Sgt. Shon W. Abeln
Agency Inspector

Cc: File



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

January 27 2018

Mrs. Shayla Platt
Florida Department of Law Enforcement
Alcohol Testing Program
PO Box 1489
Tallahassee, FL 32302

REF January 27, 2018 Inspection

Dear Mrs. Platt:

I am enclosing a copy of the Agency Inspection Reports for the month of December 2017. Instrument 80-001043 and 80-001054 both have passed their inspections. However instrument 80-111054 has a recorded in compliance record that is false. I had too much mouth alcohol and got a range exceeded error. I improperly closed the inspection without any further testing on THAT inspection and uploaded it. I then inspected the instrument again with no problems.

Sgt. G.Glasco
Agency Inspector

Cc: File

RECEIVED
FEB 01 2018
FDLE
Alcohol Testing Program