

831  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO  
Time of Inspection: 00:47

Date of Inspection: 04/26/2019

Serial Number: 80-000831  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06717080A7 Exp: 05/05/2019
/ 0.000	0.046	0.077	0.191	0.078
/ 0.000	0.047	0.077	0.193	0.077
/ 0.000	0.047	0.078	0.195	0.078

Number of Simulators Used: 5

Remarks:

00: Ambient Fail. ALLOWED ROOM TO AIR OUT AND RE-TESTED

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D/S C.  5/14/16

CHRISTOPHER L THOMAS

Signature and Printed Name

04/26/2019  
Date