

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY BEACH PD
Time of Inspection: 07:26

Date of Inspection: 05/07/2019

Serial Number: 80-000838
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes <i>AD</i>	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:989917 Exp: 07/25/2021
0.000	0.049	0.079	0.196	0.079
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.079	0.196	0.080

Number of Simulators Used: 5

Remarks:

TIME WAS 10 MINS SLOW *AD*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Diane M. Cheri

DIANE M CHERI
Signature and Printed Name

05/07/2019
Date