

Florida Department of Law Enforcement
 Alcohol Testing Program 3331X

AGENCY INSPECTION 1P

Jefferson County

Ser. 80-00954

| | | |
|--|--|--|
| | | |
|--|--|--|

Diagnostic Check (Pre-Inspection): OK No
Yes
 Mouth Alcohol Test: Slope Not Met Yes
 Interferent Detect Test: Interferent Detect Yes
 Diagnostic Check (Post-Inspection): OK *v4T Yes
No

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 00919080A2 Exp: 03/05/2021 |
|----------------------------|--|--|--|--|
| 0.000 | 0.047 | 0.078 | 0.196 | 0.000 / 0.000 |
| 0.000 | 0.046 | 0.078 | 0.196 | 0.000 / 0.079 |
| 0.000 | 0.047 | 0.078 | 0.195 | 0.000 / 0.079 |

Number of Simulators Used: 5

Remarks:


08: Control Outside Tolerance, Control Outside TOLERANCE Non-compliance: DRY GAS HOSE NOT CONNECTED DURING TEST.

I Reset the DRY GAS Bottle Before starting the 2nd Test.
 1st Test on 2nd Series resulted in 0.000, I then saw the issue being the DRY GAS HOSE NOT connected. 2nd & 3rd Tests were OK. Contacted P. Dat Murphy to inform him of the situation. He ok'd a new inspection.



The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 FRED C HENING

11/19

Date