

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS PARK PD
Time of Inspection: 10:55

Date of Inspection: 11/07/2019

Serial Number: 80-001042
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 05218080A3 Exp: 05/05/2020
0.000	0.048	0.078	0.194	0.000 / 0.079
0.000	0.049	0.079	0.195	0.080 / 0.079
0.000	0.049	0.079	0.195	0.079 / 0.079

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

I forgot to connect Dry Gas Cylinder during first attempt.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

Signature and Printed Name

JACOB M ROLLESTON

11/07/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

7/8/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS PARK PD
Time of Inspection: 16:39

Date of Inspection: 07/05/2019

Serial Number: 80-001042
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:05218080A3 Exp: 05/05/2020
0.000	0.047	0.077	0.195	0.079
0.000	0.048	0.078	0.196	0.078
0.000	0.049	0.078	0.196	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JACOB M ROLLESTON
Signature and Printed Name

07/05/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

7/8/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS PARK PD
Time of Inspection: 16:06

Date of Inspection: 07/05/2019

Serial Number: 80-001042
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048			
0.000	0.048			
0.000	0.048			

Number of Simulators Used: 5

Remarks: *Accidentally pressed Escape button during inspection*

The above instrument complies ☒ does not comply ☐ with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jacob M Rolleston

JACOB M ROLLESTON

Signature and Printed Name

07/05/2019
Date

Graham, Thomas

From: Jacob Rolleston <JRolleston@pinellas-park.com>
Sent: Friday, July 05, 2019 5:04 PM
To: Graham, Thomas
Subject: Pinellas Park PD Agency Inspection July 2019
Attachments: 1042 July 2019.pdf

80-001042

Follow Up Flag: FollowUp
Flag Status: Completed

7/8/19
JR

Sir,

Please see the attached agency inspection report. During the first attempt, the escape button was pressed due to human error and the first attempt was not successfully completed. Handwritten notes were completed on the paperwork.

--

Ofc. Jacob Rolleston #519
Special Operations Unit
Drug Recognition Expert
Traffic Homicide Investigator
Pinellas Park Police Department
7700 59th St N, Pinellas Park, FL 33781
727-541-0758

PLEASE NOTE: *Florida has a very broad public records law. Most written communications to or from city officials regarding city business are public records available to the public and media upon request. Your e-mail communications may therefore be subject to public disclosure.*