

Florida Department of Law Enforcement Alcohol Testing Program

7/24/19
JE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 21:43

Date of Inspection: 07/23/2019

Serial Number: 80-001048
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31517080A1 Exp: 01/05/2020 |
|----------------------------|--|--|--|--|
| 0.000 | 0.046 | 0.077 | 0.200 | 0.080 |
| 0.000 | 0.048 | 0.078 | 0.200 | 0.081 |
| 0.000 | 0.049 | 0.078 | 0.201 | 0.080 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

off. #25 *WARD*

ANTHONY W GETTLE

Signature and Printed Name

07/23/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

6/12/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 00:09

Date of Inspection: 06/12/2019

Serial Number: 80-001048
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31517080A1 Exp: 01/05/2020 |
|----------------------------|--|--|--|--|
| 0.000 | 0.048 | 0.077 | 0.199 | 0.080 |
| 0.000 | 0.049 | 0.078 | 0.199 | 0.081 |
| 0.000 | 0.049 | 0.078 | 0.199 | 0.080 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFC. [Signature]
#25 CAPD
ANTHONY W GETTLE

Signature and Printed Name

06/12/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

5/29/19
JE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 18:50

Date of Inspection: 05/28/2019

Serial Number: 80-001048
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020 |
|----------------------------|---|---|---|---|
| 0.000 | 0.048 | 0.077 | 0.199 | 0.081 |
| 0.000 | 0.049 | 0.078 | 0.199 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.081 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Anthony W Gettle #25 CAPD

ANTHONY W GETTLE

Signature and Printed Name

05/28/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

5/13/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 22:06

Date of Inspection: 04/27/2019

Serial Number: 80-001048
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020 |
|----------------------------|---|---|---|---|
| 0.000 | 0.048 | 0.079 | 0.199 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.081 |
| 0.000 | 0.049 | 0.080 | 0.199 | 0.081 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFF. [Signature]
#25 CAPD
ANTHONY W GETTLE

Signature and Printed Name

04/27/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

3/26/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 14:45

Date of Inspection: 03/22/2019

Serial Number: 80-001048
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31517080A1 Exp: 01/05/2020 |
|----------------------------|--|--|--|--|
| 0.000 | 0.048 | 0.081 | 0.200 | 0.082 |
| 0.000 | 0.049 | 0.081 | 0.200 | 0.082 |
| 0.000 | 0.050 | 0.082 | 0.201 | 0.082 |

Number of Simulators Used: 4

Remarks:

BACK IN SVC AFTER ANNUAL INSPECTION 2019

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OK. [Signature] #25 LAPD

ANTHONY W GETTLE

Signature and Printed Name

03/22/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

3/11/19
JW

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 21:18

Date of Inspection: 03/10/2019

Serial Number: 80-001048
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31517080A1 Exp: 01/05/2020 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.077 | 0.200 | 0.080 |
| 0.000 | 0.050 | 0.078 | 0.199 | 0.080 |
| 0.000 | 0.049 | 0.078 | 0.199 | 0.080 |

Number of Simulators Used: 4

Remarks:

Int Det: BAD ACETONE BOTTLE RETEST MADE CONTACT WITH TJ . BAD ACETONE BOTTLE/CONTACT W/TJ USED GO
OD ACETONE RETESTED

JW

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFC. CAPT

ANTHONY W GETTLE

Signature and Printed Name

03/10/2019
Date