

Florida Department of Law Enforcement Alcohol Testing Program

80-001062
J2

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Treasure Island Police

Serial Number:

Time of Inspection: 0830

Date of Inspection: 6/26/19

Software:

NO INSPECTION

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: N/A

Remarks:

Unable to perform agency inspection. The instrument will not power on. Sending instrument to FDLE for repair.

The above instrument complies () does not comply () with Chapter 11D-8, FAC. N/A *ejp220*

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name

6/26/19
 Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Treasure Island Police

Serial Number:

Time of Inspection: 0830

Date of Inspection: 6/26/19

Software: NO INSPECTION

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

ejp220

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

ejp220


Number of Simulators Used: N/A

Remarks:

Unable to perform agency inspection. The instrument will not power on. Sending instrument to FDLE for repair.

The above instrument complies () does not comply () with Chapter 11D-8, FAC. N/A *ejp220*

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Gabriella Pillucere

Signature and Printed Name

6/26/19
Date