

# Florida Department of Law Enforcement Alcohol Testing Program

6/14/19  
JTB

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLLYWOOD PD

Time of Inspection: 12:02

Date of Inspection: 06/13/2019

Serial Number: 80-001063

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 06/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1031121 Exp: 10/12/2021
0.000	0.051	0.081	0.206	0.078
0.000	0.051	0.082	0.206	0.079
0.000	0.051	0.082	0.206	0.079

Number of Simulators Used: 5

**Remarks:**

BSO Y WASHINGTON AND A FRAZIER WERE PRESENT

THE WRONG EXP. DATE WAS ENTERED FOR .08 ALCOHOL REFERENCE SOLUTION. THE CORRECT EXP. DATE IS 08/22/2020.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

#2736  
 \_\_\_\_\_  
 Signature and Printed Name

RAFAEL DERIET

06/13/2019  
 Date