

Florida Department of Law Enforcement Alcohol Testing Program

80-001071
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO
Time of Inspection: 01:28

Date of Inspection: 08/26/2019

Serial Number: 80-001071
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17618080A1 Exp: 08/05/2020
0.000	0.048	0.077	0.199	0.079
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.079

Number of Simulators Used: 5

Remarks:
Time-Date changed.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Charles M Harwood CHARLES M HARWOOD
Signature and Printed Name

08/26/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001071
BK

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: CLAY COUNTY SO
Time of Inspection: 08:24

Date of Inspection: 06/25/2019

Serial Number: 80-001071
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17618080A1 Exp: 03/05/2020
0.000	0.048	0.077	0.196	0.079
0.000	0.048	0.077	0.197	0.079
0.000	0.049	0.078	0.197	0.079

Number of Simulators Used: 5

Remarks:
RETURN TO SERVICE

RECEIVED
FDLE
02 019
Alcohol Testing Program

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Richard L. Patrone

Signature and Printed Name

RICHARD L PATRONE

06/25/2019
Date