

80-001081
TK

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:49

Date of Inspection: 11/20/2019

Serial Number: 80-001081
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | Yes | |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____ |
|----------------------------|---|---|---|--|
| 0.000 | 0.049 | 0.078 | 0.179 / 0.186 | |
| 0.000 | 0.050 | 0.079 | 0.185 / 0.192 | |
| 0.000 | 0.049 | 0.079 | 0.187 / 0.193 | |

Number of Simulators Used: 4

Remarks:

Time-Date changed. 20: Control Outside Tolerance TIGHTENED SIM LID, Control Ou. Non-compliance: .

Problem with the simulator. Inv. taken out of service until next inspection

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]
3932

BARRY NIXON

Signature and Printed Name

11/20/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001081
TSK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:57

Date of Inspection: 10/22/2019

Serial Number: 80-001081
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22419080A3 Exp: 10/05/2021 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.198 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.200 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.080 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BARRY NIXON
Signature and Printed Name

10/22/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001081
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 15:49

Date of Inspection: 08/20/2019

Serial Number: 80-001081
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#: Exp: | 0.20g/210L Test (g/210L) Lot#: Exp: | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp: |
|----------------------------|--|---|---|--|
| 0.000 | 0.000 / 0.000 | | | |
| 0.000 | 0.000 / 0.000 | | | |
| 0.000 | 0.000 / 0.000 | | | |

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:

changed 0.05g/210L solution but hit enter before the solution had come to temp.

Took unit out of service until next inspection

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 3932

BARRY NIXON

Signature and Printed Name

08/20/2019
Date