

1150  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 10:07

Date of Inspection: 11/18/2019

Serial Number: 80-001150  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1111579 Exp: 03/19/2022
0.000	0.049	0.078	0.199	0.000 / 0.082
0.000	0.048	0.078	0.200	0.000 / 0.081
0.000	0.049	0.078	0.199	0.000 / 0.082

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

DRY GAS NOT ATTACHED, re-attached / reported.  
Killy

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

GLENN CLINE

11/18/2019  
Date

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO

Time of Inspection: 08:17

Date of Inspection: 03/25/2019

Serial Number: 80-001150

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) <i>201808E</i> Lot#:201811C Exp: <i>11/13/2020</i> <i>08/22/2020</i>	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:840875 Exp: 09/19/2020
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.200	0.079


Number of Simulators Used: 5

Remarks:

Lot# for 0.08g/210L test was incorrect, Correct lot# was 201808E, EXPIRATION OF 08/22/2020

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GLENN CLINE

Signature and Printed Name

03/25/2019  
Date

1150  
SP

Florida Department of Law Enforcement  
Alcohol Testing Program

**AGENCY INSPECTION REPORT - INTOXILYZER 8000**

Agency: MANATEE COUNTY SO

Serial Number: 80-001150

Time of Inspection: 12:02

Date of Inspection: 04/24/2019

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 840875 Exp: 09/19/2020
0.000	0.048	0.079	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks: This was the second A.I. done for April. The forms had been "dumped" so I was unsure if the inspection was actually validated and/or saved.

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GLENN CLINE

Signature and Printed Name

04/24/2019

Date

1150  
8P

## Florida Department of Law Enforcement Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO

Serial Number: 80-001150

Time of Inspection: 17:03

Date of Inspection: 04/23/2019

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) <i>201811C</i> Lot#: <del>20811C</del> Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 840875 Exp: 09/19/2020
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081

Number of Simulators Used: 5

Remarks: Lot# for 0.08g/210L test was incorrect, Correct lot# was 201811C, EXPIRATION OF 11/13/2020

This was the first A.I. done for April. The forms had been "dumped" so I was unaware if the inspection was actually validated and/or done.

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GLENN CLINE

Signature and Printed Name

04/23/2019

Date

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 08:17

Date of Inspection: 03/25/2019

Serial Number: 80-001150  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) <i>201808E</i> <del>Lot#: 201811C</del> Exp: <i>08/22/2020</i>	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 240875 Exp: 09/19/2020
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.200	0.079

Number of Simulators Used: 5

Remarks:

Lot# for 0.08g/210L test was incorrect, Correct lot# was 201808E, EXPIRATION OF 08/22/2020

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name

03/25/2019  
Date