

Florida Department of Law Enforcement Alcohol Testing Program

80-001174
43K

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 10:57

Date of Inspection: 12/23/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.048	0.077	0.197	0.080
0.000	0.049	0.077	0.199	0.081
0.000	0.049	0.078	0.200	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 141

DEBORAH REASONER

Signature and Printed Name

12/23/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001174
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 11:29

Date of Inspection: 11/27/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 02618080A2 Exp: 03/05/2020
0.000	0.048	0.077	0.196	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt. [Signature] 141 DEBORAH REASONER
Signature and Printed Name

11/27/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001174
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 13:38

Date of Inspection: 10/30/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.049	0.077	0.196	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 141

DEBORAH REASONER

Signature and Printed Name

10/30/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001174
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 15:02

Date of Inspection: 09/26/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.047	0.078	0.198	0.080
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DEBORAH REASONER
Signature and Printed Name

09/26/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001174
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 13:49

Date of Inspection: 08/01/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.048	0.077	0.196	0.080
0.000	0.048	0.079	0.197	0.079
0.000	0.049	0.079	0.198	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DEBORAH REASONER

Signature and Printed Name

08/01/2019

Date

Florida Department of Law Enforcement Alcohol Testing Program

7/18/19
JED

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 12:57

Date of Inspection: 07/18/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.048	0.079	0.197	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DEBORAH REASONER

07/18/2019
Date

** AMENDED **
Florida Department of Law Enforcement
Alcohol Testing Program

7/18/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 15:16

Date of Inspection: 07/17/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000				
0.000				
0.000				

Number of Simulators Used: 4

Remarks:

Int Det: RE DO INSP. Non-compliance: INTF ISSUE.

LAST TX USED WRONG BOTTLE OF INTERFERENT. RE DID INSPECTION 12

The above instrument complies () does not comply (☒) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

Signature and Printed Name

DEBORAH REASONER

07/17/2019
Date



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Clermont PD

Instrument Serial Number: 80-001174

AGENCY INSPECTION DATA REVIEW		
Agency Inspector: <u>Deborah Reasoner</u>	Date of Inspection: <u>04/30/2019</u>	Time of Inspection: <u>00:26:18</u>
Agency Inspection Discrepancy: <input type="checkbox"/> Incomplete <input type="checkbox"/> Untimely/Not Received <input type="checkbox"/> Erroneous Information <input checked="" type="checkbox"/> Procedural <input type="checkbox"/> Other _____		
<input type="checkbox"/> Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.		
<input type="checkbox"/> Lot Number <input type="checkbox"/> Expiration Date for _____ g/ 210L <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard is <input type="checkbox"/> Incorrect <input type="checkbox"/> Expired.		
<input checked="" type="checkbox"/> FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The <input type="checkbox"/> REASON for repeating the following test(s); OR the <input checked="" type="checkbox"/> Possible Cause and Corrective Action Taken on the following test(s) was not recorded: <input type="checkbox"/> Alcohol Free Subject Test <input type="checkbox"/> Mouth Alcohol Test <input checked="" type="checkbox"/> Alcohol Free Test <input type="checkbox"/> Interferent Detect Test <input type="checkbox"/> 0.05 g/210L Test <input type="checkbox"/> 0.08 g/210L Test <input type="checkbox"/> 0.20 g/210L Test <input type="checkbox"/> 0.08 g/210L Dry Gas Standard Test		
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." <input type="checkbox"/> The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. <input type="checkbox"/> The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. <input type="checkbox"/> The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.		
<input type="checkbox"/> The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.		
<input type="checkbox"/> Other: _____		

OTHER ELECTRONIC DATA REVIEW	
<input type="checkbox"/> Login Records Date: _____	Comments: _____
<input type="checkbox"/> Cylinder Change Records Date: _____	
<input type="checkbox"/> Control Test Records Date: _____	
<input type="checkbox"/> Diagnostic Check Records Date: _____	

CORRECTIVE ACTION
<input checked="" type="checkbox"/> Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by <u>06/30/2019</u> (Date).
<input type="checkbox"/> Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
<input type="checkbox"/> Upload the Agency Inspection(s).
<input type="checkbox"/> Remove the instrument from evidentiary use until otherwise directed by the Department.
<input type="checkbox"/> No action required
<input type="checkbox"/> Other: _____


Signature of Alcohol Testing Program Staff Member

6/17/2019
Date

Graham, Thomas

From: Deborah Reasoner <DReasoner@clermontfl.org>
Sent: Thursday, June 20, 2019 7:27 PM
To: Graham, Thomas
Subject: RE: Data Review
Attachments: April 2019 amended.pdf

Follow Up Flag: FollowUp
Flag Status: Flagged



Debbie Reasoner
Police Sergeant
Clermont Police Department
Bravo Squad
DRE/THI
3600 US 27., Clermont, FL 34711
Tel 352-536-8461 | Fax 352-394-1644
Cell 352-348-0137
dreasoner@clermontfl.org
www.cityofclermontfl.gov



From: Graham, Thomas <ThomasGraham@fdle.state.fl.us>
Sent: Thursday, June 20, 2019 7:12 PM
To: Deborah Reasoner <DReasoner@clermontfl.org>
Subject: Re: Data Review

Thank you. I know that form looks confusing. If you have any further questions, please let me know.

TJ Graham, D-ABFT-FA
Quality Assurance Manager
Alcohol Testing Program
Florida Department of Law Enforcement
Office: (850) 617-1285
Cell: (850) 445-7938

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

Sent from my Verizon, Samsung Galaxy smartphone

----- Original message -----

From: Deborah Reasoner <DReasoner@clermontfl.org>
Date: 6/20/19 7:00 PM (GMT-05:00)
To: "Graham, Thomas" <ThomasGraham@fdle.state.fl.us>
Subject: RE: Data Review

Ok, I understand now. I will mark the Form as amended and put the explanation at the bottom.

I will email you the amended form. I apologize for that. I will ensure it doesn't happen again.



Debbie Reasoner
Police Sergeant
Clermont Police Department
Bravo Squad
DRE/THI
3600 US 27., Clermont, FL 34711
Tel 352-536-8461 | Fax 352-394-1644
Cell 352-348-0137
dreasoner@clermontfl.org
www.cityofclermontfl.gov



From: Graham, Thomas <ThomasGraham@fdle.state.fl.us>
Sent: Thursday, June 20, 2019 7:37 AM
To: Deborah Reasoner <DReasoner@clermontfl.org>
Subject: RE: Data Review

Sgt. Reasoner,

The email I sent you is related to an Agency Inspection you performed on instrument #80-001174 on 4/30/2019. During this inspection, you received an "Ambient Fail" while performing the Alcohol Free simulator test. Due to this operational message being obtained, you had to repeat the test. The instrument passed the second test and you continued to run simulators and deemed that 80-001174 was compliant with Rule 11D-8, F.A.C.

The ATP Form 40 has check boxes marked that indicate you need to write a comment on your Agency Inspection whenever you need to repeat a test. In this case, you had an "Ambient Fail". Did you have alcohol solutions open in the room? Is there a possibility that mouth alcohol solution vapors were detected by the instrument? You may not know the exact cause of the "Ambient Fail", but what did you do to clear the area of the ambient issue? Did you place a fan in the room to clear the air? Did you just wait a few minutes and repeat the test? Did you not wait and repeat the test?

I am requesting that you place some explanation as to why a portion of the breath test was repeated. These notes will be valuable later should this issue come up in court.

If you have any questions, please do not hesitate to call. Even if you work nights, I will try to assist you in any way I can.

Thank you in advance,

From: Deborah Reasoner [<mailto:DReasoner@clermontfl.org>]
Sent: Wednesday, June 19, 2019 6:58 PM
To: Graham, Thomas
Subject: RE: Data Review

I have no idea what that means.



Debbie Reasoner
Police Sergeant
Clermont Police Department
Bravo Squad
DRE/THI

3600 US 27., Clermont, FL 34711
Tel 352-536-8461 | Fax 352-394-1644
Cell 352-348-0137
dreasoner@clermontfl.org
www.cityofclermontfl.gov



From: Graham, Thomas <ThomasGraham@fdle.state.fl.us>
Sent: Monday, June 17, 2019 3:58 PM
To: Deborah Reasoner <DReasoner@clermontfl.org>
Subject: Data Review

Ms. Reasoner,

Please see the attachment regarding an Agency Inspection on instrument #80-001174.

If you have any questions, please feel free to call.

TJ Graham, D-ABFT-FA
Quality Assurance Manager
Alcohol Testing Program
Florida Department of Law Enforcement
Office: 850-617-1285
Cell: 850-445-7938

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

PLEASE NOTE: Under Florida law, e-mail addresses are public records. If you do not want your email address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.

Website: <https://www.clermontfl.gov/>

PLEASE NOTE: Under Florida law, e-mail addresses are public records. If you do not want your email address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.

Website: <https://www.clermontfl.gov/>

PLEASE NOTE: Under Florida law, e-mail addresses are public records. If you do not want your email address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

Agency CLERMONT P.D.	Date 04/30/2019	Type Agency Inspection	
Inspector REASONER, DEBORAH,	Time 00:26:18	Software 8100.27	
0.05 Lot # 201811C	0.08 Lot # 201808E	0.20 Lot # 201802G	0.08 Gas Lot # 02618080A2
Expiration 11/13/2020	Expiration 08/22/2020	Expiration 02/22/2020	Expiration 03/05/2020

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 05/03/2019 17:47

Printed: 17 June 2019

COBRA ©1997-2014 CMI, Inc.

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

6/21/19
JS

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 00:26

Date of Inspection: 04/30/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
/ 0.000	0.049	0.078	0.196	0.080
/ 0.000	0.048	0.078	0.200	0.079
/ 0.000	0.048	0.078	0.196	0.080

Number of Simulators Used: 4

Remarks:

00: Ambient Fail. - *Aired Room with Fan - Repeated test w/ issue. JL*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt [Signature] 141

DEBORAH REASONER

Signature and Printed Name

04/30/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

6/3/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 23:39

Date of Inspection: 05/31/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.048	0.077	0.196	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.199	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 141

DEBORAH REASONER
Signature and Printed Name

05/31/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

5/2/19
JD

Agency: CLERMONT P.D.
Time of Inspection: 00:26

Date of Inspection: 04/30/2019

Serial Number: 80-001174
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
/ 0.000	0.049	0.078	0.196	0.080
/ 0.000	0.048	0.078	0.200	0.079
/ 0.000	0.048	0.078	0.196	0.080

Number of Simulators Used: 4

Remarks:

00: Ambient Fail.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt [Signature] 14/ DEBORAH REASONER
Signature and Printed Name

04/30/2019
Date

Graham, Thomas

From: Graham, Thomas
Sent: Thursday, May 02, 2019 7:32 AM
To: Alcohol Testing Program
Subject: RE: Clermont April inspection

80-001174
5/2/19
JO

Sgt. Reasoner,

Have you been able to upload your instrument to FDLE? Also, please remember to make a note as to why you believe you received an ambient fail message during your inspection on the Form 40.

Thank you.

TJ

From: Alcohol Testing Program
Sent: Tuesday, April 30, 2019 8:15 AM
To: Graham, Thomas
Subject: FW: Clermont April inspection

From: Deborah Reasoner [<mailto:DReasoner@clermontfl.org>]
Sent: Tuesday, April 30, 2019 4:08 AM
To: Chambers, Rebecca; Bell, Danielle; Alcohol Testing Program; winterspringsbar@flhsmv.gov; wsievers@sao5.org
Subject: Clermont April inspection



Debbie Reasoner
Police Sergeant
Clermont Police Department
Bravo Squad
DRE/THI

3600 US 27., Clermont, FL 34711
Tel 352-536-8461 | Fax 352-394-1644
Cell 352-348-0137
dreasoner@clermontfl.org
www.cityofclermontfl.gov



Florida Department of Law Enforcement Alcohol Testing Program

3/29/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.
Time of Inspection: 03:00

Date of Inspection: 03/29/2019

Serial Number: 80-001174
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.049	0.078	0.194	0.079
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.079	0.197	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 141
Signature and Printed Name DEBORAH REASONER

03/29/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

2/27/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 02:42

Date of Inspection: 02/25/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.048	0.076	0.195	0.079
0.000	0.048	0.076	0.196	0.079
0.000	0.048	0.076	0.195	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DEBORAH REASONER
Signature and Printed Name

02/25/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

2/12/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.
Time of Inspection: 23:25

Date of Inspection: 01/27/2019

Serial Number: 80-001174
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 141

DEBORAH REASONER

Signature and Printed Name

01/27/2019
Date