

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SHERIFFS OFFICE

Serial Number: 80-001273

Time of Inspection: 23:00

Date of Inspection: 3-20-19

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		X
Diagnostic Check (Pre-Inspection): OK <span style="margin-left: 50px;">(MODEM TEST FAIL)</span>		X
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: 5

Remarks: "MODEM TEST FAIL" UPON PRE-INSPECTION DIAGNOSTIC CHECK  
(WOULD NOT ALLOW ME TO OBTAIN INSPECTION)

The above instrument complies (      ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. Deane

M. DEANE

Signature and Printed Name

3-20-19  
Date