

80-001287  
BK

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 03 2019

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: BAKER COUNTY SO      Date of Inspection: 04/28/2019      Serial Number: 80-001287  
Time of Inspection: 19:59      Software: 1.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g Dry Gas Test (g/210L) Lot#:18080A3 Exp: 05/2020
0.000	0.047	0.078	0.199	0.078
0.000	0.047	0.079	0.200	0.078
0.000	0.048	0.080	0.201	0.079

Number of Simulators Used: 5

Remarks:

RECEIVED

APR 24 2019

Records Dept.  
Baker County Sheriffs Office

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

04/28/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

80-061287  
BLK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO  
Time of Inspection: 08:28

Date of Inspection: 03/30/2019

Serial Number: 001287  
Software: 8100

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Stat (g/210L) Lot#:248 JA3 Exp: 10/2020
0.000	0.047	0.076	0.195	0.082
0.000	0.048	0.078	0.196	0.082
0.000	0.048	0.078	0.197	0.

Number of Simulators Used: 5

Remarks:  
A F / M A:

RECEIVED

APR 01 2019

Records Dept.  
Baker County Sheriffs Office

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

\_\_\_\_\_  
 Signature and Printed Name

03/30/2019  
 Date

# Florida Department of Law Enforcement Alcohol Testing Program

001287  
70-0007  
JBK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO  
Time of Inspection: 07:36

Date of Inspection: 03/30/2019

Serial Number: 001287  
Software: 8101

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas St est (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: 5

**Remarks:**

A F / M A: .UNKNOWN REASON .000 READING Non-compliance: WILL NOT MOVE TO NEXT TEST.

*Inspector Error* *CWS*

RECEIVED

APR 01 2019

Records Dept.  
Baker County Sheriffs Office

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*CWS*

COREY S WILSON

Signature and Printed Name

03/30/2019  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

80-001287  
BK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO  
Time of Inspection: 02:36

Date of Inspection: 02/27/2019

Serial Number: 80-001287  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24818080A3 Exp: 10/05/2020
0.000	0.047	0.078	0.196	0.082
0.000	0.048	0.078	0.198	0.082
0.000	0.048	0.078	0.198	0.082

Number of Simulators Used: 5

Remarks:

RECEIVED

MAR 01 2019

Records Dept.  
Baker County Sheriffs Office

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey

COREY S WILSON

Signature and Printed Name

02/27/2019  
Date