

Florida Department of Law Enforcement Alcohol Testing Program

5/2/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.
Time of Inspection: 07:04

Date of Inspection: 05/01/2019

Serial Number: 80-001345
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1031121 Exp: 10/12/2021
0.000 / 0.000	0.047	0.057 / 0.077	0.198	0.080
0.000 / 0.000	0.047	0.058 / 0.078	0.199	0.080
RFI / 0.000	0.047	0.060 / 0.078	0.198	0.080

Number of Simulators Used: 5

Remarks:

00: RFI Detect. 08: Control Outside Tolerance.

RADIO WAS CATCH IN ROOM DURING TEST
SIMULATOR WAS NOT TIGHT ENOUGH

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DET Robert T. Simken

ROBERT T SIMKEN

Signature and Printed Name

05/01/2019
Date

80-001345
5/2/19
JTB

Graham, Thomas

From: Alcohol Testing Program
Sent: Wednesday, May 01, 2019 7:40 AM
To: Graham, Thomas
Subject: FW: Eustis Police Department May 2019 Agency Inspection Intoxilyzer 8000
Attachments: May 2019 Eustis Police Department Intoxilyzer 8000 Agency Inspection.pdf

Importance: High

Follow Up Flag: FollowUp
Flag Status: Flagged

From: Simken, Robert [<mailto:SimkenR@ci.eustis.fl.us>]
Sent: Wednesday, May 01, 2019 7:18 AM
To: (WINTERSPRINGSBAR@FLHSMV.GOV); Kati Cosden; Alcohol Testing Program; Bell, Danielle; Bell, Danielle; Grieves, Matthew; Hunsicker, Megan; McSheehy, Shane; wsievers@sao5.org
Cc: Toler, Kenneth
Subject: Eustis Police Department May 2019 Agency Inspection Intoxilyzer 8000
Importance: High

Good morning,

During this inspection a radio was in the room and was cut off causing an RFI.

Also, the simulator was not tightened down enough causing an out of tolerance.

The Tests were repeated after making the required adjustments (removing the radio from the room, and tightening the simulator) and all were in compliance.

Thank you,

Robert

Detective Robert Simken
Senior Officer, E34
Accreditation Manager
Office of Professional Standards

Eustis Police Department
51 East Norton Avenue
Eustis, FL 32726

Desk: 352-483-5400
Fax: 352-357-1930
Email: Simkenr@ci.Eustis.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
FDLE
APR 01 2019

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: EUSTIS P.D.
Time of Inspection: 12:20

Date of Inspection: 04/01/2019

Serial Number: 80-001345
Software: 8100.27

4/8/19
JS

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1031121 Exp: 10/12/2021
0.000	0.049	0.078	0.199	0.081
0.000	0.048	0.079	0.199	0.081
0.000	0.048	0.078	0.199	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DET Robert T. Simken

ROBERT T SIMKEN

Signature and Printed Name

04/01/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

3/1/19
JS

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.
Time of Inspection: 09:28

Date of Inspection: 03/01/2019

Serial Number: 80-001345
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1031121 Exp: 10/12/2021
0.000	0.047	0.078	0.199	0.081
0.000	0.047	0.078	0.199	0.081
0.000	0.048	0.078	0.199	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Det. Robert T. Simken

ROBERT T SIMKEN

Signature and Printed Name

03/01/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

2/4/19
JE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 10:43

Date of Inspection: 02/04/2019

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1031121 Exp: 10/12/2021
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.079	0.199	0.081
0.000	0.049	0.079	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DET Robert T Simken

ROBERT T SIMKEN

Signature and Printed Name

02/04/2019
Date