

Florida Department of Law Enforcement Alcohol Testing Program

3/5/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNETH CITY PD
Time of Inspection: 18:23

Date of Inspection: 02/27/2019

Serial Number: 80-001649
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.049	0.072 / 0.073		
0.000	0.049	0.074 / 0.073		
0.000	0.049	0.074 / 0.074		

Number of Simulators Used: 3

Remarks:

08: MOISTURE IN CLR SIM TUB CHANGED TUBING

Non-compliance: .

OUT OF SERVICE

DO NOT USE

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

LT. MICHAEL J VIENO

Signature and Printed Name

02/27/2019

Date