Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD

Serial Number: 80-002463

Time of Inspection: 10:56

Date of Inspection: 05/23/2019

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	1
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
<u>-</u>	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:712031 Exp: 02/21/2020
0.000	0.043 / 0.050	0.079	0.195	0.077
0.000	0.044 / 0.050	0.080	0.195	0.077
0.000	0.046 / 0.050	0.080	0.196	0.077

Number	οf	Simulators	Used: 5	,

Remarks:

05: Control Outside Tolerance.

* Simulator hose became loose. G.M. 06/18/2019

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

6. JUB9 GENARO N MONTOYA Signature and Printed Name

05/23/2019 Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Hialeah PD Instrument Serial Number: 80-002463 AGENCY INSPECTION DATA REVIEW Time of Inspection: 10:56:32 Agency Inspector: Genaro Montoya Date of Inspection: 05/23/2019 Untimely/Not Received Erroneous Information Agency Inspection Discrepancy: Incomplete Procedural \boxtimes Other Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded. g/ 210L □Alcohol Reference Solution □Dry Gas Standard is □Incorrect □Expired. \Box Lot Number

Expiration Date for FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The REASON for repeating the following test(s); OR the Possible Cause and Corrective Action Taken on the following test(s) was not recorded: Alcohol Free Subject Test Mouth Alcohol Test □ Alcohol Free Test Interferent Detect Test П 0.08 g/210L Dry Gas Standard Test × 0.05 q/210L Test 0.08 g/210L Test П 0.20 g/210L Test FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use. The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC. Other: OTHER ELECTRONIC DATA REVIEW Comments: Login Records Date: Cylinder Change Records Control Test Records Date: **Diagnostic Check Records** Date: CORRECTIVE ACTION Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 7/12/2019 (Date). Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date). Upload the Agency Inspection(s). Remove the instrument from evidentiary use until otherwise directed by the Department. No action required Other: 6/6/2019 Signature of Alcohol Testing Program Staff Member Date

그 일 집에는 아이들은 물로를 놓아 내내고 있다. 얼마 없이 말라고요!	