

# Florida Department of Law Enforcement Alcohol Testing Program

5/3/19  
JO

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.  
Time of Inspection: 03:55

Date of Inspection: 05/02/2019

Serial Number: 80-003935  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.049	0.077	0.198	0.082
0.000	0.049	0.077	0.198	0.082
0.000	0.049	0.078	0.198	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*D Lyons*

DANIEL E LYONS

Signature and Printed Name

05/02/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

4/8/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.  
Time of Inspection: 19:36

Date of Inspection: 04/03/2019

Serial Number: 80-003935  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.049	0.078	0.197	0.082
0.000	0.049	0.079	0.198	0.082
0.000	0.049	0.079	0.199	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons

---

DANIEL E LYONS

Signature and Printed Name

04/03/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

3/19/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.  
Time of Inspection: 23:50

Date of Inspection: 03/18/2019

Serial Number: 80-003935  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.078	0.198	0.082
0.000	0.050	0.078	0.198	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*D. Lyons*

DANIEL E LYONS

Signature and Printed Name

03/18/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

2/15/19  
*[Signature]*

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.  
Time of Inspection: 19:13

Date of Inspection: 02/07/2019

Serial Number: 80-003935  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.047	0.000 / 0.077	0.197	0.081
0.000	0.048	0.000 / 0.079	0.197	0.080
0.000	0.048	0.000 / 0.079	0.197	0.081

Number of Simulators Used: 5

Remarks:  
08:Control Outside Tolerance/SIM NOT CONN/SIM CONN/RETESTED.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I, hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Handwritten Signature: D. Lyons]*

DANIEL E LYONS  
\_\_\_\_\_  
Signature and Printed Name

02/07/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

1/25/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.  
Time of Inspection: 22:05

Date of Inspection: 01/24/2019

Serial Number: 80-003935  
Software: 8100.27

B.W.

Check or Test	YES	NO
Calibration and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.041 / 0.047	0.077	0.194	0.081
0.000	0.049 / 0.048	0.078	0.196	0.081
0.000	0.049 / 0.047	0.078	0.197	0.081


Number of Simulators Used: 5

**Remarks:**

05: Control Outside Tolerance SIM LID RESET/RETESTED.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DANIEL E LYONS

---

Signature and Printed Name

01/24/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

11/7/19  
JW

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.  
Time of Inspection: 00:12

Date of Inspection: 01/07/2019

Serial Number: 80-003935  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.048	0.079	0.197	0.082
0.000	0.048	0.078	0.198	0.082
0.000	0.048	0.079	0.198	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D Lyons

\_\_\_\_\_  
Signature and Printed Name

DANIEL E LYONS

01/07/2019  
Date