

Florida Department of Law Enforcement Alcohol Testing Program

6/28/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 12:54

Date of Inspection: 06/13/2019

Serial Number: 80-006475
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 06/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:796576 Exp: 06/14/2020
0.000	0.047	0.076	0.197	0.078
0.000	0.047	0.077	0.199	0.079
0.000	0.047	0.077	0.199	0.079


Number of Simulators Used: 5

Remarks:

A. PRAZIER OBSERVED INSPECTION. THE WARNING EXPIRATION DATE WAS ENTERED AS THE .08 ALCOHOL REFERENCE SOLUTION, THE CORRECT DATE SHOULD HAVE BEEN 08/22/2020.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



YVETTE D WASHINGTON

Signature and Printed Name

06/13/2019
Date