



INSTRUMENT PROCESSING SHEET

Agency Sumter County SOS/N 80-000815Florida Department of
Law EnforcementDate In 11/14/2019DI Completion Date 12/20/2019☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Performed By <u>DP</u>	Quality Checks	Performed By <u>SP</u>	Flow Calibration	Performed By																																																										
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>198</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE DEC 30 2019 Alcohol Testing Program		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>SD1012</td><td>201905A 5/14/2021</td></tr><tr><td>0.080</td><td>DR1279</td><td>201905B 5/14/2021</td></tr><tr><td>0.200</td><td>SD1013</td><td>201904D 4/30/2021</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG916501 06/14/2021</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201905A 5/14/2021	0.080	DR1279	201905B 5/14/2021	0.200	SD1013	201904D 4/30/2021	0.080 DGS	N/A	AG916501 06/14/2021	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>SP</u> ☒ Lab Temp °C <u>21.5</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C ±.2 Serial #: <u>SD1012</u> ☒ 34°C ±.2 Serial #: <u>DR1279</u> ☒ 34°C ±.2 Serial #: <u>SD1013</u>																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	SD1012	201905A 5/14/2021																																																													
0.080	DR1279	201905B 5/14/2021																																																													
0.200	SD1013	201904D 4/30/2021																																																													
0.080 DGS	N/A	AG916501 06/14/2021																																																													
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1019</u> ID # <u>26932</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>G8144</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>G2403</td><td>19080</td><td>3/4/2021</td></tr><tr><td>0.100</td><td>G2879</td><td>19160</td><td>7/9/2021</td></tr><tr><td>0.200</td><td>G3709</td><td>19040</td><td>1/29/2021</td></tr><tr><td>0.300</td><td>G8149</td><td>19010</td><td>1/3/2021</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08819080A1</td><td>6/5/2021</td></tr></tbody></table> ☒ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>SD1012</td><td>201905A</td><td>5/14/2021</td></tr><tr><td>0.080</td><td>DR1279</td><td>201905B</td><td>5/14/2021</td></tr><tr><td>0.200</td><td>SD1013</td><td>201904D</td><td>4/30/2021</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG916501</td><td>6/14/2021</td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.000	G8144	N/A	N/A	0.040	G2403	19080	3/4/2021	0.100	G2879	19160	7/9/2021	0.200	G3709	19040	1/29/2021	0.300	G8149	19010	1/3/2021	0.080 DGS	N/A	08819080A1	6/5/2021	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1012	201905A	5/14/2021	0.080	DR1279	201905B	5/14/2021	0.200	SD1013	201904D	4/30/2021	0.080 DGS	N/A	AG916501	6/14/2021	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1027</u> Instrument <u>1026</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>G2408</td></tr><tr><td>Interferent</td><td>G2882</td></tr><tr><td>0.050</td><td>SD1012</td></tr><tr><td>0.080</td><td>DR1279</td></tr><tr><td>0.200</td><td>SD1013</td></tr></tbody></table> Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____		Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000	G8144	N/A	N/A																																																												
0.040	G2403	19080	3/4/2021																																																												
0.100	G2879	19160	7/9/2021																																																												
0.200	G3709	19040	1/29/2021																																																												
0.300	G8149	19010	1/3/2021																																																												
0.080 DGS	N/A	08819080A1	6/5/2021																																																												
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050	SD1012	201905A	5/14/2021																																																												
0.080	DR1279	201905B	5/14/2021																																																												
0.200	SD1013	201904D	4/30/2021																																																												
0.080 DGS	N/A	AG916501	6/14/2021																																																												
Simulator	Serial Number																																																														
0.000	G2408																																																														
Interferent	G2882																																																														
0.050	SD1012																																																														
0.080	DR1279																																																														
0.200	SD1013																																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>DP</u> 12/30/19 <u>Butt</u> 12/30/19 Tech Review / Date Admin Review / Date																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000815, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000815</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>SUMTER COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>12/20/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:38</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/20/2019

Date

Shayla Platt

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

OK 12/30/19

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO

Time of Inspection: 14:38

Date of Inspection: 12/20/2019

Serial Number: 80-000815

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test:			Diagnostic Check (Post-Inspection): OK	Yes	
Interferent Detect	Yes				

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.077	0.200	0.080
0.000	0.049	0.077	0.200	0.079
0.000	0.050	0.078	0.199	0.079
0.000	0.049	0.078	0.200	0.079

Standard Deviations	0.0005	0.0007	0.0008	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

Handwritten: PBM
TSK
12/30/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten Signature: Shayla Platt

SHAYLA D PLATT
Signature and Printed Name

12/20/2019
Date

Stability Checks #80-000815

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
12/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.048	10:13
Air Blank	0.000	10:13
Control Test	0.048	10:14
Air Blank	0.000	10:14
Control Test	0.048	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

POM

RK
12/30/19

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
12/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.077	10:23
Air Blank	0.000	10:23
Control Test	0.077	10:24
Air Blank	0.000	10:25
Control Test	0.078	10:25
Air Blank	0.000	10:26
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

SP

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
12/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:17
Control Test	0.199	10:18
Air Blank	0.000	10:19
Control Test	0.198	10:19
Air Blank	0.000	10:20
Control Test	0.198	10:20
Air Blank	0.000	10:21
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

SP

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
12/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:27
Control Test	0.081	10:27
Air Blank	0.000	10:28
Control Test	0.080	10:28
Air Blank	0.000	10:28
Control Test	0.080	10:29
Air Blank	0.000	10:29
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DAS

SP

Operator's Signature

SUMTER COUNTY 50

Intoxilyzer - Alcohol Analyzer

Model 8000

12/16/2019

SN 60-000815

10:30:58

Auto Calibration

Max Power Res Value = 22

Auto Range Res Value = 8

Sol Value = 0.000 g/210L ***

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12600, Sum Io = 13026

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 0.0230 (-0.0160)

Sample #2 = 0.0550 (0.0000)

Sample #3 = 0.0440 (0.0400)

Sample #4 = 0.0380 (0.0750)

Avg % Abs = 0.0457 (0.0383)

STD DEV = 0.0086 (0.0375)

REL STD DEV = 18.860 (97.899)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 0.0780 (0.0030)

Sample #2 = 0.1300 (-0.0180)

Sample #3 = 0.1290 (-0.0140)

Sample #4 = 0.1110 (0.0160)

Avg % Abs = 0.1233 (-0.0053)

STD DEV = 0.0107 (0.0186)

REL STD DEV = 8.670 (348.434)

Sol Value = 0.040 g/210L ***

Fit value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12586, Sum Io = 13022

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 0.7800 (-0.0240)

Sample #2 = 0.7890 (0.0290)

Sample #3 = 0.8120 (0.0190)

Sample #4 = 0.8070 (0.0430)

Avg % Abs = 0.8027 (0.0303)

STD DEV = 0.0121 (0.0121)

REL STD DEV = 1.507 (39.743)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 1.4810 (-0.0070)

Sample #2 = 1.4630 (0.0210)

Sample #3 = 1.4980 (0.0040)

Sample #4 = 1.5060 (0.0080)

Avg % Abs = 1.4890 (0.0110)

STD DEV = 0.0229 (0.0089)

REL STD DEV = 1.536 (80.802)

Sol Value = 0.100 g/210L ***

Fit value = 0.4762 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12574, Sum Io = 13019

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 1.8190 (-0.0220)

Sample #2 = 1.8300 (-0.0170)

Sample #3 = 1.8270 (-0.0300)

Sample #4 = 1.8570 (-0.0230)

Avg % Abs = 1.8380 (-0.0233)

STD DEV = 0.0165 (0.0065)

REL STD DEV = 0.899 (27.885)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 3.4030 (0.0090)

Sample #2 = 3.4220 (0.0060)

Sample #3 = 3.4250 (-0.0150)

Sample #4 = 3.4040 (0.0140)

Avg % Abs = 3.4170 (0.0017)

STD DEV = 0.0114 (0.0150)

REL STD DEV = 0.332 (898.665)

Sol Value = 0.200 g/210L ***

Fit value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12573, Sum Io = 13018

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 3.4990 (-0.0240)

Sample #2 = 3.5380 (0.0250)

Sample #3 = 3.5520 (0.0210)

Sample #4 = 3.5390 (0.0080)

Avg % Abs = 3.5430 (0.0180)

STD DEV = 0.0078 (0.0089)

REL STD DEV = 0.220 (49.379)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 6.4920 (0.0060)

Sample #2 = 6.5630 (0.0290)

Sample #3 = 6.5480 (0.0350)

Sample #4 = 6.6050 (0.0130)

Avg % Abs = 6.5720 (0.0257)

STD DEV = 0.0295 (0.0114)

REL STD DEV = 0.450 (44.308)

Sol Value = 0.300 g/210L ***

Fit value = 1.4286 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12571, Sum Io = 13014

***** CHANNEL 1 *****

Sample % Abs (% Abs Ref)

Sample #1 = 5.1490 (0.0050)

Sample #2 = 5.1530 (0.0330)

Sample #3 = 5.1310 (0.0410)

Sample #4 = 5.1480 (0.0380)

Avg % Abs = 5.1440 (0.0373)

STD DEV = 0.0115 (0.0040)

REL STD DEV = 0.224 (10.825)

***** CHANNEL 2 *****

Sample % Abs (% Abs Ref)

Sample #1 = 9.4960 (0.0060)

Sample #2 = 9.4850 (0.0350)

Sample #3 = 9.4790 (0.0330)

Sample #4 = 9.5100 (0.0270)

Avg % Abs = 9.4913 (0.0317)

STD DEV = 0.0164 (0.0042)

REL STD DEV = 0.173 (13.147)

***** AUTO CAL DATA *****

***** CHANNEL 1 *****

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.046

Std Dev = 0.01 Rel Std Dev = 18.88

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.803

Std Dev = 0.01 Rel Std Dev = 1.51

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.838

Std Dev = 0.02 Rel Std Dev = 0.90

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.543

Std Dev = 0.01 Rel Std Dev = 0.22

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.144

Std Dev = 0.01 Rel Std Dev = 0.22

Zero Order Coef = -137.25

First Order Coef = 2562.30

Second Order Coef = 46.94

Standard Deviation = 29.276747

***** CHANNEL 2 *****

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.123

Std Dev = 0.01 Rel Std Dev = 8.67

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.489

Std Dev = 0.02 Rel Std Dev = 1.54

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.417

Std Dev = 0.01 Rel Std Dev = 0.33

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.572

Std Dev = 0.03 Rel Std Dev = 0.45

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 9.491

Std Dev = 0.02 Rel Std Dev = 0.17

Zero Order Coef = -170.69

First Order Coef = 1383.14

Second Order Coef = 14.64

Standard Deviation = 24.486137

Solution Stats Quadratic Fit Chan 1

Act Fit Residual

g/210L g/210L g/210L

0.000 -0.000 0.0004

0.040 0.041 -0.0009

0.100 0.099 0.0007

0.200 0.200 -0.0001

0.300 0.300 0.0000

Solution Stats Quadratic Fit Chan 2

Act Fit Residual

g/210L g/210L g/210L

0.000 0.000 -0.0000

0.040 0.040 -0.0003

0.100 0.099 0.0007

0.200 0.201 -0.0006

0.300 0.300 0.0002

Sol Value = 0.080 g/210L ***

Fit value = 0.3810 mg/l %%%

Samples Taken = 4, Discarded = 1

***** CHANNEL 1 *****

Sample #1 = 3303.00

Sample #2 = 3234.00

Sample #3 = 3354.00

Sample #4 = 3277.00

Average Result = 3288.3333

STD DEV = 60.7975

REL STD DEV = 1.849

***** CHANNEL 2 *****

Sample #1 = 3363.00

Sample #2 = 3317.00

Sample #3 = 3377.00

Sample #4 = 3402.00

Average Result = 3365.3333

STD DEV = 43.6845

REL STD DEV = 1.298

***** CHANNEL 1 *****

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1019

3 um H2O Adjust (mg/l*10,000) = 521

9 um H2O Adjust (mg/l*10,000) = 444

***** AUTO CAL PASS *****

CAL ADJUSTMENT
#80-000815 SP

PDM BK 12/30/19

Post Cal Adjust Stability Checks # 80-000815

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:49
Control Test	0.048	10:50
Air Blank	0.000	10:50
Control Test	0.049	10:51
Air Blank	0.000	10:51
Control Test	0.049	10:52
Air Blank	0.000	10:53
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

SP

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:54
Control Test	0.078	10:54
Air Blank	0.000	10:55
Control Test	0.077	10:55
Air Blank	0.000	10:56
Control Test	0.078	10:57
Air Blank	0.000	10:57
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

SP

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:01
Control Test	0.200	11:02
Air Blank	0.000	11:03
Control Test	0.198	11:03
Air Blank	0.000	11:04
Control Test	0.199	11:05
Air Blank	0.000	11:05
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

SP

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:45
Control Test	0.080	10:45
Air Blank	0.000	10:46
Control Test	0.080	10:46
Air Blank	0.000	10:47
Control Test	0.080	10:47
Air Blank	0.000	10:48
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

SP

Operator's Signature

SPK 12/30/19

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 10:06

Date of Inspection: 12/10/2019

Serial Number: 80-000815
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

*N/A compliance
not determined*

*13K
12/30/19*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

12/10/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 10:30

Date of Inspection: 12/16/2019

Serial Number: 80-000815
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

*N/A Compliance
Not Determined.*

QJGM

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

12/16/2019
Date

*RSK
12/30/19*



INSTRUMENT PROCESSING SHEET

Florida Department of
Law Enforcement

Agency FDLE Sumter Co. S.O. S/N 80-000815
Date In 11/14/2019 DI Completion Date 11/18/2019 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>AA</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☐ Case ☐ Handle ☐ Keyboard ☐ Dry Gas Shelf ☐ Feet ☐ Breath Tube ☐ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Training</u>		Quality Checks Performed By <u>AA</u> ☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☒ R-Value <u>186</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.178</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks		Flow Calibration Performed By <u>AA</u> Flow Column # <u>ATP-105</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>189</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.170</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547)																																							
Final Release Date		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1018</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>G2078</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG916501 06/14/2021</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD1018	201905A 05/14/2021	0.080	SD3962	201905B 05/14/2021	0.200	G2078	201904D 04/30/2021	0.080 DGS	N/A	AG916501 06/14/2021	Maintenance Performed By <u>AA</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																								
Simulator	Serial #	Lot #/Exp																																									
0.050	SD1018	201905A 05/14/2021																																									
0.080	SD3962	201905B 05/14/2021																																									
0.200	G2078	201904D 04/30/2021																																									
0.080 DGS	N/A	AG916501 06/14/2021																																									
Calibration Adjustment Performed By <u>AA</u> Barometric Pressure Gauge <u>1010</u> ID # <u>30793</u>		Department Inspection Performed By <u>AA</u> Barometric Pressure ID# <u>28421</u> Gauge <u>10221021</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>19080</td> <td>03/04/2021</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>19160</td> <td>07/09/2021</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>19040</td> <td>01/29/2021</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>19010</td> <td>01/03/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28311080A</td> <td>11/01/2013</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	MP5086	N/A	N/A	0.040	MP5082	19080	03/04/2021	0.100	MP5083	19160	07/09/2021	0.200	MP5084	19040	01/29/2021	0.300	MP5085	19010	01/03/2021	0.080 DGS	N/A	28311080A	11/01/2013	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G7444</td> </tr> <tr> <td>Interferent</td> <td>G6621</td> </tr> <tr> <td>0.050</td> <td>SD1018</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> </tr> <tr> <td>0.200</td> <td>G2078</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	G7444	Interferent	G6621	0.050	SD1018	0.080	SD3962	0.200	G2078
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	MP5086	N/A	N/A																																								
0.040	MP5082	19080	03/04/2021																																								
0.100	MP5083	19160	07/09/2021																																								
0.200	MP5084	19040	01/29/2021																																								
0.300	MP5085	19010	01/03/2021																																								
0.080 DGS	N/A	28311080A	11/01/2013																																								
Simulator	Serial Number																																										
0.000	G7444																																										
Interferent	G6621																																										
0.050	SD1018																																										
0.080	SD3962																																										
0.200	G2078																																										
☒ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☒ Stability Checks ☐ Calibration Certificate ☒ Calibration Adjustment ☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____																																									
Notes/Suggested Service: <u>Done unintentionally by Dept.</u> <u>Inspector trainee. Will be repeated</u> <u>Repeat in correct order by certified Dept</u> <u>Inspector. BK</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																									
<u>11/19/19</u> <u>BK</u> <u>11/21/19</u> Tech Review / Date Admin Review / Date																																											

not same BK

Stability Checks

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:26
Control Test	0.048	12:27
Air Blank	0.000	12:28
Control Test	0.049	12:28
Air Blank	0.000	12:29
Control Test	0.051	12:30
Air Blank	0.000	12:30
Control Test Stats		
Average	0.0493	
Std Dev	0.0015	
Rel Std Dev(%)	3.0963	


Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:36
Control Test	0.203	12:37
Air Blank	0.000	12:38
Control Test	0.206	12:38
Air Blank	0.000	12:39
Control Test	0.208	12:40
Air Blank	0.000	12:40
Control Test Stats		
Average	0.2057	
Std Dev	0.0025	
Rel Std Dev(%)	1.2236	


Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.080	12:32
Air Blank	0.000	12:33
Control Test	0.081	12:34
Air Blank	0.000	12:34
Control Test	0.082	12:35
Air Blank	0.000	12:35
Control Test Stats		
Average	0.0810	
Std Dev	0.0010	
Rel Std Dev(%)	1.2346	

wet


Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:42
Control Test	0.081	12:42
Air Blank	0.000	12:43
Control Test	0.081	12:43
Air Blank	0.000	12:44
Control Test	0.081	12:44
Air Blank	0.000	12:45
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry


Operator's Signature


11/21/19

JA

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/18/2019 09:08:28

Auto Calibration
Max Power Res Value = 21
Auto Range Res Value = 8

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12585, Sum Io = 13023

CHANCELL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.0790	(-0.0560)
Sample #2 =	0.1010	(0.0180)
Sample #3 =	0.0760	(0.0350)
Sample #4 =	0.0710	(0.0890)
Avg % Abs =	0.0827	(0.0473)
STD DEV =	0.0161	(0.0371)
REL STD DEV =	19.443	(78.321)

CHANCELL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1590	(-0.0210)
Sample #2 =	0.1520	(0.0030)
Sample #3 =	0.1330	(0.0040)
Sample #4 =	0.1250	(0.0280)
Avg % Abs =	0.1367	(0.0117)
STD DEV =	0.0139	(0.0142)
REL STD DEV =	10.148	(121.319)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12567, Sum Io = 13017

CHANCELL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.7810	(-0.0180)
Sample #2 =	0.7910	(0.0200)
Sample #3 =	0.7870	(0.0400)
Sample #4 =	0.7660	(0.0370)
Avg % Abs =	0.7813	(0.0323)
STD DEV =	0.0134	(0.0108)
REL STD DEV =	1.719	(33.358)

CHANCELL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.4880	(0.0100)
Sample #2 =	1.4570	(0.0410)
Sample #3 =	1.4800	(0.0330)
Sample #4 =	1.4180	(0.0310)
Avg % Abs =	1.4517	(0.0350)
STD DEV =	0.0313	(0.0053)
REL STD DEV =	2.159	(15.119)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12547, Sum Io = 13005

CHANCELL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.5600	(-0.0140)
Sample #2 =	3.5700	(0.0080)
Sample #3 =	3.5360	(0.0020)
Sample #4 =	3.5600	(0.0100)
Avg % Abs =	3.5553	(0.0067)
STD DEV =	0.0175	(0.0042)
REL STD DEV =	0.491	(62.450)

CHANCELL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	6.5930	(-0.0110)
Sample #2 =	6.6000	(0.0050)
Sample #3 =	6.6140	(-0.0070)
Sample #4 =	6.5920	(0.0130)
Avg % Abs =	6.6020	(0.0037)
STD DEV =	0.0111	(0.0101)
REL STD DEV =	0.169	(274.539)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12557, Sum Io = 13008

CHANCELL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.8230	(-0.0160)
Sample #2 =	1.8490	(0.0060)
Sample #3 =	1.7960	(0.0270)
Sample #4 =	1.8170	(0.0620)
Avg % Abs =	1.8207	(0.0317)
STD DEV =	0.0267	(0.0283)
REL STD DEV =	1.466	(89.337)

CHANCELL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.4250	(-0.0350)
Sample #2 =	3.4070	(-0.0230)
Sample #3 =	3.4070	(-0.0130)
Sample #4 =	3.3890	(0.0100)
Avg % Abs =	3.4010	(-0.0087)
STD DEV =	0.0104	(0.0169)
REL STD DEV =	0.306	(195.247)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12543, Sum Io = 13001

CHANCELL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	5.2410	(-0.0220)
Sample #2 =	5.2140	(0.0200)
Sample #3 =	5.1970	(0.0090)
Sample #4 =	5.2030	(0.0190)
Avg % Abs =	5.2047	(0.0160)
STD DEV =	0.0086	(0.0061)
REL STD DEV =	0.166	(38.017)

CHANCELL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	9.5310	(-0.0150)
Sample #2 =	9.5480	(0.0200)
Sample #3 =	9.5630	(0.0050)
Sample #4 =	9.5480	(0.0020)
Avg % Abs =	9.5530	(0.0090)
STD DEV =	0.0087	(0.0096)
REL STD DEV =	0.091	(107.152)

QDM

13K
11/21/19

Optical Bench Cal. Adj.

AD

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.083
 Std Dev = 0.02 Rel Std Dev = 19.44
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.781
 Std Dev = 0.01 Rel Std Dev = 1.72
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.821
 Std Dev = 0.03 Rel Std Dev = 1.47
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.555
 Std Dev = 0.02 Rel Std Dev = 0.49
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.205
 Std Dev = 0.01 Rel Std Dev = 0.17
 Zero Order Coef = -204.48
 First Order Coef = 2671.94
 Second Order Coef = 21.05
 Standard Deviation = 27.148037

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.137
 Std Dev = 0.01 Rel Std Dev = 10.15
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.452
 Std Dev = 0.03 Rel Std Dev = 2.16
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.401
 Std Dev = 0.01 Rel Std Dev = 0.31
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.602
 Std Dev = 0.01 Rel Std Dev = 0.17
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.553
 Std Dev = 0.01 Rel Std Dev = 0.09
 Zero Order Coef = -165.96
 First Order Coef = 1396.70
 Second Order Coef = 11.94
 Standard Deviation = 37.307869

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0002
0.100	0.099	0.0007
0.200	0.201	-0.0008
0.300	0.300	0.0003

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0005
0.040	0.040	0.0004
0.100	0.099	0.0008
0.200	0.201	-0.0011
0.300	0.300	0.0004

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3288.00
 Sample #2 = 3379.00
 Sample #3 = 3225.00
 Sample #4 = 3361.00
 Average Result = 3321.6667
 STD DEV = 84.1982
 REL STD DEV = 2.535

***** CHANNEL 2

Sample #1 = 3339.00
 Sample #2 = 3308.00
 Sample #3 = 3309.00
 Sample #4 = 3350.00
 Average Result = 3322.3333
 STD DEV = 23.9653
 REL STD DEV = 0.721

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1010
 3 um H2O Adjust (mg/l*10,000) = 488
 9 um H2O Adjust (mg/l*10,000) = 487

**** AUTO CAL PASS

ADOM

TSK
 11/25/19

Post Opt. Bench Cal. Adj. Stability Checks

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:49
Control Test	0.048	09:50
Air Blank	0.000	09:51
Control Test	0.050	09:51
Air Blank	0.000	09:52
Control Test	0.050	09:52
Air Blank	0.000	09:53
Control Test Stats		
Average	0.0493	
Std Dev	0.0012	
Rel Std Dev(%)	2.3406	

SS

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.079	09:55
Air Blank	0.000	09:55
Control Test	0.079	09:56
Air Blank	0.000	09:57
Control Test	0.079	09:57
Air Blank	0.000	09:58
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet

SS

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.200	10:00
Air Blank	0.000	10:00
Control Test	0.200	10:01
Air Blank	0.000	10:02
Control Test	0.199	10:02
Air Blank	0.000	10:03
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

SS

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.081	10:11
Air Blank	0.000	10:12
Control Test	0.081	10:12
Air Blank	0.000	10:12
Control Test	0.081	10:13
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry

SS

Operator's Signature

pgm
11/21/15

Flow Rate Sensor Cali. Adj.

DA

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
11/14/2019
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 7.070

2: Rate (Liters/min) = 15

SQRT(Diff)) = 11.703

3: Rate (Liters/min) = 30

SQRT(Diff)) = 21.352

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 672

Rounded Intercept = -634881

Correlation = 0.99638

Agm

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 14:15

Date of Inspection: 11/14/2019

Serial Number: 80-000815
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.049	0.081	0.209	0.079
0.000	0.050	0.082	0.209	0.079
0.000	0.051	0.082	0.208	0.079
0.000	0.051	0.081	0.207	0.079
0.000	0.052	0.082	0.206	0.079
0.000	0.051	0.082	0.206	0.080
0.000	0.051	0.083	0.204	0.079
0.000	0.051	0.082	0.204	0.079
0.000	0.051	0.082	0.204	0.079
0.000	0.051	0.082	0.204	0.079
0.000	0.051	0.083	0.203	0.080

Standard Deviations	0.0007	0.0006	0.0022	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0009 Number of Simulators Used: 5

Remarks:

JDM
BSK
11/21/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

11/14/2019
Date