



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade Police DepartmentS/N 80-000882

Florida Department of Law Enforcement

Date In 02/04/2019DI Completion Date 02/05/2019 Ship P/U H/D CMI EE

Intake Performed By <u>Dee</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>Dee</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>171</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By <u>Dee</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Form load/Changed pass</u> Temperature Checks Performed By <u>Dee</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.97C</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD3969</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	SD3967	201707D 07/25/2019																																																												
0.080	SD3968	201707E 07/25/2019																																																												
0.200	SD3969	201707C 07/24/2019																																																												
0.080 DGS	N/A	AG805701 02/26/2020																																																												
Final Release Date FDLE FEB 14 2019 Alcohol Testing Program																																																														
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>Dee</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1016</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	SD3965																																																													
Interferent	SD3966																																																													
0.050	SD3967																																																													
0.080	SD3968																																																													
0.200	SD3969																																																													
Notes/Suggested Service: <u>E-mailed</u> <input checked="" type="checkbox"/> APPROVED <u>Approved date 2/5/2019</u> _____ DERR Digitally signed by DERR Date: 2020.05.20 11:46:58 -04'00'	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>Dee 2/13/19</u> <u>J. Deah 2/14/19</u> Tech Review / Date Admin Review / Date																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 08:14

Date of Inspection: 02/05/2019

Serial Number: 80-000882
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.050	0.081	0.200	0.079
0.000	0.049	0.081	0.201	0.079
0.000	0.050	0.082	0.202	0.080
0.000	0.050	0.081	0.201	0.080
0.000	0.051	0.081	0.202	0.080
0.000	0.050	0.082	0.201	0.080
0.000	0.050	0.081	0.201	0.080
0.000	0.051	0.082	0.201	0.080
0.000	0.051	0.082	0.201	0.080
0.000	0.051	0.082	0.201	0.080

Standard Deviations	0.0006	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

PDM

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

02/05/2019
 Date

2/14/19
DR

U180

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000882	Miami-Dade Police Department	02/05/2019	<i>Deer</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.06g 0.047 to 0.053 <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.07C 0.077 to 0.083 <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.08C 0.194 to 0.206 <input checked="" type="checkbox"/>	Lot AG805701 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
MVM-040E F0 Intoxilyzer - Alcotest Analyzer Model: 8000 02/05/2019 Software: 8100.27	MVM-040E F0 Intoxilyzer - Alcotest Analyzer Model: 8000 02/05/2019 Software: 8100.27	MVM-040E F0 Intoxilyzer - Alcotest Analyzer Model: 8000 02/05/2019 Software: 8100.27	MVM-040E F0 Intoxilyzer - Alcotest Analyzer Model: 8000 02/05/2019 Software: 8100.27																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.009</td><td>06:10</td></tr> <tr><td>Control Test</td><td>0.045</td><td>06:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:11</td></tr> <tr><td>Control Test</td><td>0.150</td><td>06:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:12</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.009	06:10	Control Test	0.045	06:11	Air Blank	0.000	06:11	Control Test	0.150	06:12	Air Blank	0.000	06:12	Control Test	0.049	06:13	Air Blank	0.000	06:14	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel. Std Dev(%)	1.1703		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:16</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:18</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7167</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:15	Control Test	0.080	06:16	Air Blank	0.000	06:16	Control Test	0.080	06:17	Air Blank	0.000	06:18	Control Test	0.081	06:18	Air Blank	0.000	06:19	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel. Std Dev(%)	0.7167		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:20</td></tr> <tr><td>Control Test</td><td>0.199</td><td>06:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:21</td></tr> <tr><td>Control Test</td><td>0.198</td><td>06:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:22</td></tr> <tr><td>Control Test</td><td>0.199</td><td>06:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2986</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:20	Control Test	0.199	06:21	Air Blank	0.000	06:21	Control Test	0.198	06:22	Air Blank	0.000	06:22	Control Test	0.199	06:23	Air Blank	0.000	06:24	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel. Std Dev(%)	0.2986		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:26</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:25	Control Test	0.079	06:25	Air Blank	0.000	06:26	Control Test	0.080	06:26	Air Blank	0.000	06:27	Control Test	0.079	06:27	Air Blank	0.000	06:27	Control Test Stats			Average	0.0793		Std Dev	0.0005		Rel. Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.009	06:10																																																																																																																																																	
Control Test	0.045	06:11																																																																																																																																																	
Air Blank	0.000	06:11																																																																																																																																																	
Control Test	0.150	06:12																																																																																																																																																	
Air Blank	0.000	06:12																																																																																																																																																	
Control Test	0.049	06:13																																																																																																																																																	
Air Blank	0.000	06:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:15																																																																																																																																																	
Control Test	0.080	06:16																																																																																																																																																	
Air Blank	0.000	06:16																																																																																																																																																	
Control Test	0.080	06:17																																																																																																																																																	
Air Blank	0.000	06:18																																																																																																																																																	
Control Test	0.081	06:18																																																																																																																																																	
Air Blank	0.000	06:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7167																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:20																																																																																																																																																	
Control Test	0.199	06:21																																																																																																																																																	
Air Blank	0.000	06:21																																																																																																																																																	
Control Test	0.198	06:22																																																																																																																																																	
Air Blank	0.000	06:22																																																																																																																																																	
Control Test	0.199	06:23																																																																																																																																																	
Air Blank	0.000	06:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2986																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:25																																																																																																																																																	
Control Test	0.079	06:25																																																																																																																																																	
Air Blank	0.000	06:26																																																																																																																																																	
Control Test	0.080	06:26																																																																																																																																																	
Air Blank	0.000	06:27																																																																																																																																																	
Control Test	0.079	06:27																																																																																																																																																	
Air Blank	0.000	06:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel. Std Dev(%)	0.7277																																																																																																																																																		
<i>Deer</i> Operator's Signature	<i>Deer</i> Operator's Signature	<i>Deer</i> Operator's Signature	<i>Deer</i> Operator's Signature																																																																																																																																																

2/14/19



Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000882, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000882</u>	UNCERTAINTY* ±
Owning Agency:	<u>MIAMI-DADE PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/05/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>08:14</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/05/2019

Date

David Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

2/14/19
EC

Calibrated



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade Police Department

S/N 80-000882

Florida Department of Law Enforcement

Date In 02/04/2019

DI Completion Date 02/05/2019

Ship P/U H/D CMI EE

Intake Performed By <u>Dee</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>Dee</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>171</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	SD3967	201707D 07/25/2019																																																													
0.080	SD3968	201707E 07/25/2019																																																													
0.200	SD3969	201707C 07/24/2019																																																													
0.080 DGS	N/A	AG805701 02/26/2020																																																													
Final Release Date <p style="text-align: center;">FDLE FEB 14 2019 Alcohol Testing Program</p>		Maintenance Performed By <u>Dee</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Form load/Changed pass</u>																																																													
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>Dee</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1016</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	SD3965																																																														
Interferent	SD3966																																																														
0.050	SD3967																																																														
0.080	SD3968																																																														
0.200	SD3969																																																														
Notes/Suggested Service: <u>E-mailed</u> <input checked="" type="checkbox"/> APPROVED _____ _____ _____ _____		Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																													
		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													
		<u>Dee</u> 2/13/19 <u>J. Deah</u> 2/14/19 Tech Review / Date Admin Review / Date																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 08:14

Date of Inspection: 02/05/2019

Serial Number: 80-000882
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.050	0.081	0.200	0.079
0.000	0.049	0.081	0.201	0.079
0.000	0.050	0.082	0.202	0.080
0.000	0.050	0.081	0.201	0.080
0.000	0.051	0.081	0.202	0.080
0.000	0.050	0.082	0.201	0.080
0.000	0.050	0.081	0.201	0.080
0.000	0.051	0.082	0.201	0.080
0.000	0.051	0.082	0.201	0.080
0.000	0.051	0.082	0.201	0.080

Standard Deviations	0.0006	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

PDM

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

02/05/2019
 Date

2/14/19
DR

U180

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000882	Miami-Dade Police Department	02/05/2019	<i>Deer</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.06g 0.047 to 0.053 <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.07C 0.077 to 0.083 <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.08C 0.194 to 0.206 <input checked="" type="checkbox"/>	Lot AG805701 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
MIAMI-DADE PD Intoxilyzer - Alcotest Analyzer Model: 8000 02/05/2019 Software: 8101.27	MIAMI-DADE PD Intoxilyzer - Alcotest Analyzer Model: 8000 02/05/2019 Software: 8101.27	MIAMI-DADE PD Intoxilyzer - Alcotest Analyzer Model: 8000 02/05/2019 Software: 8101.27	MIAMI-DADE PD Intoxilyzer - Alcotest Analyzer Model: 8000 02/05/2019 Software: 8101.27																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.009</td><td>06:10</td></tr> <tr><td>Control Test</td><td>0.045</td><td>06:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:11</td></tr> <tr><td>Control Test</td><td>0.050</td><td>06:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:12</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.009	06:10	Control Test	0.045	06:11	Air Blank	0.000	06:11	Control Test	0.050	06:12	Air Blank	0.000	06:12	Control Test	0.049	06:13	Air Blank	0.000	06:14	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel. Std Dev(%)	1.1703		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:16</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:18</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7167</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:15	Control Test	0.080	06:16	Air Blank	0.000	06:16	Control Test	0.080	06:17	Air Blank	0.000	06:18	Control Test	0.081	06:18	Air Blank	0.000	06:19	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel. Std Dev(%)	0.7167		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:20</td></tr> <tr><td>Control Test</td><td>0.199</td><td>06:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:21</td></tr> <tr><td>Control Test</td><td>0.198</td><td>06:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:22</td></tr> <tr><td>Control Test</td><td>0.199</td><td>06:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2986</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:20	Control Test	0.199	06:21	Air Blank	0.000	06:21	Control Test	0.198	06:22	Air Blank	0.000	06:22	Control Test	0.199	06:23	Air Blank	0.000	06:24	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel. Std Dev(%)	0.2986		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:26</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:25	Control Test	0.079	06:25	Air Blank	0.000	06:26	Control Test	0.080	06:26	Air Blank	0.000	06:27	Control Test	0.079	06:27	Air Blank	0.000	06:27	Control Test Stats			Average	0.0793		Std Dev	0.0005		Rel. Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.009	06:10																																																																																																																																																	
Control Test	0.045	06:11																																																																																																																																																	
Air Blank	0.000	06:11																																																																																																																																																	
Control Test	0.050	06:12																																																																																																																																																	
Air Blank	0.000	06:12																																																																																																																																																	
Control Test	0.049	06:13																																																																																																																																																	
Air Blank	0.000	06:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:15																																																																																																																																																	
Control Test	0.080	06:16																																																																																																																																																	
Air Blank	0.000	06:16																																																																																																																																																	
Control Test	0.080	06:17																																																																																																																																																	
Air Blank	0.000	06:18																																																																																																																																																	
Control Test	0.081	06:18																																																																																																																																																	
Air Blank	0.000	06:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7167																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:20																																																																																																																																																	
Control Test	0.199	06:21																																																																																																																																																	
Air Blank	0.000	06:21																																																																																																																																																	
Control Test	0.198	06:22																																																																																																																																																	
Air Blank	0.000	06:22																																																																																																																																																	
Control Test	0.199	06:23																																																																																																																																																	
Air Blank	0.000	06:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2986																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:25																																																																																																																																																	
Control Test	0.079	06:25																																																																																																																																																	
Air Blank	0.000	06:26																																																																																																																																																	
Control Test	0.080	06:26																																																																																																																																																	
Air Blank	0.000	06:27																																																																																																																																																	
Control Test	0.079	06:27																																																																																																																																																	
Air Blank	0.000	06:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel. Std Dev(%)	0.7277																																																																																																																																																		
<i>Deer</i> Operator's Signature	<i>Deer</i> Operator's Signature	<i>Deer</i> Operator's Signature	<i>Deer</i> Operator's Signature																																																																																																																																																

2/14/19
Deer



Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000882, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000882</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI-DADE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/05/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:14</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/05/2019

Date

David Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

EC
2/4/19

Calibrated