



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade Police DepartmentS/N 80-000883Florida Department of
Law EnforcementDate In 02/18/2019 DI Completion Date 02/19/2019 Ship P/U H/D CMI EE

Intake Performed By <u>DELL</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>100</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.187</u> (.156 - .190) 53 mm <u>.257</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																
Final Release Date <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FDLE</div> <div style="text-align: center; font-weight: bold; font-size: 1.2em;">MAR 04 2019</div> <div style="text-align: center;">Alcohol Testing Program</div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Maintenance Performed By <u>DELL</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Changed pass/ Batt box</u>																	
Simulator	Serial #	Lot #/Exp																																
0.050	SD3967	201707D 07/25/2019																																
0.080	SD3968	201707E 07/25/2019																																
0.200	SD3969	201707C 07/24/2019																																
0.080 DGS	N/A	AG805701 02/26/2020																																
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Department Inspection Performed By <u>DELL</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1019</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u>				
Simulator	Serial Number	Lot Number	Expiration																															
0.000		N/A	N/A																															
0.040																																		
0.100																																		
0.200																																		
0.300																																		
0.080 DGS	N/A																																	
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3963</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD3965	Interferent	SD3963	0.050	SD3967	0.080	SD3968	0.200	SD3969	Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____
Simulator	Serial Number	Lot Number	Expiration																															
0.050																																		
0.080																																		
0.200																																		
0.080 DGS	N/A																																	
Simulator	Serial Number																																	
0.000	SD3965																																	
Interferent	SD3963																																	
0.050	SD3967																																	
0.080	SD3968																																	
0.200	SD3969																																	
Notes/Suggested Service: <u>E-mailed</u> <input checked="" type="checkbox"/> APPROVED <u>Recommended instrument be sent to repair facility for a flow sensor replacement</u> _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	<table style="width:100%;"> <tr> <td style="text-align: center;"><u>Dgm 3/4/19</u></td> <td style="text-align: center;"><u>JJ Deh 3/4/19</u></td> </tr> <tr> <td style="text-align: center;">Tech Review / Date</td> <td style="text-align: center;">Admin Review / Date</td> </tr> </table>	<u>Dgm 3/4/19</u>	<u>JJ Deh 3/4/19</u>	Tech Review / Date	Admin Review / Date																												
<u>Dgm 3/4/19</u>	<u>JJ Deh 3/4/19</u>																																	
Tech Review / Date	Admin Review / Date																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 08:13

Date of Inspection: 02/19/2019

Serial Number: 80-000883
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.082	0.199	0.080
0.000	0.050	0.082	0.199	0.079
0.000	0.051	0.082	0.200	0.079
0.000	0.051	0.082	0.200	0.079
0.000	0.051	0.082	0.200	0.079
0.000	0.051	0.082	0.200	0.079
0.000	0.051	0.082	0.200	0.079
0.000	0.051	0.082	0.200	0.079
0.000	0.051	0.082	0.200	0.079
0.000	0.050	0.082	0.200	0.079

Standard Deviations	0.0005	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

Pgam

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

02/19/2019
Date

*3/4/19
JA*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000883	Miami-Dade Police Department	02/18/2019	<i>JKK</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.05C <input checked="" type="checkbox"/> 0.047 to 0.053 <input checked="" type="checkbox"/> MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000883 02/18/2019 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:15 Control Test 0.050 11:15 Air Blank 0.000 11:16 Control Test 0.050 11:17 Air Blank 0.000 11:17 Control Test 0.049 11:18 Air Blank 0.000 11:18 Control Test Status Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625 Operator's Signature <i>DELL</i> 3/4/19 <i>JKK</i>	SN: SD3968 Temp: 34.06C <input checked="" type="checkbox"/> 0.077 to 0.083 <input checked="" type="checkbox"/> MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000883 02/18/2019 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:20 Control Test 0.081 11:20 Air Blank 0.000 11:21 Control Test 0.081 11:22 Air Blank 0.000 11:22 Control Test 0.081 11:23 Air Blank 0.000 11:23 Control Test Status Average 0.0810 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>JKK</i>	SN: SD3969 Temp: 34.07C <input checked="" type="checkbox"/> 0.194 to 0.206 <input checked="" type="checkbox"/> MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000883 02/18/2019 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:25 Control Test 0.200 11:25 Air Blank 0.000 11:26 Control Test 0.198 11:27 Air Blank 0.000 11:27 Control Test 0.199 11:28 Air Blank 0.000 11:28 Control Test Status Average 0.1990 Std Dev 0.0010 Rel Std Dev(%) 0.5025 Operator's Signature <i>DELL</i>	Lot AG805701 <input checked="" type="checkbox"/> 0.077 to 0.083 <input checked="" type="checkbox"/> MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000883 02/18/2019 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:30 Control Test 0.080 11:30 Air Blank 0.000 11:30 Control Test 0.078 11:31 Air Blank 0.000 11:31 Control Test 0.079 11:32 Air Blank 0.000 11:32 Control Test Status Average 0.0790 Std Dev 0.0010 Rel Std Dev(%) 1.2658 Operator's Signature <i>JKK</i>



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000883, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000883</u>	UNCERTAINTY* ±
Owning Agency:	<u>MIAMI-DADE PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/19/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>08:13</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

<u>02/19/2019</u>	<u><i>David Reyes-Rivera</i></u>
Date	DAVID E REYES-RIVERA, Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

BDM
3/4/19
JQ