



INSTRUMENT PROCESSING SHEET

Agency Franklin County SOS/N 80-000952

Florida Department of Law Enforcement

Date In 06/11/2019 DI Completion Date 06/14/2019 Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>JD</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>223</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																												
Final Release Date <p style="text-align: center;">FDLE</p> <p style="text-align: center;">JUN 17 2019</p> <p style="text-align: center;">Alcohol Testing Program</p>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804 11/14/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 07/25/2019	0.080	DR1279	201707E 07/25/2019	0.200	SD1013	201707C 07/24/2019	0.080 DGS	N/A	AG831804 11/14/2020	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>JD</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.0</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1012</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1013</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	SD1012	201707D 07/25/2019																																																												
0.080	DR1279	201707E 07/25/2019																																																												
0.200	SD1013	201707C 07/24/2019																																																												
0.080 DGS	N/A	AG831804 11/14/2020																																																												
Calibration Adjustment Performed By <u>JD</u> Barometric Pressure Gauge <u>1019</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1016</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1022</td> <td>17410</td> <td>12/06/2019</td> </tr> <tr> <td>0.100</td> <td>SD3964</td> <td>18200</td> <td>07/03/2020</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>19040</td> <td>01/29/2021</td> </tr> <tr> <td>0.300</td> <td>SD1024</td> <td>18110</td> <td>04/02/2020</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A2</td> <td>08/05/2019</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks	Simulator	Serial Number	Lot Number	Expiration	0.000	SD1016	N/A	N/A	0.040	SD1022	17410	12/06/2019	0.100	SD3964	18200	07/03/2020	0.200	SD1025	19040	01/29/2021	0.300	SD1024	18110	04/02/2020	0.080 DGS	N/A	17817080A2	08/05/2019	Department Inspection Performed By <u>JD</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1018</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2408</td> </tr> <tr> <td>Interferent</td> <td>G2882</td> </tr> <tr> <td>0.050</td> <td>SD1012</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	SD1013	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D</td> <td>07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E</td> <td>07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201707C</td> <td>07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804</td> <td>11/14/2020</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1012	201707D	07/25/2019	0.080	DR1279	201707E	07/25/2019	0.200	SD1013	201707C	07/24/2019	0.080 DGS	N/A	AG831804	11/14/2020
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000	SD1016	N/A	N/A																																																											
0.040	SD1022	17410	12/06/2019																																																											
0.100	SD3964	18200	07/03/2020																																																											
0.200	SD1025	19040	01/29/2021																																																											
0.300	SD1024	18110	04/02/2020																																																											
0.080 DGS	N/A	17817080A2	08/05/2019																																																											
Simulator	Serial Number																																																													
0.000	G2408																																																													
Interferent	G2882																																																													
0.050	SD1012																																																													
0.080	DR1279																																																													
0.200	SD1013																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050	SD1012	201707D	07/25/2019																																																											
0.080	DR1279	201707E	07/25/2019																																																											
0.200	SD1013	201707C	07/24/2019																																																											
0.080 DGS	N/A	AG831804	11/14/2020																																																											
Notes/Suggested Service: <u>Please change the level 2</u> <u>password to something unique.</u> <u>I ran 10 additional 0.08 DGS tests with a different</u> <u>tank (Lot #831804) post Department Inspection.</u> <u>I placed a fan in the room to clear ambient ethanol</u> <u>vapor that may have been present during the</u> <u>Department Inspection.</u> <u>JD</u>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>Pagan 6/17/19</u> <u>Britt Kirkland 6/17/19</u> Tech Review / Date Admin Review / Date																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FRANKLIN COUNTY SO
Time of Inspection: 16:03

Date of Inspection: 06/14/2019

Serial Number: 80-000952
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.051	0.081	0.202	0.079
0.000	0.051	0.081	0.202	0.078
0.000	0.051	0.081	0.202	0.077
0.000	0.050	0.082	0.202	0.076
0.000	0.052	0.082	0.202	0.076
0.000	0.051	0.081	0.203	0.076
0.000	0.050	0.081	0.202	0.075
0.000	0.051	0.082	0.203	0.075
0.000	0.050	0.082	0.201	0.076
0.000	0.051	0.081	0.203	0.076

Standard Deviations	0.0006	0.0005	0.0006	0.0012
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

PGM
 BK
 6/17/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

THOMAS J GRAHAM

06/14/2019
 Date

80-000952

6/14/19
JA

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:40
Control Test	0.046	11:41
Air Blank	0.000	11:41
Control Test	0.046	11:42
Air Blank	0.000	11:43
Control Test	0.046	11:43
Air Blank	0.000	11:44
Control Test Stats		
Average	0.0460	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JA

Operator's Signature

ADAM
BK 6/14/19

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:45
Control Test	0.076	11:46
Air Blank	0.000	11:46
Control Test	0.075	11:47
Air Blank	0.000	11:48
Control Test	0.076	11:48
Air Blank	0.000	11:49
Control Test Stats		
Average	0.0757	
Std Dev	0.0006	
Rel Std Dev(%)	0.7630	

JA

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:35
Control Test	0.195	11:36
Air Blank	0.000	11:37
Control Test	0.194	11:37
Air Blank	0.000	11:38
Control Test	0.193	11:39
Air Blank	0.000	11:39
Control Test Stats		
Average	0.1940	
Std Dev	0.0010	
Rel Std Dev(%)	0.5155	

JA

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:57
Control Test	0.079	11:57
Air Blank	0.000	11:57
Control Test	0.079	11:58
Air Blank	0.000	11:58
Control Test	0.079	11:59
Air Blank	0.000	11:59
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

JA

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000952, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000952</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FRANKLIN COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/14/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:03</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/14/2019

Date


THOMAS J GRAHAM,
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Handwritten notes:
809M
TK 6/17/19

80-000952 Calibration Adjustment

6/14/19 JD

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5690 (-0.0360)
 Sample #2 = 1.5250 (-0.0300)
 Sample #3 = 1.5290 (-0.0210)
 Sample #4 = 1.5700 (-0.0000)
 Avg % Abs = 1.5413 (-0.0170)
 STD DEV = 0.0249 (0.0154)
 REL STD DEV = 1.616 (90.558)

FRANKLIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000
 SN 80-000952
 06/14/2019 12:52:31
 Auto Calibration
 Max Power Res Value = 55
 Auto Range Res Value = 39

Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12862, Sum Io = 13368
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.9800 (-0.0240)
 Sample #2 = 1.9600 (-0.0160)
 Sample #3 = 1.9720 (-0.0190)
 Sample #4 = 1.9400 (-0.0190)
 Avg % Abs = 1.9573 (-0.0053)
 STD DEV = 0.0162 (0.0211)
 REL STD DEV = 0.826 (396.124)

Sol Value = 0.000 g/210L ***
 Fit value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12877, Sum Io = 13376
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1960 (-0.0180)
 Sample #2 = 0.1760 (-0.0020)
 Sample #3 = 0.1720 (-0.0160)
 Sample #4 = 0.1680 (-0.0490)
 Avg % Abs = 0.1720 (-0.0223)
 STD DEV = 0.0040 (0.0241)
 REL STD DEV = 2.326 (108.052)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5880 (-0.0180)
 Sample #2 = 3.5000 (-0.0600)
 Sample #3 = 3.5340 (-0.0310)
 Sample #4 = 3.5490 (-0.0260)
 Avg % Abs = 3.5277 (-0.0390)
 STD DEV = 0.0251 (0.0184)
 REL STD DEV = 0.712 (47.071)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1760 (-0.0180)
 Sample #2 = 0.1730 (-0.0660)
 Sample #3 = 0.1770 (-0.0190)
 Sample #4 = 0.1650 (-0.0290)
 Avg % Abs = 0.1717 (-0.0180)
 STD DEV = 0.0061 (0.0115)
 REL STD DEV = 3.559 (64.070)

Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12862, Sum Io = 13369
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.7480 (-0.0170)
 Sample #2 = 3.6630 (-0.0530)
 Sample #3 = 3.6750 (-0.0520)
 Sample #4 = 3.6610 (-0.0910)
 Avg % Abs = 3.6663 (-0.0657)
 STD DEV = 0.0076 (0.0219)
 REL STD DEV = 0.207 (33.410)

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12870, Sum Io = 13371
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.9300 (-0.0140)
 Sample #2 = 0.8980 (-0.0310)
 Sample #3 = 0.8730 (-0.0430)
 Sample #4 = 0.9220 (-0.0250)
 Avg % Abs = 0.8977 (-0.0330)
 STD DEV = 0.0245 (0.0092)
 REL STD DEV = 2.729 (27.773)

Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 -0.000 0.0003
 0.040 0.040 -0.0003
 0.100 0.100 -0.0004
 0.200 0.199 0.0006
 0.300 0.300 -0.0002

Solution Stats Quadratic Fit Chan 2
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 -0.000 0.0004
 0.040 0.040 -0.0002
 0.100 0.101 -0.0007
 0.200 0.199 0.0008
 0.300 0.300 -0.0003

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.172
 Std Dev = 0.00 Rel Std Dev = 2.33
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.898
 Std Dev = 0.02 Rel Std Dev = 2.73
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.957
 Std Dev = 0.02 Rel Std Dev = 0.83
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.666
 Std Dev = 0.01 Rel Std Dev = 0.21
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.365
 Std Dev = 0.03 Rel Std Dev = 0.47
 Zero Order Coef = -471.75
 First Order Coef = 2645.21
 Second Order Coef = 19.98
 Standard Deviation = 21.266865

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.172
 Std Dev = 0.01 Rel Std Dev = 3.56
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.541
 Std Dev = 0.02 Rel Std Dev = 1.62
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.528
 Std Dev = 0.03 Rel Std Dev = 0.71
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.639
 Std Dev = 0.01 Rel Std Dev = 0.17
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.693
 Std Dev = 0.00 Rel Std Dev = 0.05
 Zero Order Coef = -256.75
 First Order Coef = 1391.90
 Second Order Coef = 11.34
 Standard Deviation = 28.810675

Sol Value = 0.300 g/210L ***
 Fit value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12858, Sum Io = 13361
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 5.3900 (-0.0010)
 Sample #2 = 5.3910 (-0.0180)
 Sample #3 = 5.3410 (-0.0370)
 Sample #4 = 5.3640 (-0.0300)
 Avg % Abs = 5.3653 (-0.0283)
 STD DEV = 0.0250 (0.0096)
 REL STD DEV = 0.466 (33.914)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 9.7450 (-0.0010)
 Sample #2 = 9.6870 (-0.0840)
 Sample #3 = 9.6950 (-0.0510)
 Sample #4 = 9.6960 (-0.0710)
 Avg % Abs = 9.6927 (-0.0687)
 STD DEV = 0.0049 (0.0166)
 REL STD DEV = 0.051 (24.209)

ASGM
 ASK 6/17/19

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1019
 3 um H2O Adjust (mg/l*10,000) = 1033
 9 um H2O Adjust (mg/l*10,000) = 602
 **** AUTO CAL PASS

80-000952

6/14/19
JD

Post Stability
Checks

FRANKLIN COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:52
Control Test	0.050	13:52
Air Blank	0.000	13:53
Control Test	0.050	13:53
Air Blank	0.000	13:54
Control Test	0.050	13:55
Air Blank	0.000	13:55
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

JD

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:42
Control Test	0.083	13:43
Air Blank	0.000	13:43
Control Test	0.083	13:44
Air Blank	0.000	13:45
Control Test	0.083	13:45
Air Blank	0.000	13:46
Control Test Stats		
Average	0.0830	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

Repeated
Test. 6/14/19
JD
Cleared JD
tubing. JD

JD

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:03
Control Test	0.081	14:04
Air Blank	0.000	14:04
Control Test	0.081	14:05
Air Blank	0.000	14:06
Control Test	0.081	14:06
Air Blank	0.000	14:07
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

JD

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:47
Control Test	0.202	13:47
Air Blank	0.000	13:48
Control Test	0.200	13:49
Air Blank	0.000	13:49
Control Test	0.201	13:50
Air Blank	0.000	13:51
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel. Std Dev(%)	0.4975	

JD

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:00
Control Test	0.079	14:00
Air Blank	0.000	14:01
Control Test	0.079	14:01
Air Blank	0.000	14:01
Control Test	0.079	14:02
Air Blank	0.000	14:02
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

DGS

JD

Operator's Signature

JD

6/17/19

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
06/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:06
Control Test	0.079	16:06
Air Blank	0.000	16:07
Control Test	0.079	16:07
Air Blank	0.000	16:08
Control Test	0.079	16:08
Air Blank	0.000	16:09
Control Test	0.079	16:09
Air Blank	0.000	16:09
Control Test	0.078	16:10
Air Blank	0.000	16:10
Control Test	0.079	16:11
Air Blank	0.000	16:11
Control Test	0.079	16:11
Air Blank	0.000	16:12
Control Test	0.079	16:12
Air Blank	0.000	16:13
Control Test	0.079	16:13
Air Blank	0.000	16:14
Control Test	0.078	16:14
Air Blank	0.000	16:14
Control Test Stats		
Average	0.0788	
Std Dev	0.0004	
Rel Std Dev(%)	0.5351	



Operator's Signature


BK
6/17/19



INSTRUMENT PROCESSING SHEET

Agency Franklin County SO

S/N 80-000952

Florida Department of Law Enforcement

Date In 04/11/2019 DI Completion Date 4/16/19

Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>229</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	---	--

Final Release Date

FDLE

APR 19 2019

Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	SD1021	201707D 7/25/19
0.080	DR1275	201707E 7/25/19
0.200	SD1011	201707C 7/24/19
0.080 DGS	N/A	AG831804 11/14/20

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By DP

Lab Temp °C 20.8
 External Digital Therm. ID#: 300505
 34°C +/-2 Serial #: SD1021
 34°C +/-2 Serial #: DR1275
 34°C +/-2 Serial #: SD1011

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DP

Barometric Pressure ID# 30793
 Gauge 1017 Instrument 1016
 Mouth Alcohol Solution Lot # 2018-B
 Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	611621
Interferent	DR3855
0.050	SD1021
0.080	DR1275
0.200	SD1011

Attachments

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration
<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40
<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

SP 4/18/19 J. Debra 4/19/19
 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FRANKLIN COUNTY SO
Time of Inspection: 11:18

Date of Inspection: 04/16/2019

Serial Number: 80-000952
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.047	0.079	0.195	0.080
0.000	0.048	0.079	0.197	0.080
0.000	0.048	0.080	0.197	0.079
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.080	0.197	0.078
0.000	0.048	0.080	0.197	0.078
0.000	0.048	0.080	0.196	0.077
0.000	0.048	0.080	0.197	0.077
0.000	0.048	0.080	0.196	0.077
0.000	0.048	0.079	0.197	0.076

Standard Deviations	0.0003	0.0005	0.0006	0.0013
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

04/16/2019
Date

4/19/19
JP

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
04/16/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:21
Control Test	0.048	08:22
Air Blank	0.000	08:23
Control Test	0.047	08:23
Air Blank	0.000	08:24
Control Test	0.047	08:25
Air Blank	0.000	08:25
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	

P Murphy

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
04/16/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:27
Control Test	0.079	08:28
Air Blank	0.000	08:28
Control Test	0.078	08:29
Air Blank	0.000	08:30
Control Test	0.079	08:30
Air Blank	0.000	08:31
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

P Murphy

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
04/16/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:33
Control Test	0.197	08:34
Air Blank	0.000	08:34
Control Test	0.194	08:35
Air Blank	0.000	08:36
Control Test	0.194	08:36
Air Blank	0.000	08:37
Control Test Stats		
Average	0.1950	
Std Dev	0.0017	
Rel Std Dev(%)	0.8882	

P Murphy

Operator's Signature

SP

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
04/16/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:39
Control Test	0.079	08:39
Air Blank	0.000	08:39
Control Test	0.079	08:40
Air Blank	0.000	08:40
Control Test	0.078	08:41
Air Blank	0.000	08:41
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DGS

P Murphy

Operator's Signature

4/19/19
JP



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000952, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000952</u>	UNCERTAINTY* ±
Owning Agency:	<u>FRANKLIN COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/16/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:18</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

SP

<u>04/16/2019</u>	<u><i>Patrick J Murphy</i></u>
Date	PATRICK J MURPHY, Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

4/19/19
JM