



# INSTRUMENT PROCESSING SHEET

Agency Hollywood PDS/N 80-001063Florida Department of  
Law EnforcementDate In 08/01/2019DI Completion Date 08/01/2019 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>MH</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>MH</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>219</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.187</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
<b>Final Release Date</b>  <div style="text-align: center;"> <b>FDLE</b>   <u>AUG 12 2019</u>   <b>Alcohol Testing Program</b> </div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804 11/14/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	MP4864	201905B 05/14/2021	0.200	MP5097	201904D 04/30/2021	0.080 DGS	N/A	AG831804 11/14/2020	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>MH</u> <input checked="" type="checkbox"/> Lab Temp °C <u>23.18</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5097</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP4863	201905A 05/14/2021																																																											
0.080	MP4864	201905B 05/14/2021																																																											
0.200	MP5097	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG831804 11/14/2020																																																											
<b>Calibration Adjustment</b> Performed By <u>MH</u> Barometric Pressure Gauge <u>1015</u> ID # <u>28199</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>19080</td> <td>03/04/2021</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>18200</td> <td>07/03/2020</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>19040</td> <td>01/29/2021</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>18110</td> <td>04/02/2020</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06/05/2021</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A</td> <td>05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B</td> <td>05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D</td> <td>04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804</td> <td>11/14/2020</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5098	19080	03/04/2021	0.100	MP5099	18200	07/03/2020	0.200	MP5100	19040	01/29/2021	0.300	MP5101	18110	04/02/2020	0.080 DGS	N/A	08819080A1	06/05/2021	Simulator	Serial Number	Lot Number	Expiration	0.050	MP4863	201905A	05/14/2021	0.080	MP4864	201905B	05/14/2021	0.200	MP5097	201904D	04/30/2021	0.080 DGS	N/A	AG831804	11/14/2020	<b>Department Inspection</b> Performed By <u>MH</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1015</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	MP5095	N/A	N/A																																																										
0.040	MP5098	19080	03/04/2021																																																										
0.100	MP5099	18200	07/03/2020																																																										
0.200	MP5100	19040	01/29/2021																																																										
0.300	MP5101	18110	04/02/2020																																																										
0.080 DGS	N/A	08819080A1	06/05/2021																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	MP4863	201905A	05/14/2021																																																										
0.080	MP4864	201905B	05/14/2021																																																										
0.200	MP5097	201904D	04/30/2021																																																										
0.080 DGS	N/A	AG831804	11/14/2020																																																										
Simulator	Serial Number																																																												
0.000	SD1014																																																												
Interferent	SD1015																																																												
0.050	MP4863																																																												
0.080	MP4864																																																												
0.200	MP5097																																																												
Notes/Suggested Service: <u>E-mailed</u>  <div style="border: 1px solid blue; padding: 5px; display: inline-block;"> <input checked="" type="checkbox"/> <b>APPROVED</b> <u>08/02/2019</u> </div>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  <u>PJPM 8/12/19</u> <u>Brett Keildand 8/12/19</u> Tech Review / Date Admin Review / Date																																																												

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLLYWOOD PD  
Time of Inspection: 15:00

Date of Inspection: 08/01/2019

Serial Number: 80-001063  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.080

Standard Deviations	0.0000	0.0000	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

PJM  
12K  
8/12/19

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Michael D Haughey*

MICHAEL D HAUGHEY

Signature and Printed Name

08/01/2019  
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-00 1663	Hollywood PD	08/01/2019	MH

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

HOLLYWOOD PD  
Intoxilizer - Alcohol Analyzer  
Model 8000 SN 80-001063  
08/01/2019  
Software: 8100.27

Test 9/210L Time

Air Blank 0.006 10:11  
Control Test 0.051 10:11  
Air Blank 0.000 10:12  
Control Test 0.051 10:13  
Air Blank 0.000 10:13  
Control Test 0.052 10:14  
Air Blank 0.000 10:14  
Control Test Stats  
Average 0.0513  
Std Dev 0.0006  
Rel Std Dev(%) 1.1247

MH  
Operator's Signature

HOLLYWOOD PD  
Intoxilizer - Alcohol Analyzer  
Model 8000 SN 80-001063  
08/01/2019  
Software: 8100.27

Test 9/210L Time

Air Blank 0.000 10:22  
Control Test 0.084 10:22  
Air Blank 0.000 10:23  
Control Test 0.083 10:24  
Air Blank 0.000 10:24  
Control Test 0.084 10:25  
Air Blank 0.000 10:25  
Control Test Stats  
Average 0.0837  
Std Dev 0.0006  
Rel Std Dev(%) 0.6901

MH  
Operator's Signature

HOLLYWOOD PD  
Intoxilizer - Alcohol Analyzer  
Model 8000 SN 80-001063  
08/01/2019  
Software: 8100.27

Test 9/210L Time

Air Blank 0.000 10:16  
Control Test 0.211 10:16  
Air Blank 0.000 10:17  
Control Test 0.210 10:18  
Air Blank 0.000 10:18  
Control Test 0.211 10:19  
Air Blank 0.000 10:19  
Control Test Stats  
Average 0.2107  
Std Dev 0.0006  
Rel Std Dev(%) 0.2741

MH  
Operator's Signature

HOLLYWOOD PD  
Intoxilizer - Alcohol Analyzer  
Model 8000 SN 80-001063  
08/01/2019  
Software: 8100.27

Test 9/210L Time

Air Blank 0.000 10:28  
Control Test 0.078 10:28  
Air Blank 0.000 10:28  
Control Test 0.079 10:29  
Air Blank 0.000 10:29  
Control Test 0.079 10:29  
Air Blank 0.000 10:30  
Control Test Stats  
Average 0.0787  
Std Dev 0.0006  
Rel Std Dev(%) 0.7335

MH  
Operator's Signature

CPM  
BK 8/12/19



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 4700 Terminal Drive, Suite 1  
 Ft. Myers, FL 33907

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001063, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001063</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HOLLYWOOD PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/01/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:00</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/01/2019

Date

**MICHAEL D HAUGHEY,**  
 Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*Handwritten notes:*  
 QDM  
 TSK  
 8/12/19

Innovalyzer - Alcohol Analyzer  
Model 8000  
18/01/2019  
SN 80-001063  
11:16:28

Auto Calibration  
Max Power Res Value = 81  
Auto Range Res Value = 45

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5890 (0.0010)  
Sample #2 = 1.6090 (0.0020)  
Sample #3 = 1.5770 (0.0200)  
Sample #4 = 1.5790 (0.0180)  
Avg % Abs = 1.5883 (0.0133)  
STD DEV = 0.0179 (0.0099)  
REL STD DEV = 1.129 (73.993)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 7.1350 (-0.0330)  
Sample #2 = 7.1070 (0.0190)  
Sample #3 = 7.0790 (0.0130)  
Sample #4 = 7.0750 (0.0310)  
Avg % Abs = 7.0873 (0.0210)  
STD DEV = 0.0171 (0.0092)  
REL STD DEV = 0.241 (43.644)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
<<<<< CHANNEL 1 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.120  
Std Dev = 0.02 Rel Std Dev = 16.43  
Sol Val = 0.1905 mg/l or 0.140 g/210L  
% Abs = 0.895  
Std Dev = 0.01 Rel Std Dev = 0.59  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 2.109  
Std Dev = 0.01 Rel Std Dev = 0.57  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.809  
Std Dev = 0.02 Rel Std Dev = 0.50  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.535  
Std Dev = 0.02 Rel Std Dev = 0.40  
Zero Order Coef = -302.66  
First Order Coef = 2452.85  
Second Order Coef = 33.07  
Standard Deviation = 8.933813

Solution Stats Quadratic Fit Chan 2  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 0.000 0.000  
0.040 0.040 0.0002  
0.100 0.100 0.0001  
0.200 0.200 -0.0002  
0.300 0.300 0.0001

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12426, Sum Io = 13707  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 2.0110 (-0.0320)  
Sample #2 = 2.0080 (-0.0130)  
Sample #3 = 2.0210 (0.0070)  
Sample #4 = 1.9580 (0.0210)  
Avg % Abs = 2.0090 (0.0050)  
STD DEV = 0.0115 (0.0171)  
REL STD DEV = 0.574 (341.760)

Sol Value = 0.310 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12416, Sum Io = 13699  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.5500 (0.0070)  
Sample #2 = 5.5340 (0.0150)  
Sample #3 = 5.5140 (0.0430)  
Sample #4 = 5.5580 (0.0310)  
Avg % Abs = 5.5353 (0.0297)  
STD DEV = 0.0220 (0.0140)  
REL STD DEV = 0.398 (47.351)

<<<<< CHANNEL 2 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.124  
Std Dev = 0.02 Rel Std Dev = 15.22  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.588  
Std Dev = 0.02 Rel Std Dev = 1.13  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.718  
Std Dev = 0.01 Rel Std Dev = 0.27  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 7.087  
Std Dev = 0.02 Rel Std Dev = 0.24  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 10.231  
Std Dev = 0.01 Rel Std Dev = 0.08  
Zero Order Coef = -150.83  
First Order Coef = 1267.36  
Second Order Coef = 14.00  
Standard Deviation = 8.508168

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 2980.00  
Sample #2 = 3028.00  
Sample #3 = 3112.00  
Sample #4 = 3100.00  
Average Result = 3060.0000  
STD DEV = 45.4313  
REL STD DEV = 1.475  
\*\*\*\*\*  
\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample #1 = 3353.00  
Sample #2 = 3406.00  
Sample #3 = 3423.00  
Sample #4 = 3452.00  
Average Result = 3427.0000  
STD DEV = 23.2594  
REL STD DEV = 0.679  
\*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1015  
3 um H2O Adjust (mg/l\*10,000) = 729  
9 um H2O Adjust (mg/l\*10,000) = 382  
\*\*\*\*\* AUTO CAL PASS \*\*\*\*\*

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12422, Sum Io = 13704  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.8710 (-0.0230)  
Sample #2 = 3.8170 (0.0330)  
Sample #3 = 3.8220 (0.0190)  
Sample #4 = 3.7870 (0.0550)  
Avg % Abs = 3.8187 (0.0323)  
STD DEV = 0.0189 (0.0230)  
REL STD DEV = 0.497 (71.156)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 10.2370 (-0.0090)  
Sample #2 = 10.2320 (0.0110)  
Sample #3 = 10.2220 (0.0140)  
Sample #4 = 10.2360 (0.0170)  
Avg % Abs = 10.2307 (0.0140)  
STD DEV = 0.0081 (0.0030)  
REL STD DEV = 0.079 (21.429)

<<<<< CHANNEL 2 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.124  
Std Dev = 0.02 Rel Std Dev = 15.22  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.588  
Std Dev = 0.02 Rel Std Dev = 1.13  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.718  
Std Dev = 0.01 Rel Std Dev = 0.27  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 7.087  
Std Dev = 0.02 Rel Std Dev = 0.24  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 10.231  
Std Dev = 0.01 Rel Std Dev = 0.08  
Zero Order Coef = -150.83  
First Order Coef = 1267.36  
Second Order Coef = 14.00  
Standard Deviation = 8.508168

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12436, Sum Io = 13710  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8970 (0.0130)  
Sample #2 = 0.9010 (0.0190)  
Sample #3 = 0.8930 (0.0290)  
Sample #4 = 0.8910 (0.0570)  
Avg % Abs = 0.8950 (0.0350)  
STD DEV = 0.0053 (0.0242)  
REL STD DEV = 0.591 (69.282)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12436, Sum Io = 13710  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8970 (0.0130)  
Sample #2 = 0.9010 (0.0190)  
Sample #3 = 0.8930 (0.0290)  
Sample #4 = 0.8910 (0.0570)  
Avg % Abs = 0.8950 (0.0350)  
STD DEV = 0.0053 (0.0242)  
REL STD DEV = 0.591 (69.282)

<<<<< CHANNEL 2 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.124  
Std Dev = 0.02 Rel Std Dev = 15.22  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.588  
Std Dev = 0.02 Rel Std Dev = 1.13  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.718  
Std Dev = 0.01 Rel Std Dev = 0.27  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 7.087  
Std Dev = 0.02 Rel Std Dev = 0.24  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 10.231  
Std Dev = 0.01 Rel Std Dev = 0.08  
Zero Order Coef = -150.83  
First Order Coef = 1267.36  
Second Order Coef = 14.00  
Standard Deviation = 8.508168

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12436, Sum Io = 13710  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8970 (0.0130)  
Sample #2 = 0.9010 (0.0190)  
Sample #3 = 0.8930 (0.0290)  
Sample #4 = 0.8910 (0.0570)  
Avg % Abs = 0.8950 (0.0350)  
STD DEV = 0.0053 (0.0242)  
REL STD DEV = 0.591 (69.282)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12436, Sum Io = 13710  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8970 (0.0130)  
Sample #2 = 0.9010 (0.0190)  
Sample #3 = 0.8930 (0.0290)  
Sample #4 = 0.8910 (0.0570)  
Avg % Abs = 0.8950 (0.0350)  
STD DEV = 0.0053 (0.0242)  
REL STD DEV = 0.591 (69.282)

Optical Calibration  
SN: 80-001063  
Agency: Hollywood PD  
Date: 08/01/2019  
By: WVA

*WVA*  
8/12/19

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities - Post	80-001063	Hollywood PD	08/01/2019	MH

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

HOLLYWOOD PD  
Intoxilyzer - Alconol Analyzer  
Model 8000  
08/01/2019  
SN 80-001063  
Software: 8100.27

HOLLYWOOD PD  
Intoxilyzer - Alconol Analyzer  
Model 8000  
08/01/2019  
SN 80-001063  
Software: 8100.27

HOLLYWOOD PD  
Intoxilyzer - Alconol Analyzer  
Model 8000  
08/01/2019  
SN 80-001063  
Software: 8100.27

HOLLYWOOD PD  
Intoxilyzer - Alconol Analyzer  
Model 8000  
08/01/2019  
SN 80-001063  
Software: 8100.27

g/210L

Test	Time
Air Blank	12:08
Control Test	12:09
Air Blank	12:09
Control Test	12:10
Air Blank	12:10
Control Test	12:11
Air Blank	12:12
Control Test	12:12
Average	0.0480
Std Dev	0.0010
Rel Std Dev(%)	2.0833

g/210L

Test	Time
Air Blank	12:14
Control Test	12:15
Air Blank	12:16
Control Test	12:16
Air Blank	12:17
Control Test	12:17
Air Blank	12:18
Control Test	12:18
Average	0.0780
Std Dev	0.0010
Rel Std Dev(%)	1.2821

g/210L

Test	Time
Air Blank	12:19
Control Test	12:20
Air Blank	12:20
Control Test	12:21
Air Blank	12:21
Control Test	12:22
Air Blank	12:22
Control Test	12:23
Average	0.1981
Std Dev	0.0000
Rel Std Dev(%)	0.0000

g/210L

Test	Time
Air Blank	12:24
Control Test	12:25
Air Blank	12:25
Control Test	12:26
Air Blank	12:26
Control Test	12:26
Air Blank	12:27
Control Test	12:27
Average	0.0790
Std Dev	0.0000
Rel Std Dev(%)	0.0000

MH  
Operator's Signature

MH  
Operator's Signature

MH  
Operator's Signature

MH  
Operator's Signature

ROOM  
75K  
8/12/19