



# INSTRUMENT PROCESSING SHEET

Agency Orange County SOS/N 80-001258

Florida Department of Law Enforcement

Date In 02/26/2019 DI Completion Date 3/4/19 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>JE</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>JE</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>179</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.179</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
<b>Final Release Date</b>  <div style="text-align: center;"> <b>FDLE</b>   <b>MAR 05 2019</b>   <b>Alcohol Testing Program</b> </div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG805701</u> <u>2/24/20</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	201707D 07/25/2019	0.080	<u>DR1279</u>	201707E 07/25/2019	0.200	<u>DR3856</u>	201707C 07/24/2019	0.080 DGS	N/A	<u>AG805701</u> <u>2/24/20</u>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>JE</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.8</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1012</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR3856</u>																								
Simulator	Serial #	Lot #/Exp																																							
0.050	<u>SD1012</u>	201707D 07/25/2019																																							
0.080	<u>DR1279</u>	201707E 07/25/2019																																							
0.200	<u>DR3856</u>	201707C 07/24/2019																																							
0.080 DGS	N/A	<u>AG805701</u> <u>2/24/20</u>																																							
<b>Calibration Adjustment</b> Performed By <u>JE</u> Barometric Pressure Gauge <u>1015</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G4444</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1022</u></td> <td><u>17410</u></td> <td><u>12/6/19</u></td> </tr> <tr> <td>0.100</td> <td><u>SD3964</u></td> <td><u>18070</u></td> <td><u>2/26/20</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>19040</u></td> <td><u>1/29/21</u></td> </tr> <tr> <td>0.300</td> <td><u>SD1024</u></td> <td><u>18110</u></td> <td><u>4/2/20</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>17817080A2</u></td> <td><u>8/5/19</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G4444</u>	N/A	N/A	0.040	<u>SD1022</u>	<u>17410</u>	<u>12/6/19</u>	0.100	<u>SD3964</u>	<u>18070</u>	<u>2/26/20</u>	0.200	<u>SD1025</u>	<u>19040</u>	<u>1/29/21</u>	0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>	0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>	<b>Department Inspection</b> Performed By <u>JE</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1015</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2408</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2882</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1012</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2408</u>	Interferent	<u>G2882</u>	0.050	<u>SD1012</u>	0.080	<u>DR1279</u>	0.200	<u>DR3856</u>
Simulator	Serial Number	Lot Number	Expiration																																						
0.000	<u>G4444</u>	N/A	N/A																																						
0.040	<u>SD1022</u>	<u>17410</u>	<u>12/6/19</u>																																						
0.100	<u>SD3964</u>	<u>18070</u>	<u>2/26/20</u>																																						
0.200	<u>SD1025</u>	<u>19040</u>	<u>1/29/21</u>																																						
0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>																																						
0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>																																						
Simulator	Serial Number																																								
0.000	<u>G2408</u>																																								
Interferent	<u>G2882</u>																																								
0.050	<u>SD1012</u>																																								
0.080	<u>DR1279</u>																																								
0.200	<u>DR3856</u>																																								
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG805701</u></td> <td><u>2/24/20</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1012</u>	<u>201707D</u>	<u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>	0.200	<u>DR3856</u>	<u>201707C</u>	<u>7/24/19</u>	0.080 DGS	N/A	<u>AG805701</u>	<u>2/24/20</u>	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																				
Simulator	Serial Number	Lot Number	Expiration																																						
0.050	<u>SD1012</u>	<u>201707D</u>	<u>7/25/19</u>																																						
0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>																																						
0.200	<u>DR3856</u>	<u>201707C</u>	<u>7/24/19</u>																																						
0.080 DGS	N/A	<u>AG805701</u>	<u>2/24/20</u>																																						
Notes/Suggested Service: <u>Calibration adjustment performed due to barometric sensor not being within 1% (Instrument:1002) Pressure gauge: 1013), JE 3/4/19</u> _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  <u>ppgm 3/5/19</u> <u>Britt Kirkland 3/5/19</u> Tech Review / Date Admin Review / Date																																								

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY S.O.  
Time of Inspection: 12:45

Date of Inspection: 03/04/2019

Serial Number: 80-001258  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.078
0.000	0.048	0.079	0.197	0.078
0.000	0.049	0.080	0.197	0.078
0.000	0.049	0.080	0.197	0.078
0.000	0.048	0.080	0.196	0.079
0.000	0.048	0.080	0.197	0.078
0.000	0.048	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.078

Standard Deviations	0.0005	0.0005	0.0009	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

JG  
BK  
3/5/19

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name      THOMAS J. GRAHAM

03/04/2019  
 Date

80-001258

Stability Checks

3/4/19

JD

ORANGE COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001258  
 03/04/2019  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:50
Control Test	0.049	08:50
Air Blank	0.000	08:51
Control Test	0.049	08:52
Air Blank	0.000	08:52
Control Test	0.049	08:53
Air Blank	0.000	08:53
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD

Operator's Signature

PD

BK 3/5/19

ORANGE COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001258  
 03/04/2019  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:31
Control Test	0.080	08:32
Air Blank	0.000	08:32
Control Test	0.081	08:33
Air Blank	0.000	08:34
Control Test	0.080	08:34
Air Blank	0.000	08:35
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

JD

Operator's Signature

ORANGE COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001258  
 03/04/2019  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:37
Control Test	0.198	08:38
Air Blank	0.000	08:38
Control Test	0.198	08:39
Air Blank	0.000	08:40
Control Test	0.199	08:40
Air Blank	0.000	08:41
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

JD

Operator's Signature

ORANGE COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001258  
 03/04/2019  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.079	08:44
Air Blank	0.000	08:44
Control Test	0.080	08:45
Air Blank	0.000	08:45
Control Test	0.079	08:46
Air Blank	0.000	08:46
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

JD

Operator's Signature



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001258, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001258</u>	UNCERTAINTY* ±	
Owning Agency:	<u>ORANGE COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/04/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:45</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/04/2019

Date

THOMAS J GRAHAM,  
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

RBK 3/5/19

80-001258

# Calibration Adjustment

3/4/19  
JD

ORANGE COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000  
 SN 80-001258  
 03/04/2019 09:30:18

Auto Calibration  
 Max Power Res Value = 34  
 Auto Range Res Value = 25

### <<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 1.4740 (-0.0140)  
 Sample #2 = 1.4690 (-0.0110)  
 Sample #3 = 1.4690 (-0.0010)  
 Sample #4 = 1.4530 (-0.0020)  
 Avg % Abs = 1.4637 (-0.0047)  
 STD DEV = 0.0092 (0.0055)  
 REL STD DEV = 0.631 (118.019)

### Sol Value = 0.100 g/210L \*\*\*

Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12814, Sum Io = 13504

### <<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 1.7330 (-0.0010)  
 Sample #2 = 1.7490 (0.0160)  
 Sample #3 = 1.7740 (0.0200)  
 Sample #4 = 1.7690 (0.0100)  
 Avg % Abs = 1.7640 (0.0153)  
 STD DEV = 0.0132 (0.0050)  
 REL STD DEV = 0.750 (32.825)

### Sol Value = 0.000 g/210L \*\*\*

Fit value = 0.0000 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12825, Sum Io = 13506

### <<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 0.0650 (-0.0030)  
 Sample #2 = 0.0810 (0.0140)  
 Sample #3 = 0.0760 (0.0170)  
 Sample #4 = 0.0910 (0.0500)  
 Avg % Abs = 0.0823 (0.0270)  
 STD DEV = 0.0071 (0.0200)  
 REL STD DEV = 8.617 (73.981)

### <<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 0.1150 (-0.0040)  
 Sample #2 = 0.1300 (0.0020)  
 Sample #3 = 0.1340 (-0.0010)  
 Sample #4 = 0.1250 (0.0080)  
 Avg % Abs = 0.1297 (0.0030)  
 STD DEV = 0.0045 (0.0046)  
 REL STD DEV = 3.478 (152.753)

### Sol Value = 0.040 g/210L \*\*\*

Fit value = 0.1905 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12818, Sum Io = 13504

### <<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7830 (-0.0190)  
 Sample #2 = 0.7510 (0.0230)  
 Sample #3 = 0.7910 (0.0120)  
 Sample #4 = 0.7760 (0.0110)  
 Avg % Abs = 0.7727 (0.0153)  
 STD DEV = 0.0202 (0.0067)  
 REL STD DEV = 2.615 (43.424)

### <<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 1.4740 (-0.0140)  
 Sample #2 = 1.4690 (-0.0110)  
 Sample #3 = 1.4690 (-0.0010)  
 Sample #4 = 1.4530 (-0.0020)  
 Avg % Abs = 1.4637 (-0.0047)  
 STD DEV = 0.0092 (0.0055)  
 REL STD DEV = 0.631 (118.019)

### Sol Value = 0.100 g/210L \*\*\*

Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12814, Sum Io = 13504

### <<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 1.7330 (-0.0010)  
 Sample #2 = 1.7490 (0.0160)  
 Sample #3 = 1.7740 (0.0200)  
 Sample #4 = 1.7690 (0.0100)  
 Avg % Abs = 1.7640 (0.0153)  
 STD DEV = 0.0132 (0.0050)  
 REL STD DEV = 0.750 (32.825)

### <<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 3.4080 (-0.0100)  
 Sample #2 = 3.4140 (0.0140)  
 Sample #3 = 3.4500 (0.0050)  
 Sample #4 = 3.4350 (-0.0070)  
 Avg % Abs = 3.4330 (0.0040)  
 STD DEV = 0.0181 (0.0105)  
 REL STD DEV = 0.527 (263.391)

### Sol Value = 0.200 g/210L \*\*\*

Fit value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12809, Sum Io = 13502

### <<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 3.3660 (-0.0090)  
 Sample #2 = 3.4050 (-0.0020)  
 Sample #3 = 3.4080 (-0.0080)  
 Sample #4 = 3.4290 (0.0060)  
 Avg % Abs = 3.4140 (-0.0013)  
 STD DEV = 0.0131 (0.0079)  
 REL STD DEV = 0.383 (526.783)

### <<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 1.4740 (-0.0140)  
 Sample #2 = 1.4690 (-0.0110)  
 Sample #3 = 1.4690 (-0.0010)  
 Sample #4 = 1.4530 (-0.0020)  
 Avg % Abs = 1.4637 (-0.0047)  
 STD DEV = 0.0092 (0.0055)  
 REL STD DEV = 0.631 (118.019)

### Sol Value = 0.100 g/210L \*\*\*

Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12814, Sum Io = 13504

### <<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 1.7330 (-0.0010)  
 Sample #2 = 1.7490 (0.0160)  
 Sample #3 = 1.7740 (0.0200)  
 Sample #4 = 1.7690 (0.0100)  
 Avg % Abs = 1.7640 (0.0153)  
 STD DEV = 0.0132 (0.0050)  
 REL STD DEV = 0.750 (32.825)

### <<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 3.4080 (-0.0100)  
 Sample #2 = 3.4140 (0.0140)  
 Sample #3 = 3.4500 (0.0050)  
 Sample #4 = 3.4350 (-0.0070)  
 Avg % Abs = 3.4330 (0.0040)  
 STD DEV = 0.0181 (0.0105)  
 REL STD DEV = 0.527 (263.391)

### Sol Value = 0.200 g/210L \*\*\*

Fit value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12809, Sum Io = 13502

### <<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)  
 Sample #1 = 3.3660 (-0.0090)  
 Sample #2 = 3.4050 (-0.0020)  
 Sample #3 = 3.4080 (-0.0080)  
 Sample #4 = 3.4290 (0.0060)  
 Avg % Abs = 3.4140 (-0.0013)  
 STD DEV = 0.0131 (0.0079)  
 REL STD DEV = 0.383 (526.783)

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0004
0.100	0.099	0.0005
0.200	0.200	-0.0004
0.300	0.300	0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0003
0.100	0.100	0.0002
0.200	0.200	-0.0004
0.300	0.300	0.0002

Sol Value = 0.080 g/210L \*\*\*  
 Fit value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sample #1 = 3241.00  
 Sample #2 = 3398.00  
 Sample #3 = 3174.00  
 Sample #4 = 3154.00

Average Result = 3242.0000  
 STD DEV = 135.4696  
 REL STD DEV = 4.179

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sample #1 = 3442.00  
 Sample #2 = 3472.00  
 Sample #3 = 3455.00  
 Sample #4 = 3402.00

Average Result = 3443.0000  
 STD DEV = 36.5103  
 REL STD DEV = 1.060

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1015  
 3 um H2O Adjust (mg/l\*10,000) = 567  
 9 um H2O Adjust (mg/l\*10,000) = 366  
 \*\*\*\*\* AUTO CAL PASS \*\*\*\*\*

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
 <<<< CHANNEL 1 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.082  
 Std Dev = 0.01 Rel Std Dev = 8.62

Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 0.773  
 Std Dev = 0.02 Rel Std Dev = 2.62

Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 1.764  
 Std Dev = 0.01 Rel Std Dev = 0.75

Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 3.414  
 Std Dev = 0.01 Rel Std Dev = 0.38

Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 4.993  
 Std Dev = 0.01 Rel Std Dev = 0.29

Zero Order Coef = -231.75  
 First Order Coef = 2766.67  
 Second Order Coef = 28.05  
 Standard Deviation = 18.162160

<<<< CHANNEL 2 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.130  
 Std Dev = 0.00 Rel Std Dev = 3.48

Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 1.464  
 Std Dev = 0.01 Rel Std Dev = 0.63

Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 3.433  
 Std Dev = 0.02 Rel Std Dev = 0.53

Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 6.586  
 Std Dev = 0.02 Rel Std Dev = 0.36

Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 9.546  
 Std Dev = 0.02 Rel Std Dev = 0.19

Zero Order Coef = -167.79  
 First Order Coef = 1387.70  
 Second Order Coef = 13.16  
 Standard Deviation = 14.496857

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 6.5330 (-0.0090)  
 Sample #2 = 6.5590 (0.0020)  
 Sample #3 = 6.6030 (0.0030)  
 Sample #4 = 6.5960 (0.0290)  
 Avg % Abs = 6.5860 (0.0113)  
 STD DEV = 0.0236 (0.0153)  
 REL STD DEV = 0.359 (135.070)

Sol Value = 0.300 g/210L \*\*\*  
 Fit value = 1.4286 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12807, Sum Io = 13498

<<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 4.9420 (-0.0060)  
 Sample #2 = 5.0010 (0.0040)  
 Sample #3 = 4.9760 (0.0180)  
 Sample #4 = 5.0010 (-0.0040)  
 Avg % Abs = 4.9927 (0.0060)  
 STD DEV = 0.0144 (0.0111)  
 REL STD DEV = 0.289 (185.592)

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 9.5060 (-0.0140)  
 Sample #2 = 9.5650 (-0.0110)  
 Sample #3 = 9.5290 (0.0170)  
 Sample #4 = 9.5440 (-0.0020)  
 Avg % Abs = 9.5460 (0.0013)  
 STD DEV = 0.0181 (0.0143)  
 REL STD DEV = 0.189 (1072.089)

PPM BK 3/5/19

80-001258 Post Stabilities

3/4/19  
JD

ORANGE COUNTY S.O.  
Intoxilyzer - Alcnol Analyzer  
Model 8000 SN 80-001258  
03/04/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.050	10:11
Air Blank	0.000	10:12
Control Test	0.050	10:13
Air Blank	0.000	10:13
Control Test	0.050	10:14
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD  
Operator's Signature

ORANGE COUNTY S.O.  
Intoxilyzer - Alcnol Analyzer  
Model 8000 SN 80-001258  
03/04/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.081	10:16
Air Blank	0.000	10:16
Control Test	0.080	10:17
Air Blank	0.000	10:18
Control Test	0.080	10:18
Air Blank	0.000	10:19
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

JD  
Operator's Signature

ORANGE COUNTY S.O.  
Intoxilyzer - Alcnol Analyzer  
Model 8000 SN 80-001258  
03/04/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:25
Control Test	0.198	10:26
Air Blank	0.000	10:26
Control Test	0.198	10:27
Air Blank	0.000	10:28
Control Test	0.198	10:28
Air Blank	0.000	10:29
Control Test Stats		
Average	0.1960	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD  
Operator's Signature

ORANGE COUNTY S.O.  
Intoxilyzer - Alcnol Analyzer  
Model 8000 SN 80-001258  
03/04/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.078	10:30
Air Blank	0.000	10:30
Control Test	0.079	10:31
Air Blank	0.000	10:31
Control Test	0.078	10:32
Air Blank	0.000	10:32
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

JD  
Operator's Signature

JDAM  
13K 3/5/19