



INSTRUMENT PROCESSING SHEET

Agency Baker County Sheriff's OfficeS/N 80-001287Florida Department of
Law EnforcementDate In 04/08/2019DI Completion Date 04/17/2019☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Performed By <u>DP</u>	Quality Checks	Performed By <u>JD</u>	Flow Calibration	Performed By																																																											
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>195</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																												
Final Release Date FDLE APR 18 2019 Alcohol Testing Program		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>SD1012</td><td>201707D 07/25/2019</td></tr><tr><td>0.080</td><td>DR1279</td><td>201707E 07/25/2019</td></tr><tr><td>0.200</td><td>SD1013</td><td>201707C 07/24/2019</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG831804 11/14/2020</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 07/25/2019	0.080	DR1279	201707E 07/25/2019	0.200	SD1013	201707C 07/24/2019	0.080 DGS	N/A	AG831804 11/14/2020	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>JD</u> ☒ Lab Temp °C <u>20.9</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C +/- .2 Serial #: <u>SD1021</u> ☒ 34°C +/- .2 Serial #: <u>DR1275</u> ☒ 34°C +/- .2 Serial #: <u>SD1011</u>																																													
Simulator	Serial #	Lot #/Exp																																																														
0.050	SD1012	201707D 07/25/2019																																																														
0.080	DR1279	201707E 07/25/2019																																																														
0.200	SD1013	201707C 07/24/2019																																																														
0.080 DGS	N/A	AG831804 11/14/2020																																																														
Calibration Adjustment Performed By <u>JD</u> Barometric Pressure Gauge <u>1018</u> ID # <u>30793</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>DR2035</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>SD1022</td><td>17410</td><td>12/06/2019</td></tr><tr><td>0.100</td><td>SD3964</td><td>18200</td><td>07/03/2020</td></tr><tr><td>0.200</td><td>SD1025</td><td>19040</td><td>01/29/2021</td></tr><tr><td>0.300</td><td>SD1024</td><td>18110</td><td>04/02/2020</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>17817080A2</td><td>08/05/2019</td></tr></tbody></table> ☒ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>SD1021</td><td>201707D</td><td>07/25/2019</td></tr><tr><td>0.080</td><td>DR1275</td><td>201707E</td><td>07/25/2019</td></tr><tr><td>0.200</td><td>SD1011</td><td>201707C</td><td>07/24/2019</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG831804</td><td>11/14/2020</td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.000	DR2035	N/A	N/A	0.040	SD1022	17410	12/06/2019	0.100	SD3964	18200	07/03/2020	0.200	SD1025	19040	01/29/2021	0.300	SD1024	18110	04/02/2020	0.080 DGS	N/A	17817080A2	08/05/2019	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1021	201707D	07/25/2019	0.080	DR1275	201707E	07/25/2019	0.200	SD1011	201707C	07/24/2019	0.080 DGS	N/A	AG831804	11/14/2020	Department Inspection Performed By <u>JD</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1018</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>G11621</td></tr><tr><td>Interferent</td><td>DR3855</td></tr><tr><td>0.050</td><td>SD1021</td></tr><tr><td>0.080</td><td>DR1275</td></tr><tr><td>0.200</td><td>SD1011</td></tr></tbody></table> Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☒ Calibration Adjustment ☐ Other _____ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>SP 4/17/19</u> <u>Butt Kirkland</u> <u>4/18/19</u> Tech Review / Date Admin Review / Date			Simulator	Serial Number	0.000	G11621	Interferent	DR3855	0.050	SD1021	0.080	DR1275	0.200	SD1011
Simulator	Serial Number	Lot Number	Expiration																																																													
0.000	DR2035	N/A	N/A																																																													
0.040	SD1022	17410	12/06/2019																																																													
0.100	SD3964	18200	07/03/2020																																																													
0.200	SD1025	19040	01/29/2021																																																													
0.300	SD1024	18110	04/02/2020																																																													
0.080 DGS	N/A	17817080A2	08/05/2019																																																													
Simulator	Serial Number	Lot Number	Expiration																																																													
0.050	SD1021	201707D	07/25/2019																																																													
0.080	DR1275	201707E	07/25/2019																																																													
0.200	SD1011	201707C	07/24/2019																																																													
0.080 DGS	N/A	AG831804	11/14/2020																																																													
Simulator	Serial Number																																																															
0.000	G11621																																																															
Interferent	DR3855																																																															
0.050	SD1021																																																															
0.080	DR1275																																																															
0.200	SD1011																																																															
Notes/Suggested Service: <u>Please change level 2 password to something unique. JD</u> <u>Calibration adjustment performed due to pressure not being within 1% of pressure gauge during quality checks (Gauge 1015 / Intoxilyzer 998). JD</u>																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 10:06

Date of Inspection: 04/17/2019

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.050	0.079	0.196	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.080	0.198	0.079
0.000	0.048	0.080	0.198	0.079
0.000	0.049	0.080	0.199	0.079

Standard Deviations	0.0007	0.0004	0.0008	0.0000
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP
BK
4/18/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


THOMAS J GRAHAM

 Signature and Printed Name

04/17/2019
 Date

80-001287

4/15/19

JD

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/15/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:48
Control Test	0.049	08:49
Air Blank	0.000	08:49
Control Test	0.049	08:50
Air Blank	0.000	08:50
Control Test	0.049	08:51
Air Blank	0.000	08:52
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/15/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:54
Control Test	0.079	08:55
Air Blank	0.000	08:55
Control Test	0.078	08:56
Air Blank	0.000	08:56
Control Test	0.079	08:57
Air Blank	0.000	08:58
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

JD

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/15/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:59
Control Test	0.191	09:00
Air Blank	0.000	09:01
Control Test	0.196	09:01
Air Blank	0.000	09:02
Control Test	0.195	09:03
Air Blank	0.000	09:03
Control Test Stats		
Average	0.1940	
Std Dev	0.0026	
Rel Std Dev(%)	1.3638	

Out
JD

JD

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/15/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.082	08:44
Air Blank	0.000	08:44
Control Test	0.082	08:45
Air Blank	0.000	08:45
Control Test	0.083	08:46
Air Blank	0.000	08:46
Control Test Stats		
Average	0.0823	
Std Dev	0.0006	
Rel Std Dev(%)	0.7012	

DGS

JD

Operator's Signature

0.20 g/210L is outside of 0.003 or 3%. Instrument requires calibration adjustment due to barometric pressure prior to this reading. JD

SP BK 4/18/19



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001287, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001287</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>BAKER COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/17/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>10:06</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/17/2019 Date
THOMAS J GRAHAM,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

SP BK 4/18/19

Auto Calibration

Max Power Res Value = 31

Auto Range Res Value = 25

Sol Value = 0.000 g/210L ***

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12752, Sum Io = 12380

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.0930 (-0.0100)

Sample #2 = 0.1020 (-0.0150)

Sample #3 = 0.1140 (0.0390)

Sample #4 = 0.0880 (0.0600)

Avg % Abs = 0.1013 (0.0380)

STD DEV = 0.0130 (0.0225)

REL STD DEV = 12.842 (59.254)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1220 (-0.0090)

Sample #2 = 0.1340 (-0.0080)

Sample #3 = 0.1300 (0.0070)

Sample #4 = 0.1170 (0.0170)

Avg % Abs = 0.1270 (0.0053)

STD DEV = 0.0089 (0.0126)

REL STD DEV = 6.999 (235.932)

Sol Value = 0.040 g/210L ***

Fit value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12735, Sum Io = 12375

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8370 (-0.0050)

Sample #2 = 0.8490 (-0.0230)

Sample #3 = 0.8350 (0.0430)

Sample #4 = 0.8450 (0.0650)

Avg % Abs = 0.8430 (0.0437)

STD DEV = 0.0062 (0.0210)

REL STD DEV = 0.741 (48.110)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 1.5490 (-0.0010)

Sample #2 = 1.5720 (0.0030)

Sample #3 = 1.5690 (0.0000)

Sample #4 = 1.5710 (0.0200)

Avg % Abs = 1.5707 (0.0077)

STD DEV = 0.0015 (0.0108)

REL STD DEV = 0.097 (140.684)

Sol Value = 0.100 g/210L ***

Fit value = 0.4762 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12729, Sum Io = 12370

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 1.9160 (0.0000)

Sample #2 = 1.9280 (0.0120)

Sample #3 = 1.9290 (0.0190)

Sample #4 = 1.9140 (0.0470)

Avg % Abs = 1.9237 (0.0260)

STD DEV = 0.0084 (0.0185)

REL STD DEV = 0.436 (71.232)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.6000 (0.0050)

Sample #2 = 3.6340 (0.0110)

Sample #3 = 3.6560 (0.0060)

Sample #4 = 3.6200 (0.0190)

Avg % Abs = 3.6367 (0.0120)

STD DEV = 0.0181 (0.0066)

REL STD DEV = 0.499 (54.645)

Sol Value = 0.200 g/210L ***

Fit value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12723, Sum Io = 12367

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.7230 (-0.0060)

Sample #2 = 3.6590 (0.0270)

Sample #3 = 3.6900 (0.0320)

Sample #4 = 3.6870 (0.0460)

Avg % Abs = 3.6787 (0.0350)

STD DEV = 0.0171 (0.0098)

REL STD DEV = 0.465 (28.140)

80-001287

Calibration
Adjustment

yD

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.101

Std Dev = 0.01 Rel Std Dev = 12.84

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.843

Std Dev = 0.01 Rel Std Dev = 0.74

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.924

Std Dev = 0.01 Rel Std Dev = 0.44

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.679

Std Dev = 0.02 Rel Std Dev = 0.46

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.375

Std Dev = 0.04 Rel Std Dev = 0.69

Zero Order Coef = -263.89

First Order Coef = 2558.04

Second Order Coef = 27.76

Standard Deviation = 4.607818

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.127

Std Dev = 0.01 Rel Std Dev = 7.00

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.571

Std Dev = 0.00 Rel Std Dev = 0.10

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.637

Std Dev = 0.02 Rel Std Dev = 0.50

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.912

Std Dev = 0.03 Rel Std Dev = 0.43

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 10.007

Std Dev = 0.06 Rel Std Dev = 0.58

Zero Order Coef = -171.99

First Order Coef = 1306.14

Second Order Coef = 13.86

Standard Deviation = 5.957672

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0002
0.100	0.100	0.0000
0.200	0.200	0.0000
0.300	0.300	-0.0000

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0002
0.100	0.100	0.0000
0.200	0.200	0.0001
0.300	0.300	-0.0000

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3296.00

Sample #2 = 3320.00

Sample #3 = 3314.00

Sample #4 = 3287.00

Average Result = 3307.0000

STD DEV = 17.5784

REL STD DEV = 0.532

***** CHANNEL 2

Sample #1 = 3333.00

Sample #2 = 3365.00

Sample #3 = 3374.00

Sample #4 = 3319.00

Average Result = 3352.6667

STD DEV = 29.5014

REL STD DEV = 0.880

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1018
 3 um H2O Adjust (mg/l*10,000) = 502
 9 um H2O Adjust (mg/l*10,000) = 457
 ***** AUTO CAL PASS

SP

B3K

4/18/19

80-001287

Post

4/17/19

Stabilities

JD

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001287
04/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:00
Control Test	0.050	08:00
Air Blank	0.000	08:01
Control Test	0.049	08:02
Air Blank	0.000	08:02
Control Test	0.050	08:03
Air Blank	0.000	08:03
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

JD

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001287
04/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:05
Control Test	0.081	08:05
Air Blank	0.000	08:06
Control Test	0.080	08:07
Air Blank	0.000	08:07
Control Test	0.081	08:08
Air Blank	0.000	08:08
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

JD

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001287
04/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:10
Control Test	0.198	08:10
Air Blank	0.000	08:11
Control Test	0.200	08:12
Air Blank	0.000	08:12
Control Test	0.200	08:13
Air Blank	0.000	08:13
Control Test Stats		
Average	0.1983	
Std Dev	0.0012	
Rel Std Dev(%)	0.5793	

JD

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001287
04/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:14
Control Test	0.080	08:15
Air Blank	0.000	08:15
Control Test	0.079	08:16
Air Blank	0.000	08:16
Control Test	0.079	08:16
Air Blank	0.000	08:17
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

SP
BK
4/18/19

JD

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 08:38

Date of Inspection: 04/15/2019

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

NOT A COMPLIANCE CHECK. BYPASSED AI TO OPERATE INSTRUMENT.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

THOMAS J GRAHAM
Signature and Printed Name

04/15/2019
Date

8P BK 4/18/19

April 3, 2019

80-001287
4/17/19
JQ

To Whom It May Concern:

Please find that the Intoxilyzer Machine is having an issue of uploading the agency inspection for the Calendar month of March 2019.

If you have any questions, please give me a call at 904-437-7426.

Sincerely

Lt. Corey Wilson

Baker County Sheriffs Office

RECEIVED

APR 08 2019

FDLE
Alcohol Testing Program

SR
BK
4/18/19