FDLE

Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Pasco County SO S/N 80-003347

Date In 01/22/2019 DI Completion Date 01/22/2019 ☐Ship ☑P/U ☐H/D ☐CMI ☐EE

Intake Performed By JD

☑ Annual
☐ Registration
☐ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☐ Power cord ☐ Printer Cable
☐ Static Bag ☐ 12V DC Cable

Notes:

Quality Checks Performed By JD

☑ Breath Tube Screen
☑ Replace External O-Rings
☑ Instrument Set Up Verified
☑ R-Value 169
☑ Flow Verification (L/s)
Flow Column # ATP-105
32 mm .156 (.139 - .169)
36 mm .175 (.156 - .190)
53 mm .246 (.228 - .278)
103 mm .527 (.447 - .547)
☑ Barometric Pressure Check
Gauge ID # 28421
☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1012	201707D 07/25/2019
0.080	DR1279	201707E 07/25/2019
0.200	DR3856	201707C 07/24/2019
0.080 DGS	N/A	AG805701 02/26/2020

Flow Calibration Performed By

Flow Column #
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value
☐ Post Calibration Verification (L/s)
Flow Column #
32 mm (.139 - .169)
36 mm (.156 - .190)
53 mm (.228 - .278)
103 mm (.447 - .547)

Maintenance Performed By

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other

Temperature Checks Performed By JD

☑ Lab Temp °C 22.0
External Digital Therm. ID#: 300504
☑ 34°C +-.2 Serial #: SD1012
☑ 34°C +-.2 Serial #: DR1279
☑ 34°C +-.2 Serial #: DR3856

Final Release Date

FDLE
JAN 22 2019
Alcohol Testing Program

Calibration Adjustment Performed By

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By JD

Barometric Pressure ID# 26938
Gauge 1028 Instrument 1027
Mouth Alcohol Solution Lot # 2018-B
Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G2408
Interferent	G2882
0.050	SD1012
0.080	DR1279
0.200	DR3856

Attachments

☑ Form 41
☑ Stability Checks
☑ Calibration Certificate
☐ Calibration Adjustment
☐ Post-Stability Checks
☐ Flow Calibration
☐ Form 40
☐ Other

Notes/Suggested Service: Please change level 2 password to something unique.

☑ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
☑ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**
☑ **Conduct an Agency Inspection Before Evidentiary Use**

SP 1/22/19 Brett Kirkland 1/22/19
Tech Review / Date Admin Review / Date

80-003347

1/22/19
JD

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
01/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:50
Control Test	0.049	10:50
Air Blank	0.000	10:51
Control Test	0.048	10:52
Air Blank	0.000	10:52
Control Test	0.048	10:53
Air Blank	0.000	10:53
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

JD
Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
01/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:00
Control Test	0.078	11:01
Air Blank	0.000	11:02
Control Test	0.078	11:02
Air Blank	0.000	11:03
Control Test	0.079	11:04
Air Blank	0.000	11:04
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

JD
Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
01/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:44
Control Test	0.196	10:45
Air Blank	0.000	10:45
Control Test	0.197	10:46
Air Blank	0.000	10:46
Control Test	0.197	10:47
Air Blank	0.000	10:48
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

JD
Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
01/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:39
Control Test	0.081	10:39
Air Blank	0.000	10:39
Control Test	0.081	10:40
Air Blank	0.000	10:40
Control Test	0.081	10:41
Air Blank	0.000	10:41
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP
BK
1/22/19

JD
Operator's Signature

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 13:57
Date of Inspection: 01/22/2019
Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.077	0.194	0.080
0.000	0.048	0.076	0.194	0.080
0.000	0.048	0.077	0.194	0.080
0.000	0.048	0.076	0.194	0.080
0.000	0.047	0.077	0.194	0.080
0.000	0.048	0.077	0.193	0.080
0.000	0.048	0.078	0.193	0.081
0.000	0.048	0.078	0.194	0.080
0.000	0.048	0.078	0.193	0.080
0.000	0.048	0.078	0.193	0.080
Standard Deviations	0.0003	0.0007	0.0005	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP
BK
1/22/19

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

TJ Graham

THOMAS J GRAHAM
Signature and Printed Name

01/22/2019
Date

FDLE/ATP Form 41 –Revised August 2005

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003347 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-003347
Owning Agency:	PASCO COUNTY SO
Calibration Date:	01/22/2019
Calibration Time:	13:57

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/22/2019
Date

THOMAS J GRAHAM,
Department Inspector

SP
BK
1/22/19

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality