




INSTRUMENT PROCESSING SHEET

Agency Dade City PDS/N 80-005477Florida Department of
Law EnforcementDate In 10/3/2019DI Completion Date 10/17/19☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake		Quality Checks		Flow Calibration																																																													
Performed By <u>DP</u>		Performed By <u>SP</u>		Performed By _____																																																													
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>197</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.156</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>SD1012</td><td>201905A 5-14-21</td></tr><tr><td>0.080</td><td>DR1279</td><td>201905B 5-14-21</td></tr><tr><td>0.200</td><td>SD1013</td><td>201904D 4-30-21</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG916501 6-14-21</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201905A 5-14-21	0.080	DR1279	201905B 5-14-21	0.200	SD1013	201904D 4-30-21	0.080 DGS	N/A	AG916501 6-14-21	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	SD1012	201905A 5-14-21																																																															
0.080	DR1279	201905B 5-14-21																																																															
0.200	SD1013	201904D 4-30-21																																																															
0.080 DGS	N/A	AG916501 6-14-21																																																															
Final Release Date <u>FDLE</u> OCT 22 2019 Alcohol Testing Program		Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>SP</u> ☒ Lab Temp °C <u>21.9</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C ±.2 Serial #: <u>SD1012</u> ☒ 34°C ±.2 Serial #: <u>DR1279</u> ☒ 34°C ±.2 Serial #: <u>SD1013</u>																																																															
Calibration Adjustment Performed By _____		Department Inspection Performed By <u>SP</u>																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>26932</u> Gauge <u>1012</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2018-A</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>G2408</td></tr><tr><td>Interferent</td><td>G2882</td></tr><tr><td>0.050</td><td>SD1012</td></tr><tr><td>0.080</td><td>DR1279</td></tr><tr><td>0.200</td><td>SD1013</td></tr></tbody></table>				Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number	Lot Number	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	G2408																																																																
Interferent	G2882																																																																
0.050	SD1012																																																																
0.080	DR1279																																																																
0.200	SD1013																																																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date <u>10/17/19</u> Admin Review / Date <u>Brett Kirkland 10/18/19</u>																																																															



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005477, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005477</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>DADE CITY PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/17/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:23</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/17/2019

Date

Shayla Platt

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Pgm
13K10/18/19

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD

Time of Inspection: 12:23

Date of Inspection: 10/17/2019

Serial Number: 80-005477

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.076	0.202	0.079
0.000	0.049	0.076	0.203	0.079
0.000	0.049	0.075	0.202	0.079
0.000	0.049	0.077	0.203	0.080
0.000	0.050	0.076	0.202	0.079
0.000	0.049	0.076	0.201	0.079
0.000	0.050	0.076	0.201	0.079
0.000	0.051	0.077	0.200	0.079
0.000	0.051	0.076	0.200	0.079

Standard Deviations	0.0007	0.0008	0.0011	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

OGM
1SK
10/18/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT
Signature and Printed Name

10/17/2019
Date

Stability Checks # 80 - 005477

DADE CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005477
10/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:43
Control Test	0.050	09:44
Air Blank	0.000	09:44
Control Test	0.050	09:45
Air Blank	0.000	09:46
Control Test	0.050	09:46
Air Blank	0.000	09:47
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

DADE CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005477
10/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:36
Control Test	0.080	09:36
Air Blank	0.000	09:37
Control Test	0.080	09:38
Air Blank	0.000	09:38
Control Test	0.080	09:39
Air Blank	0.000	09:39
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

DADE CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005477
10/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:26
Control Test	0.199	09:26
Air Blank	0.000	09:27
Control Test	0.199	09:28
Air Blank	0.000	09:28
Control Test	0.200	09:29
Air Blank	0.000	09:29
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

SP

Operator's Signature

DADE CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005477
10/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.081	09:31
Air Blank	0.000	09:31
Control Test	0.082	09:31
Air Blank	0.000	09:32
Control Test	0.081	09:32
Air Blank	0.000	09:33
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

SP

SP

Operator's Signature

SP 10/18/19



INSTRUMENT PROCESSING SHEET

Agency Dade City PDS/N 80-005477Florida Department of
Law EnforcementDate In 04/04/2019 DI Completion Date _____☐ Ship ☐ P/U ☐ H/D ☐ CMI ☒ EE

Intake Performed By <u>KP</u>		Quality Checks Performed By <u>JB</u>		Flow Calibration Performed By _____																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>196</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
Final Release Date FDLE APR 10 2019 Alcohol Testing Program		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>SD1012</td><td>201707D 07/25/2019</td></tr><tr><td>0.080</td><td>DR1279</td><td>201707E 07/25/2019</td></tr><tr><td>0.200</td><td>SD1013</td><td>201707C 07/24/2019</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG805701 02/26/2020</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 07/25/2019	0.080	DR1279	201707E 07/25/2019	0.200	SD1013	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By _____ ☐ Lab Temp °C _____ External Digital Therm. ID#: _____ ☐ 34°C +/- .2 Serial #: _____ ☐ 34°C +/- .2 Serial #: _____ ☐ 34°C +/- .2 Serial #: _____														
Simulator	Serial #	Lot #/Exp																															
0.050	SD1012	201707D 07/25/2019																															
0.080	DR1279	201707E 07/25/2019																															
0.200	SD1013	201707C 07/24/2019																															
0.080 DGS	N/A	AG805701 02/26/2020																															
Calibration Adjustment Performed By _____		Department Inspection Performed By _____																															
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# _____																															
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____			
Simulator	Serial Number	Lot Number	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
☐ Post Calibration Adjustment Stability Checks		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table>				Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200																	
Simulator	Serial Number																																
0.000																																	
Interferent																																	
0.050																																	
0.080																																	
0.200																																	
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☐ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☐ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____											
Simulator	Serial Number	Lot Number	Expiration																														
0.050																																	
0.080																																	
0.200																																	
0.080 DGS	N/A																																
Notes/Suggested Service: <u>Please change level 2 password to something unique. JB</u>		☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																															
The instrument is unable to analyze alcohol reference solutions. No bubbles can be seen in the simulator during the analysis of the samples. Needs to be sent to a repair facility. <u>JB</u>		8P 4/10/19 Tech Review / Date <u>Brett Kirkland 4/10/19</u> Admin Review / Date																															

80-005477

4/10/19

JD

DADE CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005477
04/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:52
Control Test	0.000	07:52
Air Blank	0.000	07:53
Control Test	0.000	07:53
Air Blank	0.000	07:54
Control Test	0.000	07:55
Air Blank	0.000	07:55
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.05
ARS

JD

Operator's Signature

DADE CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005477
04/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:41
Control Test	0.000	07:41
Air Blank	0.000	07:42
Control Test	0.000	07:43
Air Blank	0.000	07:43
Control Test	0.000	07:44
Air Blank	0.000	07:44
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.08
ARS

JD

Operator's Signature

DADE CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005477
04/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:56
Control Test	0.000	07:57
Air Blank	0.000	07:57
Control Test	0.000	07:58
Air Blank	0.000	07:59
Control Test	0.000	07:59
Air Blank	0.000	08:00
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.20
ARS

JD

Operator's Signature

DADE CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005477
04/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:47
Control Test	0.079	07:48
Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:49
Control Test	0.079	07:49
Air Blank	0.000	07:50
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

JD

Operator's Signature

All alcohol reference solutions read 0.000. No bubbles can be seen in the simulator during sample analysis. Probable pump issue.

4/10/19
JD

SP
BK
4/10/19

Return Material Authorization

Ship to:

☐ CMI, Inc.

☒ Enforcement Electronics

Shipment to repair facility authorized by: Dean Baldwin on 04/10/2019

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005477

Bill To Address:

Dade City Police Department

Attn: Dean Baldwin

38042 Pasco Ave.

Dade City, FL 33525

Ship to Address:

FDLE Off-Site Mail Facility

Alcohol Testing Program

813-B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

The instrument will not aerate simulator solutions. All ARS solutions are reading 0.000%. Dry gas is reading accurately.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Dean Baldwin

Phone #: 352-521-1493

Email: dbaldwin@dadecityfl.com

ATP Contact Name: TJ Graham

ATP Email: thomasgraham@fdle.state.fl.us