



INSTRUMENT PROCESSING SHEET

Agency Brevard CountyS/N 80-006233

Florida Department of Law Enforcement

Date In 8/8/2019DI Completion Date 8/13/19 Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>230</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>1148</u> (.139 - .169) 36 mm <u>1167</u> (.156 - .190) 53 mm <u>1230</u> (.228 - .278) 103 mm <u>1472</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date <div style="text-align: center;"> FDLE <u>AUG 15 2019</u> Alcohol Testing Program </div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201905A</u> <u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201905B</u> <u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1013</u></td> <td><u>201904P</u> <u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG916501</u> <u>6-14-21</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	<u>201905A</u> <u>5-14-21</u>	0.080	<u>DR1279</u>	<u>201905B</u> <u>5-14-21</u>	0.200	<u>SD1013</u>	<u>201904P</u> <u>4-30-21</u>	0.080 DGS	N/A	<u>AG916501</u> <u>6-14-21</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.2</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1012</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1013</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD1012</u>	<u>201905A</u> <u>5-14-21</u>																																																											
0.080	<u>DR1279</u>	<u>201905B</u> <u>5-14-21</u>																																																											
0.200	<u>SD1013</u>	<u>201904P</u> <u>4-30-21</u>																																																											
0.080 DGS	N/A	<u>AG916501</u> <u>6-14-21</u>																																																											
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1011</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G8144</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>G2403</u></td> <td><u>19080</u></td> <td><u>3-4-21</u></td> </tr> <tr> <td>0.100</td> <td><u>G2879</u></td> <td><u>18200</u></td> <td><u>7-3-20</u></td> </tr> <tr> <td>0.200</td> <td><u>G3709</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>G8149</u></td> <td><u>18110</u></td> <td><u>4-2-20</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>SP 19817080A2</u></td> <td><u>6-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201905A</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201905B</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1013</u></td> <td><u>SP 20190440</u></td> <td><u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG916501</u></td> <td><u>6-14-21</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G8144</u>	N/A	N/A	0.040	<u>G2403</u>	<u>19080</u>	<u>3-4-21</u>	0.100	<u>G2879</u>	<u>18200</u>	<u>7-3-20</u>	0.200	<u>G3709</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>G8149</u>	<u>18110</u>	<u>4-2-20</u>	0.080 DGS	N/A	<u>SP 19817080A2</u>	<u>6-5-21</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1012</u>	<u>201905A</u>	<u>5-14-21</u>	0.080	<u>DR1279</u>	<u>201905B</u>	<u>5-14-21</u>	0.200	<u>SD1013</u>	<u>SP 20190440</u>	<u>4-30-21</u>	0.080 DGS	N/A	<u>AG916501</u>	<u>6-14-21</u>	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1011</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2408</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2882</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1012</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1013</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2408</u>	Interferent	<u>G2882</u>	0.050	<u>SD1012</u>	0.080	<u>DR1279</u>	0.200	<u>SD1013</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>G8144</u>	N/A	N/A																																																										
0.040	<u>G2403</u>	<u>19080</u>	<u>3-4-21</u>																																																										
0.100	<u>G2879</u>	<u>18200</u>	<u>7-3-20</u>																																																										
0.200	<u>G3709</u>	<u>19040</u>	<u>1-29-21</u>																																																										
0.300	<u>G8149</u>	<u>18110</u>	<u>4-2-20</u>																																																										
0.080 DGS	N/A	<u>SP 19817080A2</u>	<u>6-5-21</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>SD1012</u>	<u>201905A</u>	<u>5-14-21</u>																																																										
0.080	<u>DR1279</u>	<u>201905B</u>	<u>5-14-21</u>																																																										
0.200	<u>SD1013</u>	<u>SP 20190440</u>	<u>4-30-21</u>																																																										
0.080 DGS	N/A	<u>AG916501</u>	<u>6-14-21</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>G2408</u>																																																												
Interferent	<u>G2882</u>																																																												
0.050	<u>SD1012</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>SD1013</u>																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		<u>Program 8/14/19</u> <u>Scott Kirkland 8/14/19</u> Tech Review / Date Admin Review / Date																																																											

 DGS Lot #
 08819080A2

Stability Checks # 80-006233

BREWARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006233
 08/13/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:09
Control Test	0.046	10:10
Air Blank	0.000	10:10
Control Test	0.046	10:11
Air Blank	0.000	10:11
Control Test	0.046	10:12
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0460	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
 Operator's Signature

BREWARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006233
 08/13/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.077	09:59
Air Blank	0.000	10:00
Control Test	0.077	10:01
Air Blank	0.000	10:01
Control Test	0.077	10:02
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
 Operator's Signature

BREWARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006233
 08/13/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.199	10:04
Air Blank	0.000	10:05
Control Test	0.198	10:05
Air Blank	0.000	10:06
Control Test	0.197	10:07
Air Blank	0.000	10:07
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

SP
 Operator's Signature

BREWARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006233
 08/13/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.077	10:15
Air Blank	0.000	10:16
Control Test	0.076	10:16
Air Blank	0.000	10:16
Control Test	0.077	10:17
Air Blank	0.000	10:17
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

DGS
 Operator's Signature

SPAM
 JSK 8/14/19

BERNARD COUNTY S.O.
Intoxilyzer - Alconol Analyzer
Model 8000
08/13/2019

SN 80-005233
11:28:44

Auto Calibration
Max Power Res Value = 83
Auto Range Res Value = 56

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12378, Sum Io = 12668
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0760 (-0.0200)
Sample #2 = 0.0990 (-0.0110)
Sample #3 = 0.0880 (-0.0050)
Sample #4 = 0.0340 (-0.0460)
Aug % Abs = 0.0737 (-0.0133)
STD DEV = 0.0348 (0.0294)
REL STD DEV = 47.226 (220.497)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1110 (-0.0010)
Sample #2 = 0.1340 (-0.0000)
Sample #3 = 0.1720 (-0.0230)
Sample #4 = 0.1360 (-0.0030)
Aug % Abs = 0.1473 (-0.0067)
STD DEV = 0.0214 (0.0142)
REL STD DEV = 14.515 (213.366)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12372, Sum Io = 12666
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8280 (-0.0200)
Sample #2 = 0.8230 (-0.0070)
Sample #3 = 0.7930 (-0.0310)
Sample #4 = 0.7790 (-0.0450)
Aug % Abs = 0.7983 (0.0230)
STD DEV = 0.0225 (0.0269)
REL STD DEV = 2.816 (116.988)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5790 (-0.0190)
Sample #2 = 1.5780 (-0.0150)
Sample #3 = 1.5430 (-0.0040)
Sample #4 = 1.5590 (-0.0070)
Aug % Abs = 1.5600 (-0.0060)
STD DEV = 0.0175 (0.0095)
REL STD DEV = 1.123 (156.990)

<<<<< CHANNEL 1 >>>>>
Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12366, Sum Io = 12667
<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8760 (-0.0060)
Sample #2 = 1.8360 (-0.0230)
Sample #3 = 1.8550 (-0.0130)
Sample #4 = 1.8960 (-0.0040)
Aug % Abs = 1.8623 (-0.0107)
STD DEV = 0.0307 (0.0137)
REL STD DEV = 1.647 (127.972)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6080 (-0.0070)
Sample #2 = 3.6190 (-0.0040)
Sample #3 = 3.6340 (-0.0150)
Sample #4 = 3.6340 (-0.0050)
Aug % Abs = 3.6290 (0.0080)
STD DEV = 0.0087 (0.0051)
REL STD DEV = 0.239 (76.035)

<<<<< CHANNEL 1 >>>>>
Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12364, Sum Io = 12665
<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.4100 (-0.0110)
Sample #2 = 3.4830 (-0.0270)
Sample #3 = 3.5420 (-0.0130)
Sample #4 = 3.5040 (-0.0320)
Aug % Abs = 3.5097 (0.0240)
STD DEV = 0.0299 (0.0198)
REL STD DEV = 0.862 (41.037)

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.074
Std Dev = 0.03 Rel Std Dev = 47.23
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.798
Std Dev = 0.02 Rel Std Dev = 2.82
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.862
Std Dev = 0.03 Rel Std Dev = 1.65
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.510
Std Dev = 0.03 Rel Std Dev = 0.85
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.268
Std Dev = 0.03 Rel Std Dev = 0.59
Zero Order Coef = -249.66
First Order Coef = 2728.09
Second Order Coef = 7.21
Standard Deviation = 79.708847

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.147
Std Dev = 0.02 Rel Std Dev = 14.51
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.560
Std Dev = 0.02 Rel Std Dev = 1.12
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.629
Std Dev = 0.01 Rel Std Dev = 0.24
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.760
Std Dev = 0.03 Rel Std Dev = 0.50
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.966
Std Dev = 0.06 Rel Std Dev = 0.61
Zero Order Coef = -235.42
First Order Coef = 1365.33
Second Order Coef = 9.52
Standard Deviation = 67.043839

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.001 0.0010
0.040 0.041 -0.0006
0.100 0.102 -0.0020
0.200 0.198 0.0023
0.300 0.301 -0.0008

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.001 0.0007
0.040 0.040 -0.0003
0.100 0.102 -0.0017
0.200 0.198 0.0020
0.300 0.301 -0.0006

Sol Value = 0.060 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3188.00
Sample #2 = 3147.00
Sample #3 = 3191.00
Sample #4 = 3102.00
Average Result = 3146.6667
STD DEV = 44.5109
REL STD DEV = 1.414
***** CHANNEL 2 *****
Sample #1 = 3315.00
Sample #2 = 3319.00
Sample #3 = 3295.00
Sample #4 = 3287.00
Average Result = 3300.3333
STD DEV = 16.6533
REL STD DEV = 0.505

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1011
3 um H2O Adjust (mg/l*10,000) = 663
9 um H2O Adjust (mg/l*10,000) = 509
***** AUTO CAL PASS *****

CAL ADJUSTMENT
#80-000233 SP

POAM
Bk 8/14/19

Post Cal-A djust Stability Checks # 80-006233

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006233
 08/13/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:11
Control Test	0.048	12:11
Air Blank	0.000	12:12
Control Test	0.049	12:13
Air Blank	0.000	12:13
Control Test	0.049	12:14
Air Blank	0.000	12:14
Control Test	0.000	12:14
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

SP

 Operator's Signature

OBM

TSK
 8/14/19

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006233
 08/13/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:25
Control Test	0.079	12:25
Air Blank	0.000	12:26
Control Test	0.080	12:26
Air Blank	0.000	12:27
Control Test	0.080	12:28
Air Blank	0.000	12:28
Control Test	0.000	12:28
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP

 Operator's Signature

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006233
 08/13/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:06
Control Test	0.197	12:07
Air Blank	0.000	12:07
Control Test	0.199	12:08
Air Blank	0.000	12:09
Control Test	0.199	12:09
Air Blank	0.000	12:10
Control Test	0.000	12:10
Average	0.1983	
Std Dev	0.0012	
Rel Std Dev(%)	0.5822	

SP

 Operator's Signature

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006233
 08/13/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:18
Control Test	0.080	12:19
Air Blank	0.000	12:19
Control Test	0.080	12:19
Air Blank	0.000	12:20
Control Test	0.081	12:20
Air Blank	0.000	12:21
Control Test	0.000	12:21
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

SP

 Operator's Signature

DES

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 15:35

Date of Inspection: 08/13/2019

Serial Number: 80-006233
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.079	0.199	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.079	0.199	0.079

Standard Deviations	0.0005	0.0004	0.0007	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

JPM
BK
8/14/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

08/13/2019
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006233, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006233</u>	UNCERTAINTY* ±	
Owning Agency:	<u>BREVARD COUNTY S.O.</u>	0.050 g/210 L	0.004
Calibration Date:	<u>08/13/2019</u>	0.080 g/210 L	0.004
Calibration Time:	<u>15:35</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

08/13/2019 _____ Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Platt
SK 8/14/19