



INSTRUMENT PROCESSING SHEET

Agency Palm Beach County SO S/N 80-006478  
Date In 6/11/19 DI Completion Date 6/11/19  Ship  P/U  H/D  CMI  EE

Florida Department of Law Enforcement

<b>Intake</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____	<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>240</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.488</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>U8639</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	---	--

**Final Release Date**

FDLE

JUN 17 2019

Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	SD3967	201707D 7-25-19
0.080	SD3968	201707E 7-25-19
0.200	SD3969	201707C 7-24-19
0.080 DGS	N/A	AG831804 11-14-20

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By SP

Lab Temp °C 21.9  
 External Digital Therm. ID#: 300504  
 34°C +/- .2 Serial #: SD3967  
 34°C +/- .2 Serial #: SD3968  
 34°C +/- .2 Serial #: SD3969

**Calibration Adjustment** Performed By \_\_\_\_\_

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

**Department Inspection** Performed By SP

Barometric Pressure ID# U8639  
 Gauge 1015 Instrument 1014  
 Mouth Alcohol Solution Lot # 2019-B  
 Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	SD3965
Interferent	SD3966
0.050	SD3967
0.080	SD3968
0.200	SD3969

**Attachments**

Form 41  Post-Stability Checks  
 Stability Checks  Flow Calibration  
 Calibration Certificate  Form 40  
 Calibration Adjustment  Other \_\_\_\_\_

Notes/Suggested Service: \_\_\_\_\_

**PROVED**  
6/11/19

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

MX 06/11/19 J. Deha 6/11/19  
 Tech Review / Date Admin Review / Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO  
Time of Inspection: 12:37

Date of Inspection: 06/11/2019

Serial Number: 80-006478  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.048	0.080	0.198	0.080
0.000	0.048	0.080	0.199	0.080
0.000	0.048	0.080	0.199	0.080
0.000	0.048	0.081	0.199	0.080
0.000	0.048	0.080	0.199	0.080
0.000	0.048	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.198	0.080
0.000	0.050	0.081	0.199	0.080

Standard Deviations	0.0007	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

06/11/2019  
Date

*MX*

*1730  
6/11/19  
JD*



<b>TYPE OF TEST</b>	<b>SERIAL NUMBER</b>	<b>AGENCY</b>	<b>DATE</b>	<b>PERFORMED BY</b>
Stabilities	80-006478	Palm Beach County SO	6/11/19	SP

0.05g/210L	0.08g/210L	0.20g/210L	0.08g/210L	DGS 0.08g/210L																																																																																																																																															
PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006478 06/11/2019 Software: 8100.27	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006478 06/11/2019 Software: 8100.27	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006478 06/11/2019 Software: 8100.27	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006478 06/11/2019 Software: 8100.27	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006478 06/11/2019 Software: 8100.27																																																																																																																																															
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:45</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:46</td></tr> <tr><td>Control Test</td><td>0.047</td><td>08:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:47</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:49</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0477</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2112</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:45	Control Test	0.048	08:46	Air Blank	0.000	08:46	Control Test	0.047	08:47	Air Blank	0.000	08:47	Control Test	0.048	08:48	Air Blank	0.000	08:49	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel Std Dev(%)	1.2112		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:51</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:52</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:50	Control Test	0.079	08:51	Air Blank	0.000	08:51	Control Test	0.080	08:52	Air Blank	0.000	08:52	Control Test	0.080	08:53	Air Blank	0.000	08:54	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:40</td></tr> <tr><td>Control Test</td><td>0.197</td><td>08:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:41</td></tr> <tr><td>Control Test</td><td>0.197</td><td>08:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:43</td></tr> <tr><td>Control Test</td><td>0.197</td><td>08:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:44</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1970</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:40	Control Test	0.197	08:41	Air Blank	0.000	08:41	Control Test	0.197	08:42	Air Blank	0.000	08:43	Control Test	0.197	08:43	Air Blank	0.000	08:44	Control Test Stats			Average	0.1970		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:56</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:57</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:57</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:58</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:56	Control Test	0.080	08:56	Air Blank	0.000	08:57	Control Test	0.080	08:57	Air Blank	0.000	08:57	Control Test	0.080	08:58	Air Blank	0.000	08:58	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:45																																																																																																																																																	
Control Test	0.048	08:46																																																																																																																																																	
Air Blank	0.000	08:46																																																																																																																																																	
Control Test	0.047	08:47																																																																																																																																																	
Air Blank	0.000	08:47																																																																																																																																																	
Control Test	0.048	08:48																																																																																																																																																	
Air Blank	0.000	08:49																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2112																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:50																																																																																																																																																	
Control Test	0.079	08:51																																																																																																																																																	
Air Blank	0.000	08:51																																																																																																																																																	
Control Test	0.080	08:52																																																																																																																																																	
Air Blank	0.000	08:52																																																																																																																																																	
Control Test	0.080	08:53																																																																																																																																																	
Air Blank	0.000	08:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:40																																																																																																																																																	
Control Test	0.197	08:41																																																																																																																																																	
Air Blank	0.000	08:41																																																																																																																																																	
Control Test	0.197	08:42																																																																																																																																																	
Air Blank	0.000	08:43																																																																																																																																																	
Control Test	0.197	08:43																																																																																																																																																	
Air Blank	0.000	08:44																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1970																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:56																																																																																																																																																	
Control Test	0.080	08:56																																																																																																																																																	
Air Blank	0.000	08:57																																																																																																																																																	
Control Test	0.080	08:57																																																																																																																																																	
Air Blank	0.000	08:57																																																																																																																																																	
Control Test	0.080	08:58																																																																																																																																																	
Air Blank	0.000	08:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
SP Operator's Signature	SP Operator's Signature	SP Operator's Signature	SP Operator's Signature	SP Operator's Signature																																																																																																																																															

6/12/19  
SP

SP

DGS



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006478, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006478</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PALM BEACH CO SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/11/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:37</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/11/2019

Date

*Shayla D Platt*

SHAYLA D PLATT,  
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*6/17/19*  
*[Signature]*

*[Signature]*