



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Miami

S/N 80-006630

Florida Department of Law Enforcement

Date In 01/07/2019 DI Completion Date 01/07/2019

Ship P/U H/D CMI EE

Intake Performed By DEU Quality Checks Performed By DEU Flow Calibration Performed By DEU
Annual Registration Return from CMI / EE
Visual Inspection: Case Handle Keyboard Dry Gas Shelf Feet Breath Tube Ports Screws Tight
Other Equipment/ Accessories: Power cord Printer Cable Static Bag 12V DC Cable
Notes:
Breath Tube Screen Replace External O-Rings Instrument Set Up Verified R-Value 237
Flow Verification (L/s) Flow Column # ATP 106
32 mm .148 (.139 - .169)
36 mm .171 (.156 - .190)
53 mm .238 (.228 - .278)
103 mm .507 (.447 - .547)
Barometric Pressure Check Gauge ID # 28663
Stability Checks

Table with 3 columns: Simulator, Serial #, Lot #/Exp. Rows include 0.050 SD3967 201707D 07/25/2019, 0.080 SD3968 201707E 07/25/2019, 0.200 SD3969 201707C 07/24/2019, 0.080 DGS N/A AG805701 02/26/2020

Final Release Date
FDLE
JAN 15 2019
Alcohol Testing Program

Maintenance Performed By DEU
Battery Replacement Dry Gas Regulator Replacement Breath Tube Replacement Other Form Load/Change Pass
Temperature Checks Performed By DEU
Lab Temp °C 22.07C
External Digital Therm. ID#: 300918
34°C +/- .2 Serial #: SD3967
34°C +/- .2 Serial #: SD3968
34°C +/- .2 Serial #: SD3969

Calibration Adjustment Performed By DEU
Barometric Pressure Gauge ID #
Simulator Serial Number Lot Number Expiration
0.000 N/A N/A
0.040
0.100
0.200
0.300
0.080 DGS N/A
Post Calibration Adjustment Stability Checks
Simulator Serial Number Lot Number Expiration
0.050
0.080
0.200
0.080 DGS N/A

Department Inspection Performed By DEU
Barometric Pressure ID# 68639
Gauge 1020 Instrument 1020
Mouth Alcohol Solution Lot # 2017-B
Acetone Stock Solution Lot # 2018-A
Simulator Serial Number
0.000 SD3965
Interferent SD3966
0.050 SD3967
0.080 SD3968
0.200 SD3969

Attachments
Form 41 Post-Stability Checks
Stability Checks Flow Calibration
Calibration Certificate Form 40
Calibration Adjustment Other

Notes/Suggested Service: E-mailed [checked] APPROVED

Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use
Tech Review / Date Admin Review / Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP MIAMI

Time of Inspection: 15:21

Date of Inspection: 01/07/2019

Serial Number: 80-006630

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.080	0.197	0.081
0.000	0.048	0.080	0.198	0.081
0.000	0.049	0.080	0.198	0.081
0.000	0.048	0.080	0.198	0.081
0.000	0.048	0.080	0.199	0.081
0.000	0.048	0.080	0.199	0.080
0.000	0.048	0.080	0.199	0.081
0.000	0.048	0.080	0.200	0.081
0.000	0.048	0.080	0.200	0.081
0.000	0.048	0.080	0.200	0.081

Standard Deviations	0.0003	0.0000	0.0010	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

*EDM*

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David E Reyes Rivera* DAVID E REYES-RIVERA  
Signature and Printed Name

01/07/2019  
Date

*1/15/19  
22*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006630	Florida Highway Patrol Miami	01/07/2019	<i>DELL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.07c	SN: SD3968 Temp: 34.07c	SN: SD3969 Temp: 34.07c	Lot AG805701
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

<p>FHP MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 01/07/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:23</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:23	Control Test	0.048	12:24	Air Blank	0.000	12:25	Control Test	0.048	12:25	Air Blank	0.000	12:26	Control Test	0.048	12:27	Air Blank	0.000	12:27	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FHP MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 01/07/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:31</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:29	Control Test	0.079	12:30	Air Blank	0.000	12:30	Control Test	0.079	12:31	Air Blank	0.000	12:31	Control Test	0.079	12:32	Air Blank	0.000	12:33	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FHP MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 01/07/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:34</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:34	Control Test	0.199	12:35	Air Blank	0.000	12:35	Control Test	0.198	12:36	Air Blank	0.000	12:37	Control Test	0.199	12:37	Air Blank	0.000	12:38	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		<p>FHP MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 01/07/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:40</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:41</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:39	Control Test	0.080	12:39	Air Blank	0.000	12:40	Control Test	0.080	12:40	Air Blank	0.000	12:41	Control Test	0.080	12:41	Air Blank	0.000	12:41	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:23																																																																																																																																																	
Control Test	0.048	12:24																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test	0.048	12:25																																																																																																																																																	
Air Blank	0.000	12:26																																																																																																																																																	
Control Test	0.048	12:27																																																																																																																																																	
Air Blank	0.000	12:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:29																																																																																																																																																	
Control Test	0.079	12:30																																																																																																																																																	
Air Blank	0.000	12:30																																																																																																																																																	
Control Test	0.079	12:31																																																																																																																																																	
Air Blank	0.000	12:31																																																																																																																																																	
Control Test	0.079	12:32																																																																																																																																																	
Air Blank	0.000	12:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:34																																																																																																																																																	
Control Test	0.199	12:35																																																																																																																																																	
Air Blank	0.000	12:35																																																																																																																																																	
Control Test	0.198	12:36																																																																																																																																																	
Air Blank	0.000	12:37																																																																																																																																																	
Control Test	0.199	12:37																																																																																																																																																	
Air Blank	0.000	12:38																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2906																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:39																																																																																																																																																	
Control Test	0.080	12:39																																																																																																																																																	
Air Blank	0.000	12:40																																																																																																																																																	
Control Test	0.080	12:40																																																																																																																																																	
Air Blank	0.000	12:41																																																																																																																																																	
Control Test	0.080	12:41																																																																																																																																																	
Air Blank	0.000	12:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

<p><i>DELL</i> Operator's Signature</p>	<p><i>DELL</i> Operator's Signature</p>	<p><i>DELL</i> Operator's Signature</p>	<p><i>DELL</i> Operator's Signature</p>
---	---	---	---

*DELL*

*1/5/19*  
*DELL*



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006630, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006630</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FHP MIAMI</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/07/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:21</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/07/2019

Date

*David Reyes-Rivera*

DAVID E REYES-RIVERA,  
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*UDB*

*1/15/19  
JD*