



INSTRUMENT PROCESSING SHEET

Agency Pensacola PDS/N 80-007159

Florida Department of Law Enforcement

Date In 7/18/2019 DI Completion Date 7/18/19 Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>gjm</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>189</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	--	--

Final Release Date
FDLE
 JUL 18 2019
 Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	5D1018	201707D 7/25/19
0.080	5D3962	201707E 7/25/19
0.200	62078	201707C 7/24/19
0.080 DGS	N/A	AG916501 6/14/21

Maintenance Performed By _____
 Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By gjm
 Lab Temp °C 22.0
 External Digital Therm. ID#: 300503
 34°C +/-2 Serial #: 5D1018
 34°C +/-2 Serial #: 5D3962
 34°C +/-2 Serial #: 62078

Calibration Adjustment Performed By gjm
 Barometric Pressure Gauge 1017 ID # 26932

Simulator	Serial Number	Lot Number	Expiration
0.000	68144	N/A	N/A
0.040	62403	19080	3/4/21
0.100	62879	18200	7/3/20
0.200	5D1019	19040	1/29/21
0.300	68149	18110	4/2/20
0.080 DGS	N/A	17817080R2	8/5/19

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	5D1018	201707D	7/25/19
0.080	5D3962	201707E	7/25/19
0.200	62078	201707C	7/24/19
0.080 DGS	N/A	AG916501	6/14/21

Department Inspection Performed By gjm
 Barometric Pressure ID# 28427
 Gauge 1017 Instrument 1015
 Mouth Alcohol Solution Lot # 2018-B
 Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	64444
Interferent	66621
0.050	5D1018
0.080	5D3962
0.200	62078

Attachments

<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Post-Stability Checks
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration
<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40
<input checked="" type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

gjm 7/18/19 Brett Kunkel 7/18/19
 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PENSACOLA P.D.

Serial Number: 80-007159

Time of Inspection: 12:24

Date of Inspection: 07/18/2019

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.049	0.081	0.198	0.080
0.000	0.050	0.081	0.199	0.079
0.000	0.050	0.081	0.199	0.079
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.079
0.000	0.050	0.081	0.199	0.079
0.000	0.050	0.082	0.199	0.079
0.000	0.050	0.082	0.199	0.079
0.000	0.050	0.082	0.199	0.080
0.000	0.050	0.081	0.200	0.079

Standard Deviations	0.0003	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

SP
BK
7/18/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


PATRICK J MURPHY

 Signature and Printed Name

07/18/2019
 Date

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.045	08:37
Air Blank	0.000	08:38
Control Test	0.045	08:38
Air Blank	0.000	08:39
Control Test	0.045	08:40
Air Blank	0.000	08:40
Control Test Stats		
Average	0.0450	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P. Murphy
Operator's Signature

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:41
Control Test	0.076	08:42
Air Blank	0.000	08:42
Control Test	0.077	08:43
Air Blank	0.000	08:44
Control Test	0.077	08:44
Air Blank	0.000	08:45
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

P. Murphy
Operator's Signature

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:47
Control Test	0.191	08:47
Air Blank	0.000	08:48
Control Test	0.193	08:49
Air Blank	0.000	08:49
Control Test	0.193	08:50
Air Blank	0.000	08:50
Control Test Stats		
Average	0.1923	
Std Dev	0.0012	
Rel Std Dev(%)	0.6004	

P. Murphy
Operator's Signature

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.080	08:52
Air Blank	0.000	08:52
Control Test	0.079	08:53
Air Blank	0.000	08:53
Control Test	0.080	08:53
Air Blank	0.000	08:54
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

P. Murphy
Operator's Signature

SP
BK
7/18/19

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
07/18/2019 09:01:25

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 69

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12563, 9um lo = 12848

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1090 (-0.0060)
Sample #2 = 0.0960 (0.0310)
Sample #3 = 0.1010 (0.0580)
Sample #4 = 0.1150 (0.0600)
Avg % Abs = 0.1040 (0.0497)
STD DEV = 0.0098 (0.0162)
REL STD DEV = 9.470 (32.611)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1210 (-0.0130)
Sample #2 = 0.1060 (0.0020)
Sample #3 = 0.1010 (0.0070)
Sample #4 = 0.1090 (0.0040)
Avg % Abs = 0.1053 (0.0043)
STD DEV = 0.0040 (0.0025)
REL STD DEV = 3.837 (58.076)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12547, 9um lo = 12841

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8670 (-0.0150)
Sample #2 = 0.8800 (0.0060)
Sample #3 = 0.8380 (0.0380)
Sample #4 = 0.8600 (0.0560)
Avg % Abs = 0.8593 (0.0333)
STD DEV = 0.0210 (0.0253)
REL STD DEV = 2.445 (75.974)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.4910 (-0.0180)
Sample #2 = 1.5060 (-0.0040)
Sample #3 = 1.4710 (0.0160)
Sample #4 = 1.4780 (0.0290)
Avg % Abs = 1.4850 (0.0137)
STD DEV = 0.0185 (0.0166)
REL STD DEV = 1.247 (121.634)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12536, 9um lo = 12836

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.9740 (-0.0120)
Sample #2 = 1.9520 (0.0020)
Sample #3 = 1.9520 (0.0180)
Sample #4 = 1.9680 (0.0180)
Avg % Abs = 1.9573 (0.0127)
STD DEV = 0.0092 (0.0092)
REL STD DEV = 0.472 (72.928)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.4850 (-0.0030)
Sample #2 = 3.4800 (-0.0060)
Sample #3 = 3.4760 (0.0000)
Sample #4 = 3.4540 (0.0220)
Avg % Abs = 3.4700 (0.0053)
STD DEV = 0.0140 (0.0147)
REL STD DEV = 0.403 (276.417)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12529, 9um lo = 12828

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.7540 (-0.0110)
Sample #2 = 3.7310 (0.0200)
Sample #3 = 3.7490 (0.0100)
Sample #4 = 3.7890 (0.0100)
Avg % Abs = 3.7563 (0.0133)
STD DEV = 0.0297 (0.0058)
REL STD DEV = 0.790 (43.301)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.6080 (-0.0210)
Sample #2 = 6.6220 (-0.0040)
Sample #3 = 6.6180 (0.0000)
Sample #4 = 6.6300 (0.0000)
Avg % Abs = 6.6233 (-0.0013)
STD DEV = 0.0061 (0.0023)
REL STD DEV = 0.092 (173.205)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12525, 9um lo = 12831

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.4530 (0.0000)
Sample #2 = 5.4850 (-0.0080)
Sample #3 = 5.5270 (-0.0110)
Sample #4 = 5.4620 (0.0310)
Avg % Abs = 5.4913 (0.0040)
STD DEV = 0.0330 (0.0234)
REL STD DEV = 0.600 (585.769)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 9.5710 (0.0070)
Sample #2 = 9.6010 (0.0000)
Sample #3 = 9.6100 (0.0150)
Sample #4 = 9.5890 (0.0200)
Avg % Abs = 9.6000 (0.0117)
STD DEV = 0.0105 (0.0104)
REL STD DEV = 0.110 (89.214)

Auto Cal Data:
Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.104
Std Dev = 0.01 Rel Std Dev = 9.47
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.859
Std Dev = 0.02 Rel Std Dev = 2.44
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.957
Std Dev = 0.01 Rel Std Dev = 0.47
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.756
Std Dev = 0.03 Rel Std Dev = 0.79
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.491
Std Dev = 0.03 Rel Std Dev = 0.60
Zero Order Coef = -265.36
First Order Coef = 2515.71
Second Order Coef = 24.38
Standard Deviation = 7.509191

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.105
Std Dev = 0.00 Rel Std Dev = 3.84
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.485
Std Dev = 0.02 Rel Std Dev = 1.25
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.470
Std Dev = 0.01 Rel Std Dev = 0.40
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.623
Std Dev = 0.01 Rel Std Dev = 0.09
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.600
Std Dev = 0.01 Rel Std Dev = 0.11
Zero Order Coef = -146.59
First Order Coef = 1363.24
Second Order Coef = 14.60
Standard Deviation = 3.215292

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0002
0.100	0.100	0.0002
0.200	0.200	-0.0001
0.300	0.300	0.0000

SPK
7/18/19
SP

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0001
0.100	0.100	0.0000
0.200	0.200	0.0000
0.300	0.300	-0.0000

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 3185.00
 Sample #2 = 3145.00
 Sample #3 = 3166.00
 Sample #4 = 3128.00
 Average Result = 3146.3333
 STD DEV = 19.0351
 REL STD DEV = 0.605

 ***** CHANNEL 2
 Sample #1 = 3434.00
 Sample #2 = 3413.00
 Sample #3 = 3417.00
 Sample #4 = 3415.00
 Average Result = 3415.0000
 STD DEV = 2.0000
 REL STD DEV = 0.059

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1017
 3 um H2O Adjust (mg/l*10,000) = 663
 9 um H2O Adjust (mg/l*10,000) = 394

 **** AUTO CAL PASS

PENSACOLA P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-007159
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:52
Control Test	0.050	09:53
Air Blank	0.000	09:53
Control Test	0.050	09:54
Air Blank	0.000	09:54
Control Test	0.050	09:55
Air Blank	0.000	09:56
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P. Murphy
 Operator's Signature

PENSACOLA P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-007159
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.081	09:58
Air Blank	0.000	09:58
Control Test	0.081	09:59
Air Blank	0.000	10:00
Control Test	0.081	10:00
Air Blank	0.000	10:01
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P. Murphy
 Operator's Signature

PENSACOLA P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-007159
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.079	10:08
Air Blank	0.000	10:08
Control Test	0.079	10:09
Air Blank	0.000	10:09
Control Test	0.079	10:10
Air Blank	0.000	10:10
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P. Murphy
 Operator's Signature

PENSACOLA P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-007159
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:02
Control Test	0.199	10:03
Air Blank	0.000	10:03
Control Test	0.200	10:04
Air Blank	0.000	10:04
Control Test	0.199	10:05
Air Blank	0.000	10:06
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

P. Murphy
 Operator's Signature

80-007159

POST CAL ADJUST
 STABILITIES

SP
 ASK
 7/18/19



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007159, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007159</u>	UNCERTAINTY * ±	
Owning Agency:	<u>PENSACOLA P.D.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/18/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:24</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.
Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/18/2019 Date
Patrick J Murphy
PATRICK J MURPHY,
Department Inspector

SP
BR
7/18/19



INSTRUMENT PROCESSING SHEET

Agency Pensacola PD

S/N 80-007159

Florida Department of Law Enforcement

Date In 03/28/2019 DI Completion Date 4/2/19

Ship P/U H/D CMI EE

Intake Performed By <u>JE</u> <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input checked="" type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>Q9M</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>203</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.246</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>501021</u></td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>A6831804</u> 11/14/20</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>501021</u>	201707D 07/25/2019	0.080	<u>DR1275</u>	201707E 07/25/2019	0.200	<u>SD1011</u>	201707C 07/24/2019	0.080 DGS	N/A	<u>A6831804</u> 11/14/20	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.050	<u>501021</u>	201707D 07/25/2019															
0.080	<u>DR1275</u>	201707E 07/25/2019															
0.200	<u>SD1011</u>	201707C 07/24/2019															
0.080 DGS	N/A	<u>A6831804</u> 11/14/20															

Final Release Date

FDLE

APR 08 2019

Alcohol Testing Program

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By Q9M

Lab Temp °C 21.7
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: 501021
 34°C +/- .2 Serial #: DR1275
 34°C +/- .2 Serial #: SD1011

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By Q9M

Barometric Pressure ID# 30793
 Gauge 1019 Instrument 1017
 Mouth Alcohol Solution Lot # 2018-B
 Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	<u>G11621</u>
Interferent	<u>DR3855</u>
0.050	<u>501021</u>
0.080	<u>DR1275</u>
0.200	<u>SD1011</u>

Attachments

Form 41
 Stability Checks
 Calibration Certificate
 Calibration Adjustment

Post-Stability Checks
 Flow Calibration
 Form 40
 Other form 47

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

JE 4/8/19 Brett Kirkland 4/8/19
 Tech Review / Date Admin Review / Date 4/8/19 BK

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PENSACOLA P.D.
Time of Inspection: 11:09

Date of Inspection: 04/02/2019

Serial Number: 80-007159
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.046	0.077	0.193	0.080
0.000	0.047	0.078	0.196	0.080
0.000	0.047	0.078	0.196	0.080
0.000	0.047	0.079	0.196	0.080
0.000	0.047	0.079	0.196	0.080
0.000	0.047	0.079	0.196	0.080
0.000	0.047	0.079	0.195	0.080
0.000	0.047	0.079	0.196	0.079
0.000	0.047	0.079	0.196	0.080
0.000	0.048	0.079	0.196	0.080

Standard Deviations	0.0004	0.0006	0.0009	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

04/02/2019
Date

*4/8/19
JA
BK*

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
04/02/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:53
Control Test	0.047	07:54
Air Blank	0.000	07:55
Control Test	0.047	07:55
Air Blank	0.000	07:56
Control Test	0.047	07:56
Air Blank	0.000	07:57
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
04/02/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:59
Control Test	0.077	07:59
Air Blank	0.000	08:00
Control Test	0.077	08:01
Air Blank	0.000	08:01
Control Test	0.077	08:02
Air Blank	0.000	08:02
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
04/02/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:08
Control Test	0.196	08:09
Air Blank	0.000	08:10
Control Test	0.195	08:10
Air Blank	0.000	08:11
Control Test	0.196	08:11
Air Blank	0.000	08:12
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	


Operator's Signature

PENSACOLA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007159
04/02/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:13
Control Test	0.080	08:14
Air Blank	0.000	08:14
Control Test	0.080	08:15
Air Blank	0.000	08:15
Control Test	0.079	08:15
Air Blank	0.000	08:16
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS


Operator's Signature

4/8/19
JEB
BK



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007159, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007159</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>PENSACOLA P.D.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/02/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:09</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/02/2019 Date
Patrick J Murphy
PATRICK J MURPHY,
Department Inspector

FDLE/ATP Form 69 July 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

4/8/19
 JQ
 BK



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007159

OWNING AGENCY: Pensacola P.D.

DATE OF DEPARTMENT INSPECTION: 04/02/2019

AGENCY INSPECTOR: James Daniels

ADDRESS: 711 North Hayne Street

CITY, STATE, ZIP: Pensacola, FL 32501

TELEPHONE NUMBER: 850-221-8050

FAX NUMBER: _____

EMAIL ADDRESS (if available): JDaniels@cityofpensacola.com

For Program Office Use Only:

- Registration Issued *BK*
- Instrument Added to Evidentiary Instrument Database
- Instrument Added to Monthly Statistics Database *SP*
- Contact Information Added to Instrument Database

*4/8/19
JO*

Florida Department of Law Enforcement Alcohol Testing Program

REGISTRATION OF EVIDENTIARY BREATH TEST INSTRUMENT

MANUFACTURER: CMI, Inc.
MODEL: Intoxilyzer 8000
SERIAL NUMBER: 80-007159
OWNER: Pensacola Police Department
DATE OF REGISTRATION: April 8, 2019

The above instrument is hereby approved for evidentiary breath alcohol testing in the State of Florida pursuant to Chapter 11D-8, Florida Administrative Code. This instrument and related records are subject to inspection at any time by the Florida Department of Law Enforcement.



Authorized Representative
Alcohol Testing Program
Florida Department of Law Enforcement