



INSTRUMENT PROCESSING SHEET

Agency FFWCCS/N 80-007165Florida Department of
Law EnforcementDate In 8/5/2019DI Completion Date 8/8/19☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Performed By <u>DP</u>	Quality Checks	Performed By <u>DP</u>	Flow Calibration	Performed By <u>DP</u>																																																										
☒ Annual ☒ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☒ 12V DC Cable Notes: <u>2 Power cords in box.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>231</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.246</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE AUG 15 2019 Alcohol Testing Program		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td><u>SD1018</u></td><td><u>201905A</u> <u>5/14/21</u></td></tr><tr><td>0.080</td><td><u>SD3962</u></td><td><u>201905B</u> <u>5/14/21</u></td></tr><tr><td>0.200</td><td><u>G2678</u></td><td><u>201904D</u> <u>4/30/21</u></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td><u>A6916501</u> <u>6/14/21</u></td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1018</u>	<u>201905A</u> <u>5/14/21</u>	0.080	<u>SD3962</u>	<u>201905B</u> <u>5/14/21</u>	0.200	<u>G2678</u>	<u>201904D</u> <u>4/30/21</u>	0.080 DGS	N/A	<u>A6916501</u> <u>6/14/21</u>	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>DP</u> ☒ Lab Temp °C <u>21.3</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>SD1018</u> ☒ 34°C +/- .2 Serial #: <u>SD3962</u> ☒ 34°C +/- .2 Serial #: <u>G2078</u>																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	<u>SD1018</u>	<u>201905A</u> <u>5/14/21</u>																																																													
0.080	<u>SD3962</u>	<u>201905B</u> <u>5/14/21</u>																																																													
0.200	<u>G2678</u>	<u>201904D</u> <u>4/30/21</u>																																																													
0.080 DGS	N/A	<u>A6916501</u> <u>6/14/21</u>																																																													
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>DP</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1013</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td><u>G4444</u></td></tr><tr><td>Interferent</td><td><u>G6621</u></td></tr><tr><td>0.050</td><td><u>SD1018</u></td></tr><tr><td>0.080</td><td><u>SD3962</u></td></tr><tr><td>0.200</td><td><u>G2078</u></td></tr></tbody></table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>SP 8/12/19</u> <u>David R. Ruskland</u> <u>8/12/19</u> Tech Review / Date Admin Review / Date		Simulator	Serial Number	0.000	<u>G4444</u>	Interferent	<u>G6621</u>	0.050	<u>SD1018</u>	0.080	<u>SD3962</u>	0.200	<u>G2078</u>
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	<u>G4444</u>																																																														
Interferent	<u>G6621</u>																																																														
0.050	<u>SD1018</u>																																																														
0.080	<u>SD3962</u>																																																														
0.200	<u>G2078</u>																																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____																																																															

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 12:22

Date of Inspection: 08/08/2019

Serial Number: 80-007165

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.078

Standard Deviations	0.0005	0.0004	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP
BK
8/12/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

08/08/2019
Date

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
08/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.047	09:33
Air Blank	0.000	09:34
Control Test	0.048	09:34
Air Blank	0.000	09:35
Control Test	0.048	09:36
Air Blank	0.000	09:36
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

P. Murphy
Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
08/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:38
Control Test	0.078	09:39
Air Blank	0.000	09:40
Control Test	0.078	09:40
Air Blank	0.000	09:41
Control Test	0.078	09:42
Air Blank	0.000	09:42
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P. Murphy
Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
08/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.196	09:44
Air Blank	0.000	09:45
Control Test	0.197	09:46
Air Blank	0.000	09:46
Control Test	0.199	09:47
Air Blank	0.000	09:47
Control Test Stats		
Average	0.1973	
Std Dev	0.0015	
Rel Std Dev(%)	0.7741	

P. Murphy
Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
08/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.080	09:53
Air Blank	0.000	09:54
Control Test	0.080	09:54
Air Blank	0.000	09:55
Control Test	0.079	09:55
Air Blank	0.000	09:56
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

P. Murphy
Operator's Signature

SP
BK
8/12/19



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007165, manufactured by CML, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007165</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FFWCC</u>	0.050 g/210 L	0.004
Calibration Date:	<u>08/08/2019</u>	0.080 g/210 L	0.004
Calibration Time:	<u>12:22</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/08/2019

Date


PATRICK J MURPHY,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality


8/12/19

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Brett Kirkland on 6/13/19

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: 8000 Serial Number: 80-007165

Bill To Address:

New Instrument

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Department of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Unable to print to an external printer.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: _____

Phone #: _____ Email: _____

ATP Contact Name: Brett Kirkland ATP Email: brettkirkland@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency FFWCCS/N 80-007165Florida Department of
Law EnforcementDate In 6/11/2019 DI Completion Date 6/13/19☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake Performed By <u>DP</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☒ Printer Cable ☒ Static Bag ☒ 12V DC Cable Notes: _____ _____ _____ Final Release Date _____ _____ _____	Quality Checks Performed By <u>BK</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>230</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1018</td> <td>201707D 7/25/19</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> <td>201707E 7/25/19</td> </tr> <tr> <td>0.200</td> <td>G2078</td> <td>201707C 7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804 11/14/20</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1018	201707D 7/25/19	0.080	SD3962	201707E 7/25/19	0.200	G2078	201707C 7/24/19	0.080 DGS	N/A	AG831804 11/14/20	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>BK</u> ☒ Lab Temp °C <u>21.0</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>SD1018</u> ☒ 34°C +/- .2 Serial #: <u>SD3962</u> ☒ 34°C +/- .2 Serial #: <u>G2078</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	SD1018	201707D 7/25/19																																																
0.080	SD3962	201707E 7/25/19																																																
0.200	G2078	201707C 7/24/19																																																
0.080 DGS	N/A	AG831804 11/14/20																																																
Calibration Adjustment Performed By <u>BK</u> Barometric Pressure Gauge <u>1013</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1016</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1022</td> <td>17410</td> <td>12/6/19</td> </tr> <tr> <td>0.100</td> <td>SD3964</td> <td>18200</td> <td>7/3/20</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>19040</td> <td>1/29/21</td> </tr> <tr> <td>0.300</td> <td>SD1024</td> <td>18110</td> <td>4/2/20</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A2</td> <td>8/5/19</td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1018</td> <td>201707D</td> <td>7/25/19</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> <td>201707E</td> <td>7/25/19</td> </tr> <tr> <td>0.200</td> <td>G2078</td> <td>201707C</td> <td>7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804</td> <td>11/14/20</td> </tr> </tbody> </table> Notes/Suggested Service: _____ Unable to print to external printer. Send to CMI. _____ _____ _____ _____			Simulator	Serial Number	Lot Number	Expiration	0.000	SD1016	N/A	N/A	0.040	SD1022	17410	12/6/19	0.100	SD3964	18200	7/3/20	0.200	SD1025	19040	1/29/21	0.300	SD1024	18110	4/2/20	0.080 DGS	N/A	17817080A2	8/5/19	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1018	201707D	7/25/19	0.080	SD3962	201707E	7/25/19	0.200	G2078	201707C	7/24/19	0.080 DGS	N/A	AG831804	11/14/20
Simulator	Serial Number	Lot Number	Expiration																																															
0.000	SD1016	N/A	N/A																																															
0.040	SD1022	17410	12/6/19																																															
0.100	SD3964	18200	7/3/20																																															
0.200	SD1025	19040	1/29/21																																															
0.300	SD1024	18110	4/2/20																																															
0.080 DGS	N/A	17817080A2	8/5/19																																															
Simulator	Serial Number	Lot Number	Expiration																																															
0.050	SD1018	201707D	7/25/19																																															
0.080	SD3962	201707E	7/25/19																																															
0.200	G2078	201707C	7/24/19																																															
0.080 DGS	N/A	AG831804	11/14/20																																															
Department Inspection Performed By <u>BK</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1013</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G4444</td> </tr> <tr> <td>Interferent</td> <td>G6621</td> </tr> <tr> <td>0.050</td> <td>SD1018</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> </tr> <tr> <td>0.200</td> <td>G2078</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment ☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use <u>SP 8/12/19</u> Tech Review / Date <u>Qjm 8/12/19</u> Admin Review / Date			Simulator	Serial Number	0.000	G4444	Interferent	G6621	0.050	SD1018	0.080	SD3962	0.200	G2078																																				
Simulator	Serial Number																																																	
0.000	G4444																																																	
Interferent	G6621																																																	
0.050	SD1018																																																	
0.080	SD3962																																																	
0.200	G2078																																																	



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007165, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007165</u>	UNCERTAINTY * $\pm$
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/13/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:38</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/13/2019

Date

Brett Kirkland
BRETT H KIRKLAND,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Serial Number: 80-007165

Time of Inspection: 11:38

Date of Inspection: 06/13/2019

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#: 201707E Exp: 07/27/2019	0.20g/210L Test (g/210L) Lot#: 201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG831804 Exp: 11/14/2020
0.000	0.048	0.078	0.197	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079

Standard Deviations	0.0003	0.0006	0.0009	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005

Number of Simulators Used: 5

Remarks:

A F / M A: DID NOT USE ENOUGH MA

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brett Kirkland

BRETT H KIRKLAND

Signature and Printed Name

06/13/2019

Date

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
06/12/2019
Software: 8100.27

Test g/210L Time

Air Blank 0.000 11:45
Control Test 0.046 11:46
Air Blank 0.000 11:46
Control Test 0.047 11:47
Air Blank 0.000 11:47
Control Test 0.047 11:48
Air Blank 0.000 11:49
Control Test Stats
Average 0.0467
Std Dev 0.0006
Rel Std Dev(%) 1.2372

Possible cold tubing - retest

Operator's Signature

YBK

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
06/12/2019
Software: 8100.27

Test g/210L Time

Air Blank 0.000 11:50
Control Test 0.047 11:51
Air Blank 0.000 11:51
Control Test 0.047 11:52
Air Blank 0.000 11:52
Control Test 0.047 11:53
Air Blank 0.000 11:54
Control Test Stats
Average 0.0470
Std Dev 0.0000
Rel Std Dev(%) 0.0000

Operator's Signature

YBK

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
06/12/2019
Software: 8100.27

Test g/210L Time

Air Blank 0.000 11:55
Control Test 0.077 11:55
Air Blank 0.000 11:56
Control Test 0.078 11:57
Air Blank 0.000 11:57
Control Test 0.079 11:58
Air Blank 0.000 11:58
Control Test Stats
Average 0.0780
Std Dev 0.0010
Rel Std Dev(%) 1.2821

Operator's Signature

YBK

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
06/12/2019
Software: 8100.27

Test g/210L Time

Air Blank 0.000 11:59
Control Test 0.193 12:00
Air Blank 0.000 12:01
Control Test 0.195 12:02
Air Blank 0.000 12:03
Control Test 0.197 12:03
Air Blank 0.000 12:03
Control Test Stats
Average 0.1950
Std Dev 0.0020
Rel Std Dev(%) 1.0256

Operator's Signature

YBK

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
06/12/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:05
Control Test	0.079	12:05
Air Blank	0.000	12:06
Control Test	0.080	12:06
Air Blank	0.000	12:07
Control Test	0.079	12:07
Air Blank	0.000	12:07
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

YBK
Operator's Signature

Stability Check
80-007165

YBK
6/12/19

Sp

Pgan

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007165
06/13/2019 08:18:29

Auto Calibration
Max Power Res Value = 69
Auto Range Res Value = 43

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12595, 9um lo = 12969

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.1340 (0.0000)
Sample #2 = 0.2060 (0.0240)
Sample #3 = 0.1630 (0.0800)
Sample #4 = 0.1610 (0.1160)
Avg % Abs = 0.1767 (0.0733)
STD DEV = 0.0254 (0.0464)
REL STD DEV = 14.390 (63.219)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1750 (-0.0040)
Sample #2 = 0.2150 (0.0000)
Sample #3 = 0.1890 (0.0230)
Sample #4 = 0.1930 (0.0300)
Avg % Abs = 0.1990 (0.0177)
STD DEV = 0.0140 (0.0157)
REL STD DEV = 7.035 (88.840)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12571, 9um lo = 12961
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.8960 (-0.0140)
Sample #2 = 0.9060 (0.0080)
Sample #3 = 0.8600 (0.0540)
Sample #4 = 0.8520 (0.0780)
Avg % Abs = 0.8727 (0.0467)
STD DEV = 0.0291 (0.0356)
REL STD DEV = 3.340 (76.225)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5390 (-0.0020)
Sample #2 = 1.5730 (-0.0070)
Sample #3 = 1.5300 (0.0200)
Sample #4 = 1.5430 (0.0230)
Avg % Abs = 1.5487 (0.0120)
STD DEV = 0.0221 (0.0165)
REL STD DEV = 1.424 (137.689)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12559, 9um lo = 12956
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.9260 (-0.0160)
Sample #2 = 1.8920 (0.0390)
Sample #3 = 1.9060 (0.0520)
Sample #4 = 1.9080 (0.0630)
Avg % Abs = 1.9020 (0.0513)
STD DEV = 0.0087 (0.0120)
REL STD DEV = 0.458 (23.404)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.4880 (-0.0150)
Sample #2 = 3.4910 (0.0050)
Sample #3 = 3.4970 (0.0140)
Sample #4 = 3.5060 (0.0180)
Avg % Abs = 3.4980 (0.0123)
STD DEV = 0.0075 (0.0067)
REL STD DEV = 0.216 (53.986)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12547, 9um lo = 12951
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.6120 (-0.0020)
Sample #2 = 3.6130 (0.0260)
Sample #3 = 3.6840 (0.0030)
Sample #4 = 3.6600 (0.0460)
Avg % Abs = 3.6523 (0.0250)
STD DEV = 0.0361 (0.0215)
REL STD DEV = 0.989 (86.070)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 6.6120 (-0.0030)
Sample #2 = 6.6140 (0.0340)
Sample #3 = 6.6880 (0.0080)
Sample #4 = 6.6630 (0.0390)
Avg % Abs = 6.6550 (0.0270)
STD DEV = 0.0376 (0.0166)
REL STD DEV = 0.566 (61.642)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12540, 9um lo = 12947

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.2350 (-0.0120)
Sample #2 = 5.2640 (0.0070)
Sample #3 = 5.2260 (0.0390)
Sample #4 = 5.2840 (0.0190)
Avg % Abs = 5.2580 (0.0217)
STD DEV = 0.0295 (0.0162)
REL STD DEV = 0.560 (74.611)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 9.5410 (-0.0130)
Sample #2 = 9.5480 (0.0290)
Sample #3 = 9.5490 (0.0380)
Sample #4 = 9.5950 (0.0210)
Avg % Abs = 9.5640 (0.0293)
STD DEV = 0.0269 (0.0085)
REL STD DEV = 0.281 (28.994)

***** AUTO CAL DATA *****

Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.177
Std Dev = 0.03 Rel Std Dev = 14.39
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.873
Std Dev = 0.03 Rel Std Dev = 3.34
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.902
Std Dev = 0.01 Rel Std Dev = 0.46
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.652
Std Dev = 0.04 Rel Std Dev = 0.99
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.258
Std Dev = 0.03 Rel Std Dev = 0.56
Zero Order Coef = -436.34
First Order Coef = 2642.58
Second Order Coef = 28.91
Standard Deviation = 55.766502

Channel 2
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.199
Std Dev = 0.01 Rel Std Dev = 7.04
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.549
Std Dev = 0.02 Rel Std Dev = 1.42
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.498
Std Dev = 0.01 Rel Std Dev = 0.22
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.655
Std Dev = 0.04 Rel Std Dev = 0.57
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.564
Std Dev = 0.03 Rel Std Dev = 0.28
Zero Order Coef = -262.80
First Order Coef = 1374.53
Second Order Coef = 15.19
Standard Deviation = 24.396532

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0007
0.040 0.040 0.0003
0.100 0.099 0.0014
0.200 0.202 -0.0016
0.300 0.299 0.0006

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0001
0.100 0.099 0.0006
0.200 0.201 -0.0007
0.300 0.300 0.0003

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
Channel 1
Sample #1 = 2895.00
Sample #2 = 2785.00
Sample #3 = 2871.00
Sample #4 = 2856.00
Average Result = 2837.3333
STD DEV = 45.9384
REL STD DEV = 1.619

Channel 2
Sample #1 = 3277.00
Sample #2 = 3237.00
Sample #3 = 3252.00
Sample #4 = 3252.00
Average Result = 3247.0000
STD DEV = 8.6603
REL STD DEV = 0.267

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1013
3 um H2O Adjust (mg/l*10,000) = 972
9 um H2O Adjust (mg/l*10,000) = 562
**** AUTO CAL PASS

Optical BenchCa
80-007165

BK
6/13/19
SP

PJAM

FFWCC

Intoxilyzer - Alcohol Analyzer

Model 8000 SN 80-007165

06/13/2019

Software: 8100.27

Test g/210L Time

Air Blank	0.000	09:06
Control Test	0.049	09:07
Air Blank	0.000	09:08
Control Test	0.048	09:08
Air Blank	0.000	09:09
Control Test	0.049	09:09
Air Blank	0.000	09:10
Control Test	0.000	09:10
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

FFWCC

Intoxilyzer - Alcohol Analyzer

Model 8000 SN 80-007165

06/13/2019

Software: 8100.27

Test g/210L Time

Air Blank	0.000	09:11
Control Test	0.079	09:12
Air Blank	0.000	09:12
Control Test	0.080	09:13
Air Blank	0.000	09:13
Control Test	0.079	09:14
Air Blank	0.000	09:15
Control Test	0.000	09:15
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

FFWCC

Intoxilyzer - Alcohol Analyzer

Model 8000 SN 80-007165

06/13/2019

Software: 8100.27

Test g/210L Time

Air Blank	0.000	09:16
Control Test	0.197	09:16
Air Blank	0.000	09:17
Control Test	0.197	09:18
Air Blank	0.000	09:18
Control Test	0.199	09:19
Air Blank	0.000	09:19
Control Test	0.000	09:19
Average	0.1977	
Std Dev	0.0012	
Rel Std Dev(%)	0.5842	

FFWCC

Intoxilyzer - Alcohol Analyzer

Model 8000 SN 80-007165

06/13/2019

Software: 8100.27

Test g/210L Time

Air Blank	0.000	09:25
Control Test	0.079	09:26
Air Blank	0.000	09:26
Control Test	0.080	09:26
Air Blank	0.000	09:27
Control Test	0.080	09:27
Air Blank	0.000	09:28
Control Test	0.000	09:28
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature

FFWCC

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-007165

06/13/2019

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:03
Control Test	INT*	09:04
Air Blank	0.000	09:04
Control Test	INT*	09:05
Air Blank	0.000	09:05
Control Test	0.000	09:05
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*Interferent Detect

QC
check
of Int. Sim

Operator's Signature

Post-stabilities
80-007165

JK

6/13/19

0800

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Brett Kirkland on 5/10/2019

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: 8000 Serial Number: 80-007165

Bill To Address:

New Instrument

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Department Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Diagnostic fail upon startup (DSP). Unable to get instrument out of Standby Mode.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: _____

Phone #: _____ Email: _____

ATP Contact Name: Brett Kirkland ATP Email: brettkirkland@fdle.state.fl.us

Pjm



INSTRUMENT PROCESSING SHEET

Agency FWCCS/N 80-007165Florida Department of
Law EnforcementDate In 5/1/19

DI Completion Date _____

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Performed By <u>DP</u>	Quality Checks	Performed By <u>BK</u>	Flow Calibration	Performed By <u>BK</u>																																							
☒ Annual ☒ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☒ Printer Cable ☒ Static Bag ☒ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>212</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>.167</u> (.139 - .169) 36 mm <u>.203</u> (.156 - .190) 53 mm <u>.269</u> (.228 - .278) 103 mm <u>.546</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☐ Stability Checks		Flow Column # <u>ATP 103</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>228</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547)																																								
Final Release Date		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A			Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																									
Simulator	Serial #	Lot #/Exp																																										
0.050																																												
0.080																																												
0.200																																												
0.080 DGS	N/A																																											
				Temperature Checks Performed By _____ ☐ Lab Temp °C _____ External Digital Therm. ID#: _____ ☐ 34°C +/- .2 Serial #: _____ ☐ 34°C +/- .2 Serial #: _____ ☐ 34°C +/- .2 Serial #: _____																																								
Calibration Adjustment	Performed By _____	Department Inspection Performed By _____																																										
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table>			Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial Number	Lot Number	Expiration																																									
0.000		N/A	N/A																																									
0.040																																												
0.100																																												
0.200																																												
0.300																																												
0.080 DGS	N/A																																											
Simulator	Serial Number																																											
0.000																																												
Interferent																																												
0.050																																												
0.080																																												
0.200																																												
☐ Post Calibration Adjustment Stability Checks		Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>	Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				<table border="1"><tbody><tr><td>☐ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☐ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☐ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>			☐ Form 41	☐ Post-Stability Checks	☐ Stability Checks	☐ Flow Calibration	☐ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____												
Simulator	Serial Number	Lot Number	Expiration																																									
0.050																																												
0.080																																												
0.200																																												
0.080 DGS	N/A																																											
☐ Form 41	☐ Post-Stability Checks																																											
☐ Stability Checks	☐ Flow Calibration																																											
☐ Calibration Certificate	☐ Form 40																																											
☐ Calibration Adjustment	☐ Other _____																																											
Notes/Suggested Service: _____ <u>Diagnostic fail on startup--DSP</u> <u>Unable to bring out of standby mode</u> <u>- Send to repair - CMI BK</u> _____ _____ _____		☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																										
		Tech Review / Date _____ Admin Review / Date _____																																										



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007165

OWNING AGENCY: FFWCC

DATE OF DEPARTMENT INSPECTION: 8/8/2019

AGENCY INSPECTOR: Nathan Brock

ADDRESS: FWC Regional Office, 8535 Northlake Blvd

CITY, STATE, ZIP: West Palm Beach, FL 33412

TELEPHONE NUMBER: 954-702-9909

FAX NUMBER: _____

EMAIL ADDRESS (if available): nathan.brock@myFWC.com

For Program Office Use Only:

- ☒ Registration Issued *13K*
- ☒ Instrument Added to Evidentiary Instrument Database *SP*
- ☒ Instrument Added to Monthly Statistics Database *SP*
- ☒ Contact Information Added to Instrument Database *SP*

Florida Department of Law Enforcement Alcohol Testing Program

REGISTRATION OF EVIDENTIARY BREATH TEST INSTRUMENT

MANUFACTURER: CMI, Inc.
MODEL: Intoxilyzer 8000
SERIAL NUMBER: 80-007165
OWNER: FFWCC
DATE OF REGISTRATION: August 12, 2019

The above instrument is hereby approved for evidentiary breath alcohol testing in the State of Florida pursuant to Chapter 11D-8, Florida Administrative Code. This instrument and related records are subject to inspection at any time by the Florida Department of Law Enforcement.



Authorized Representative
Alcohol Testing Program
Florida Department of Law Enforcement