

Kirkland, Brett

From: Rath, Robert <rrath@marionso.com>
Sent: Friday, August 16, 2019 8:03 AM
To: Kirkland, Brett
Subject: Re: Marion county Intoxilyzer deregistration

There have not been, we just got the instrument back last month and conducted an agency inspection upon return.

Sent from my iPhone

On Aug 16, 2019, at 7:49 AM, Kirkland, Brett <BrettKirkland@fdle.state.fl.us> wrote:

Corporal Rath,

Have there been any breath tests performed on this instrument since 7/22/19? If yes, then you will need to complete another agency inspection before I can deregister the instrument.

Thank you,

Brett

From: Rath, Robert [<mailto:rrath@marionso.com>]
Sent: Thursday, August 15, 2019 2:09 PM
To: Kirkland, Brett
Subject: Marion county Intoxilyzer deregistration

Good afternoon, per our phone conversation you will find the most recent monthly inspection for the instrument we would like to be deregistered. We recently obtained two new instruments and wish to turn this one into a Trainer. Please let me know if anything further is needed. Thank you, Rob

Corporal R. Rath 5536
Marion County Sheriff's Office
352-572-1781 (cell)
352-732-8181 (office)
rrath@marionso.com

Under Florida Law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing. CONFIDENTIALITY NOTICE: This message and any attachments are for the sole use of the intended recipient(s) and may contain confidential and privileged information that is exempt from public disclosure. Any unauthorized review, use, disclosure, or distribution is prohibited. If you have received this message in error please contact the sender (by phone or reply electronic mail).

Under Florida Law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing. CONFIDENTIALITY NOTICE: This message and any attachments are for the sole use of

80-000752
SOL 1/18/19

752

- BREATH TUBE NOT HEATING UP. Temp REG FAIL DURING DIAGNOSTIC CHECK.
- UPLOADED THE SUBJECT BLOW REPORTS STORED. UNABLE TO CONDUCT MONTHLY INSPECTION.

LT. JEFFERY RAKER
(352) 843-4215

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SHERIFF'S OFFICE

Serial Number: 80-000752

Time of Inspection:

Date of Inspection: JANUARY 2019

Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

UNABLE TO CONDUCT MONTHLY INSPECTION DUE TO TEMP. REG. FAIL.
INSTRUMENT TAKEN OUT OF SERVICE AND SENT TO FOLE.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JEFFREY RAKER
Signature and Printed Name

1/14/2019
Date

Kirkland, Brett

From: Raker, Jeff <jraker@marionso.com>
Sent: Saturday, January 26, 2019 5:44 AM
To: Kirkland, Brett
Subject: RE: Alcohol Testing Program

Ok, you can send it to CMI, Inc. for repair. I already ordered a breath tube for it. I'll see if I can cancel the order. If you need me the best way to get me is to call my cell phone. 352-843-4215. We work alternating shifts and days; however, I am always available by cell phone, call me anytime. Thanks

From: Kirkland, Brett <BrettKirkland@fdle.state.fl.us>
Sent: Wednesday, January 23, 2019 11:59 AM
To: Raker, Jeff <jraker@marionso.com>
Subject: Alcohol Testing Program

Lt. Raker,

Would you please give me a call. I have an update about instrument 80-000752. It will also need to be sent to repair. The breath tube seems to be fine but still not heating up. 850-617-1275

Regards

Brett Kirkland, Ph.D.

Program Manager
Alcohol Testing Program
Florida Department of Law Enforcement
Office: (850) 617-1275
Cell: (850) 631-1351

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

Under Florida Law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing. CONFIDENTIALITY NOTICE: This message and any attachments are for the sole use of the intended recipient(s) and may contain confidential and privileged information that is exempt from public disclosure. Any unauthorized review, use, disclosure, or distribution is prohibited. If you have received this message in error please contact the sender (by phone or reply electronic mail).