



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Wakulla County S.O. Instrument Serial Number: 80-000962

Department Inspection, Agency Inspection/Agency Contact, Records Audit. Includes checkboxes for Registration, Agency Inspection Notification, and Barometric Pressure.

Comments: Agency Inspector reported a failed inspection, attributing that to an inoperative pump, as he noticed that the simulators were not being aerated. He is bringing the instrument in for our appraisal.

Instrument/Area, Equipment, Supplies. Includes checkboxes for Clean/Dry, Proper Number of Simulators, and Distilled/Deionized Water.

Comments:

Action: Instrument Complies with Chapter 11D-8, FAC; Instrument Does Not Comply with Chapter 11D-8, FAC; Remain in Evidentiary Use; Return to/Place into Evidentiary Use; Remove from Evidentiary Use; Remain Out of Evidentiary Use; Conduct an Agency Inspection Before Evidentiary Use.

Signature of Alcohol Testing Program Staff Member (Patrick Murphy)

Date (3/22/19)

Original - FDLE Copy - Agency