

Graham, Thomas

From: Lawrence Kolbicka <LKolbicka@pinellas-park.com>
Sent: Thursday, April 04, 2019 3:36 PM
To: Graham, Thomas
Subject: PPPD Intox Issues

Follow Up Flag: FollowUp
Flag Status: Flagged

Mr Graham,

I completed both Pinellas Park Police Department agency inspections today and had two issues.

1 - instrument 1042. I did not get a interferent detect on my interferent sim. I checked and the top was on loosely. I tightened it and had no further insuses.

2 - Instrument 1732. I got an ambient fail on the first dry gas standard. I found that I did not connect the hose to the instrument from the dry gas. I connected the hose and had no further issues.

Thank You,

Larry

--

OFC Lawrence R Kolbicka II
Special Operations Unit
Traffic Homicide Investigator
Traffic Enforcement Officer
Agency Inspector

Pinellas Park Police Department