



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

July 04, 2019

Mrs. Shayla Platt
Florida Department of Law Enforcement
Alcohol Testing Program
PO Box 1489
Tallahassee, FL 32302

REF: July 04, 2019 Inspection

Dear Mrs. Platt:

I am mailing a copy of the Agency Inspection Reports for the month of July 04, 2019. Instrument 80-001054 and 80-001043 have passed inspection.

However for 1054 during the .020 test at the end I accidentally hit the enter instead of the escape to indicate I had an error which of course I did not. So I entered 00 as the reason for the error and aborted the entire test. I performed a second inspection with only 1 minor problem. During the mouth alcohol part I had to much solution in my mouth and gave an improper sample message. Cleared/rinsed my mouth and repeated that part with no issue.

Sgt. Glasco
Agency Inspector

Cc: File