

Graham, Thomas

From: Alcohol Testing Program
Sent: Friday, April 26, 2019 10:29 AM
To: Graham, Thomas
Subject: FW: Intoxilyzer 8000

80-001060
4/26/19
JA

FYI

From: Bret Hebson [mailto:bhebson@mywinterhaven.com]
Sent: Friday, April 26, 2019 10:28 AM
To: Alcohol Testing Program
Subject: Intoxilyzer 8000

To whom it may concern,

I attempted to conduct my monthly inspection of our agency Intoxilyzer and found no power to the unit. I checked the plug and fuse and both are in serviceable order. There is power to the plug as well. At this point I am unable to conduct the inspection. I have boxed the unit up to be sent in for repair and the annual inspection as well (if possible). I can be reached at the below numbers or by email.

B. Hebson, Officer #521
Motors/ Traffic Unit
Winter Haven Police Department
125 North Lake Silver Dr NW
Winter Haven, Fl. 33881
Main (863) 291-5858 Ext. 2333
Cell (863) 514-8609
Office (863) 291-5334

Please note:

This email may contain legally privileged and confidential information intended only for the individual or entity named within the message. If the reader of this message is not the intended recipient or the agent responsible to deliver it to the intended recipient, you are hereby notified that any review, dissemination, distribution or copying of this communication is prohibited. If this email was received in error, please notify us and destroy the original email and any attachments.

“Don’t let schooling interfere with your education” – Mark Twain