

Graham, Thomas

From: Graham, Thomas
Sent: Wednesday, March 27, 2019 7:05 AM
To: Platt, Shayla; Jason Engel
Cc: Petersen, Karalyn; Smith, Dorothy
Subject: RE: New Port Richey Intox Questions
Attachments: 80-001064 Agency Inspection 3-26-2019.pdf

80-001064

3/27/19
JE

Sgt. Engel,

I have attached a copy of the Agency Inspection for instrument #80-001064 performed on 3/26/2019. If there is anything further I can assist you with, please do not hesitate to contact me.

TJ Graham
(850) 445-7938 (cell)

From: Platt, Shayla
Sent: Tuesday, March 26, 2019 11:24 PM
To: Jason Engel
Cc: Graham, Thomas
Subject: Re: New Port Richey Intox Questions

Good Evening,

Thank you for sending the report with the notes. Currently Thomas Graham who is our quality manager is temporarily handling the department inspector duties for your region. I will forward a copy of the inspection in the morning. Once the instrument is uploaded after it clears the data stored on it.

Shayla

Sent from my Verizon, Samsung Galaxy smartphone

----- Original message -----

From: Jason Engel <EngelJ@CityofNewPortRichey.org>
Date: 3/26/19 11:13 PM (GMT-05:00)
To: "Platt, Shayla" <ShaylaPlatt@fdle.state.fl.us>
Subject: New Port Richey Intox Questions

Hello,

I am not sure who to send the agency inspection reports to now with notes when there is an issue. I did the inspection and had no results for the .08 solution. After running it twice I inspected the jar and realized the jar cracked and broke along the top threaded area not allowing a seal. I have attached the failed inspection with hand written notes. I was unable to print the second inspection after replacing the jar and everything going fine due to a printer error. After I reset the printer I was unable to recall the inspection. Is there a way to do that as when I did recall test the instrument said nothing to recall. If you could email me a copy that would be great as it did appear to upload fine. Thanks for the help!!

Regards,

Jason Engel
Sergeant/Patrol Division
New Port Richey Police
6739 Adams Street
New Port Richey, FL 34652
727-841-4550 ext 124
fax 727-816-1132

This correspondence is from a law enforcement agency and may contain confidential, secret, classified, and/or privileged communications. Unauthorized use of this information may subject you to criminal and civil penalties. This document is the property of the New Port Richey Police Department and may not be disseminated without authorization. As such this document may be exempt from public record as defined in Florida State Statute Chapter 119. It is the responsibility of the recipient of this correspondence to ensure applicable Laws, Statutes, Rules and Regulations are followed.

******* NOTICE: This electronic transmission is confidential and intended only for the person(s) to whom it is addressed. If you have received this transmission in error, please notify the sender by return e-mail and destroy this message in its entirety (including all attachments). *******