

From: [Graham, Thomas](#)
To: [Hernandez-Ginarte, Alain](#)
Subject: Re: INTOX 80-006638
Date: Wednesday, February 27, 2019 9:14:18 PM

Trooper Hernandez,

Thank you for reporting the issue. I will evaluate the instrument once it arrives in Tallahassee.

TJ Graham, D-ABFT-FA
Quality Assurance Manager
Alcohol Testing Program
Florida Department of Law Enforcement
Office: (850) 617-1285
Cell: (850) 445-7938

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

Sent from my Verizon, Samsung Galaxy smartphone

----- Original message -----

From: "Hernandez-Ginarte, Alain" <AlainHernandez-Ginarte@flhsmv.gov>
Date: 2/27/19 8:10 PM (GMT-05:00)
To: "Graham, Thomas" <ThomasGraham@fdle.state.fl.us>
Subject: INTOX 80-006638

Good Evening Sir,

As I was doing the inspections for our intoxes, Intox **80-006638** gave me "Ambient Fail" on the first two (2) samples for the Alcohol Free Test. On the "Repeat" portion of it, it gave me yet again another "Ambient Fail" and the inspection was terminated.

I then moved the intox to a different location and re-attempted the inspection test. As it was going thru the alcohol free test, it gave me "Ambient Fail" and two (2) different attempts. I discontinued any further attempts to do a third inspection test.

This intox was in the same place as my other seven (7) intoxes and no issues were given to me by those other 7 intoxes.

Lastly, I did send this same intox to FDLE on 09/27/2018 for the same issue, however, nothing wrong was found. I am sending it up back to you so we can figure out if something internally is wrong with it.

Thank you very much.

Trooper Alain Hernandez
Florida Highway Patrol
Troop C- Special Operations District / DUI Enforcement

Drug Recognition Expert – IACP#26438

DUI Instructor

11305 North Mckinley Drive

Tampa, FL. 33612

Office: 813.558.1800

AlainHernandez-Ginarte@flhsmv.gov

<http://www.flhsmv.gov/fhp>