

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
Time of Inspection: 18:03

Date of Inspection: 11/29/2020

Serial Number: 80-000770
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A3 Exp: 03/05/2021
0.000	0.047	0.077	0.197	0.074 / 0.076
0.000	0.046	0.077	0.198	0.072 / 0.076
0.000	0.047	0.077	0.198	0.074 / 0.076

Number of Simulators Used: 4

Amended

Remarks:

[Interf.] TURNED CELL PHONE OFF } DRY GAS DISC AND REC HOSES DRY GAS ~~W/DBS CON AND RECON DRY GAS REPEAT TEST~~

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

William C Curry

WILLIAM C CURRY

Signature and Printed Name

11/29/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
Time of Inspection: 17:48

Date of Inspection: 07/30/2020

Serial Number: 80-000770
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:N Exp: 03/05/2021
0.000	0.046	0.077	0.196	0.076
0.000	0.046	0.077	0.196	0.076
0.000	0.046	0.077	0.195	0.075

Number of Simulators Used: 4

Remarks: *Admendment Dry Gas Lot #000919080A3
Exp 03/05/2021*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

William C Curry

 Signature and Printed Name
 07/30/2020
 Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
Time of Inspection: 17:48

Date of Inspection: 07/30/2020

Serial Number: 80-000770
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: N Exp: 03/05/2021
0.000	0.046	0.077	0.196	0.076
0.000	0.046	0.077	0.196	0.076
0.000	0.046	0.077	0.195	0.075

Number of Simulators Used: 4

Remarks: *Admendment Dry Gas Lot #000919080A3*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

William C Curry

Signature and Printed Name WILLIAM C CURRY

07/30/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
Time of Inspection: 18:18

Date of Inspection: 04/28/2020

Serial Number: 80-000770
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 00919080A3 Exp: 03/05/2021
0.000	0.046	0.078	0.196	
0.000	0.047	0.078	0.196	/ 0.076
0.000	0.047	0.078	0.196	/ 0.076
				/ 0.077

Number of Simulators Used: 4 *Revised Admin MPD +*

Remarks:

AF / MA: Ambient Fail. 08: RFI Detect.
*AF/MA Ambient + Kai - Aired out Room re conducted test
RFI Detect Turned my radio off*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

W. C. Curry

Signature and Printed Name WILLIAM C CURRY

04/28/2020
Date