

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.  
Time of Inspection: 17:07

Date of Inspection: 12/04/2020

Serial Number: 80-000824  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.199	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shannon Walsh*

SHANNON L WALSH

Signature and Printed Name

12/04/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.  
Time of Inspection: 00:06

Date of Inspection: 07/10/2020

Serial Number: 80-000824  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
0.000	0.048	0.078	0.202	0.081
0.000	0.049	0.079	0.203	0.082
0.000	0.049	0.079	0.203	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shannon L Walsh SHANNON L WALSH  
Signature and Printed Name

07/10/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.  
Time of Inspection: 20:13

Date of Inspection: 06/25/2020

Serial Number: 80-000824  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:261908A Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
0.000	0.047	0.077	0.199	0.080
0.000	0.047	0.078	0.200	0.080
0.000	0.048	0.078	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shannon L Walsh*

SHANNON L WALSH

Signature and Printed Name

06/25/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEEsburg P.D.  
Time of Inspection: 20:20

Date of Inspection: 05/22/2020

Serial Number: 80-000824  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908E Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30219060A3 Exp: 01/05/2022
0.000	0.050	0.079	0.194	0.081
0.000	0.050	0.080	0.196	0.080
0.000	0.051	0.080	0.198	0.080

Number of Simulators Used: 4

Remarks:  
In. DETAIL LEAK .

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shannon I. Walsh*

SHANNON I. WALSH

Signature and Printed Name

05/22/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.

Serial Number: 80-000824

Time of Inspection: 09:23

Date of Inspection: 04/24/2020

Software: 8100.27

Check or Test	YES	NO
Data and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201918D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
0.000	0.047	0.078	0.199	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080


Number of Simulators Used: 4

Remarks:

A F / M A: N O M A.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name

SHANNON L. WALSH

04/24/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.  
Time of Inspection: 00:29

Date of Inspection: 03/19/2020

Serial Number: 80-000824  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30219080A3 Exp: 01/05/2022
0.000	0.047	0.080	0.199	0.081
0.000	0.047	0.080	0.200	0.081
0.000	0.048	0.081	0.201	0.081

Number of Simulators Used: 4

Remarks:  
NONE

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JOHN G SOMMERSDORF

Signature and Printed Name

03/19/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

80-000824  
BK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.

Serial Number: 80-000824

Time of Inspection: 00:03

Date of Inspection: 02/25/2020

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30219980A3 Exp: 01/05/2022
0.000	0.050	0.077	0.198	0.081
0.000	0.050	0.079	0.200	0.082
0.000	0.050	0.079	0.200	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cpl. J. Sommersdore

JOHN G. SOMMERSDORE  
Signature and Printed Name

02/25/2020  
Date

824  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESSBURG P.D.

Serial Number: 80-000824

Time of Inspection: 13:56

Date of Inspection: 01/19/2020

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 2018110 Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 3491708HA7 Exp: 02/05/2020
0.000	0.050	0.079	0.195	0.080
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.200	0.079

Number of Simulators Used: 4

Remarks:

FIL EXCHANGED SOLUTION NEW SIMULATOR

AFTER 2 PREVIOUS FAILURES TO Detect Acetone, MIXED New Solution in SAME simulator. FAILED AGAIN. RECOVERED Acetone From LECO, Ran in Different simulator and it PASSED.

Ran 3 WET Control TESTS with original simulator and had Acetone Detected on All 3. Simulator will be taken out of SERVICE until I can obtain New hoses and fittings for it.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*John G. Sommerdurf*  
Signature and Printed Name

01/19/2020  
Date



824 SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.O.  
Time of Inspection: 09:37

Date of Inspection: 01/19/2020

Serial Number: 80-000824  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				
0.000				
0.000				

Number of Simulators Used: 1

Remarks:

Int. Set: . .

NEW BOTTLE OF ACETONE, FAILED TO DETECT. RETIRED JAR STILL FAILED.  
STARTED NEW INSPECTION WITH NEW ACETONE SOLUTION.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JOHN G. SOMMERSDORF

Signature and Printed Name


01/19/2020  
Date

LEESBURG P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000824  
01/19/2020  
Software: 8100.27

-----  
WEP CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	14:03
Control Test	.INT*	14:04
Air Blank	0.000	14:04

\*Interferent Detect

  
-----  
Operator's Signature

LEESBURG P.D.

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000824


01/19/2020

Software: 8100.27

| WET CONTROL TEST |

Test	g/210L	Time
Air Blank	0.000	14:05
Control Test	INT*	14:06
Air Blank	0.000	14:06

\*Interferent Detect


  
Operator's Signature

LEESBURG P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000824  
01/19/2020  
Software: 8100.27

-----  
WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	14:07
Control Test	INT*	14:08
Air Blank	0.000	14:08

\*Interferent Detect



-----  
Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.  
Time of Inspection: 01:04

Date of Inspection: 11/12/2020

Serial Number: 80-000824  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
0.000	0.047	0.078	0.198	0.080
0.000	0.048	0.078	0.200	0.080
0.000	0.048	0.079	0.201	0.079

Number of Simulators Used: 4

**Remarks:**

A F / M A:TOO MUCH MA CONSUMED.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SHANNON L WALSH  
Signature and Printed Name

11/12/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.  
Time of Inspection: 19:44

Date of Inspection: 10/28/2020

Serial Number: 80-000824  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
0.000	0.048	0.080	0.199	0.080
0.000	0.048	0.081	0.200	0.080
0.000	0.048	0.081	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SHANNON L WALSH

Signature and Printed Name

10/28/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.

Time of Inspection: 17:57

Date of Inspection: 09/28/2020

Serial Number: 80-000824

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910B Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30219080A3 Exp: 01/05/2022
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SHANNON L WALSH

Signature and Printed Name

09/28/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.

Time of Inspection: 17:10

Date of Inspection: 08/07/2020

Serial Number: 80-000824

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30219080A3 Exp: 01/05/2022
0.000	0.000 / 0.047	0.081	0.198	0.080
0.000	0.000 / 0.048	0.081	0.200	0.080
0.000	0.000 / 0.048	0.082	0.201	0.080

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance.

*Air leak with simulator with .05. tightened jar reran test after allowing 20 minutes*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shannon Walsh*

SHANNON L WALSH  
Signature and Printed Name

08/07/2020  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.  
Time of Inspection: 17:10

Date of Inspection: 08/07/2020

Serial Number: 80-000824  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30219080A3 Exp: 01/05/2022
0.000	0.000 / 0.047	0.081	0.198	0.080
0.000	0.000 / 0.048	0.081	0.200	0.080
0.000	0.000 / 0.048	0.082	0.201	0.080

Number of Simulators Used: 5

**Remarks:**

05: Control Outside Tolerance.

*Air leak with simulator with .05. tightened jar reran test after allowing 20 minutes.*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shannon L Walsh*

SHANNON L WALSH

Signature and Printed Name

08/07/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

80-000824  
BK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.

Time of Inspection: 13:56

Date of Inspection: 01/19/2020

Serial Number: 80-000824

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34917088A7 Exp: 02/05/2020
0.000	0.050	0.079	0.195	0.080
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.200	0.079

Number of Simulators Used: 4

**Remarks:**


FOR DELICIOUS SOLUTION NEW SIMULATOR

AFTER 2 PREVIOUS FAILURES TO DETECT ACETONE, MIXED NEW SOLUTION IN SAME SIMULATOR. FAILED AGAIN. RECOVERED ALCOHOL FROM LCSD, RAN IN DIFFERENT SIMULATOR AND IT PASSED.

RAN 3 WET CONTROL TESTS WITH ORIGINAL SIMULATOR AND HAD ACETONE DETECTED ON ALL 3. SIMULATOR WILL BE TAKEN OUT OF SERVICE UNTIL I CAN OBTAIN NEW BUSES AND FITTINGS FOR IT.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JOHN G. SOMMERENDE

Signature and Printed Name

01/19/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

80-000824  
BK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.

Time of Inspection: 09:37

Date of Inspection: 01/19/2020

Serial Number: 80-000824

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				
0.000				
0.000				

Number of Simulators Used: 4


Remarks:

Int. Ref: . . .

NEW BOTTLE OF ACETONE, FAILED TO DETECT. RETIREMENT JAR STILL FOLDED.  
STARTED NEW INSPECTION WITH NEW ACETONE SOLUTION.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JOHN G. SOMMERSDORF

Signature and Printed Name

01/19/2020  
Date