Florida Department of Law Enforcement **Alcohol Testing Program**

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BLOUNTSTOWN PD Time of Inspection: 10:21

Date of Inspection: 10/15/2020

Serial Number: 80-000862

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1076482 Exp: 01/15/2022
0.000	0.050	0.081	0.203	/ 0.081
0.000	0.050	0.080	0.203	
0.000	0.050	0.081	0.203	

Number	٥f	Simulators	IIaod.	_	
nmmer	OL	PIMUTATOLS	usea:		

Remarks:

18: RFI Detect, RFI Detect. UNKNOWN ROOM CLEAREDNON-compliance: RFI DURING AIR BLANK MISS

Another intox \$759 was next to it \$

As RFI's.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I) hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

SUSAN M BARGE

10/15/2020

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BLOUNTSTOWN PD
Time of Inspection: 09:38 Date of Inspection: 10/15/2020

Serial Number: 80-000862

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
		No
Interferent Detect Test: Interferent Detect		
		No
Diagnostic Check (Post-Inspection): OK		
-		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5	
Remarks:	
A F / M A: RFI Detect. Non-compliance:	
IPhon Watch	
	\cap
Stated Inspection	Ove-

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

SUSAN M BARGE

10/15/2020 Date