

964 SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Orange County Sheriff's Office

Serial Number: 80-000964

Time of Inspection: N/A

Date of Inspection: 02/11/2020

Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted	N/A	N/A
Diagnostic Check (Pre-Inspection): OK	N/A	N/A
Alcohol Free Subject Test: 0.000	N/A	N/A
Mouth Alcohol Test: Slope Not Met	N/A	N/A
Interferent Detect Test: Interferent Detect	N/A	N/A
Diagnostic Check (Post-Inspection): OK	N/A	N/A

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.08g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.20g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: N/A Exp: N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A

Number of Simulators Used:     N/A    

Remarks: Not able to perform an Agency Inspection due to the instrument not powering up when the "start test" button is pushed. I will notify Department Inspector Shayla Platt via telephone of this issue. I will send the instrument to CMI for analysis.

(All of the information entered on this Agency Inspection Report was hand-typed by Agency Inspector K. Melville)

The above instrument complies (   N/A   ) does not comply (   N/A   ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OK

Kelly M. Melville

Signature and Printed Name

02/11/2020  
Date