

Florida Department of Law Enforcement Alcohol Testing Program

80-001043
TSK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 02:00

Date of Inspection: 02/13/2020

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 13513080A6 Exp: 08/05/2020
0.000	0.046	0.073 / 0.077	0.195	0.079
0.000	0.046	0.074 / 0.077	0.196	0.079
0.000	0.047	0.077 / 0.077	0.198	0.080

Number of Simulators Used: 5

Remarks:
OR: Control Outside Tolerance.

*

Repositioned Simulator
Did swab second test.

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY _____ D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO

Signature and Printed Name

02/13/2020
Date