

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO  
Time of Inspection: 19:40

Date of Inspection: 05/30/2020

Serial Number: 80-001077  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1076482 Exp: 01/15/2022
/ 0.000	0.047	0.078	0.196	0.078
/ 0.000	0.048	0.078	0.198	0.078
/ 0.000	0.048	0.079	0.199	0.077

Number of Simulators Used: 5

Remarks:

00: Ambient Fail. Everything was connected unsure why this happened.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY W CARTWRIGHT JR

Signature and Printed Name

05/30/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO  
Time of Inspection: 11:05

Date of Inspection: 07/04/2020

Serial Number: 80-001077  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1076482 Exp: 01/15/2022
0.000	0.047	0.077	0.195	0.077
0.000	0.047	0.077	0.197	0.078
0.000	0.047	0.078	0.198	0.078

Number of Simulators Used: 5

Remarks:

JUNE INSPECTION Had to conduct second inspection due to 5 day period expiring

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY W CARTWRIGHT JR

Signature and Printed Name

07/04/2020  
Date